

ABX Minotrol Retic



LOT RX 416
Rev 1

CONTROL

(Exp.) 2019-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	4.7	4.7	4.7	± 1.9	11.8	11.8	11.8	± 3.5
RETIC	%	1.60	1.72	1.72	± 0.86	5.18	5.46	5.46	± 1.37	10.50	11.10	11.10	± 2.20
RETIC	10 ⁶ /µl	0.055	0.059	0.059	± 0.030	0.180	0.189	0.189	± 0.047	0.359	0.379	0.379	± 0.076
MFI MRI *		28.9	27.9	27.9	± 8.0	47.4	46.8	46.8	± 10.0	51.9	52.2	52.2	± 12.0
GR RBC	10 ⁶ /µl	3.43	3.42	3.42	± 0.30	3.47	3.47	3.47	± 0.30	3.42	3.41	3.41	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 416
Rev 1

CONTROL

(Exp.) 2019-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	4.7			± 1.9	11.8			± 3.5
RETIC	%	1.68			± 0.84	5.34			± 1.34	10.90			± 2.20
RETIC	10 ⁶ /µl	0.057			± 0.029	0.183			± 0.046	0.367			± 0.073
MFI MRI *		26.9			± 8.0	45.6			± 10.0	50.9			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.42			± 0.30	3.37			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA