

ABX Minotrol Retic



LOT RX 417
Rev 1

CONTROL

(Exp.) 2019-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.3	1.3	1.3	± 0.7	5.4	5.4	5.4	± 2.2	10.9	10.9	10.9	± 3.3
RETIC	%	1.95	2.11	2.11	± 1.06	5.50	5.89	5.89	± 1.48	10.40	11.20	11.20	± 2.30
RETIC	10 ⁶ /µl	0.068	0.073	0.073	± 0.037	0.190	0.202	0.202	± 0.051	0.359	0.384	0.384	± 0.077
MFI MRI *		32.5	32.2	32.2	± 8.0	47.4	48.1	48.1	± 10.0	50.6	50.1	50.1	± 12.0
GR RBC	10 ⁶ /µl	3.47	3.45	3.45	± 0.30	3.45	3.43	3.43	± 0.30	3.45	3.43	3.43	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 417
Rev 1

CONTROL

(Exp.) 2019-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	5.4			± 2.2	10.9			± 3.3
RETIC	%	1.98			± 0.99	5.68			± 1.42	10.83			± 2.17
RETIC	10 ⁶ /µl	0.068			± 0.034	0.194			± 0.049	0.367			± 0.073
MFI MRI *		31.4			± 8.0	47.0			± 10.0	50.0			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.41			± 0.30	3.39			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA