

ABX Minotrol Retic



LOT RX 427
Rev 1

CONTROL

(Exp.) 2021-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.9	6.0	6.0	6.0	± 2.7	11.4	11.4	11.4	± 3.4
RETIC	%	1.66	1.66	1.66	± 0.83	5.06	5.06	5.06	± 1.27	10.30	10.30	10.30	± 2.10
RETIC	10 ⁶ /µl	0.057	0.057	0.057	± 0.029	0.172	0.172	0.172	± 0.043	0.348	0.348	0.348	± 0.070
MFI MRI *	%	32.6	32.6	32.6	± 8.0	45.8	45.8	45.8	± 10.0	50.5	50.5	50.5	± 12.0
GR RBC	10 ⁶ /µl	3.41	3.41	3.41	± 0.30	3.40	3.40	3.40	± 0.30	3.38	3.38	3.38	± 0.30
PIC RANGE	-	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 427
Rev 1

CONTROL

(Exp.) 2021-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.5			± 0.9	6.0			± 2.7	11.4			± 3.4
RETIC	%	1.83			± 0.92	5.16			± 1.29	10.36			± 2.08
RETIC	10 ⁶ /µl	0.062			± 0.031	0.174			± 0.044	0.347			± 0.070
MFI MRI *	%	29.3			± 8.0	43.5			± 10.0	47.8			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.37			± 0.30	3.35			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA