

ABX Minotrol Retic



LOT RX 428
Rev 1

CONTROL

(Exp.) 2021-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus		DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%		1.6	1.6	± 1.0		5.6	5.6	± 2.2		11.3	11.3	± 3.4
RETIC	%		1.58	1.58	± 0.79		5.00	5.00	± 1.25		10.60	10.60	± 2.10
RETIC	10 ⁶ /µl		0.053	0.053	± 0.027		0.169	0.169	± 0.042		0.360	0.360	± 0.072
MFI MRI *	%		29.7	29.7	± 8.0		46.0	46.0	± 10.0		50.3	50.3	± 12.0
GR RBC	10 ⁶ /µl		3.35	3.35	± 0.30		3.38	3.38	± 0.30		3.40	3.40	± 0.30
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 428
Rev 1

CONTROL

(Exp.) 2021-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.6			± 1.0	5.6			± 2.2	11.3			± 3.4
RETIC	%	1.71			± 0.86	5.10			± 1.28	10.70			± 2.10
RETIC	10 ⁶ /µl	0.057			± 0.029	0.170			± 0.043	0.358			± 0.072
MFI MRI *	%	25.4			± 8.0	42.3			± 10.0	48.6			± 12.0
GR RBC	10 ⁶ /µl	3.31			± 0.30	3.33			± 0.30	3.35			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA