

ABX Minotrol Retic



LOT RX 421
Rev 1

CONTROL

(Exp.) 2020-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.7	1.7	1.7	± 0.9	7.1	7.1	7.1	± 2.8	11.3	11.3	11.3	± 3.4
RETIC	%	1.77	1.99	1.99	± 1.00	5.15	5.47	5.47	± 1.37	10.40	11.00	11.00	± 2.20
RETIC	10 ⁶ /µl	0.060	0.068	0.068	± 0.034	0.176	0.187	0.187	± 0.047	0.359	0.380	0.380	± 0.076
MFI MRI *		35.2	34.8	34.8	± 8.0	47.0	46.2	46.2	± 10.0	51.5	50.7	50.7	± 12.0
GR RBC	10 ⁶ /µl	3.41	3.42	3.42	± 0.30	3.42	3.42	3.42	± 0.30	3.45	3.45	3.45	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 421
Rev 1

CONTROL

(Exp.) 2020-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.7			± 0.9	7.1			± 2.8	11.3			± 3.4
RETIC	%	2.13			± 1.07	5.54			± 1.39	10.97			± 2.19
RETIC	10 ⁶ /µl	0.072			± 0.036	0.186			± 0.047	0.372			± 0.074
MFI MRI *		33.5			± 8.0	45.1			± 10.0	50.0			± 12.0
GR RBC	10 ⁶ /µl	3.36			± 0.30	3.36			± 0.30	3.39			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA