

# ABX Minotrol Retic



**LOT** RX 095  
Rev 1

**CONTROL**

(Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.8	5.7	5.7	5.7	± 2.3	10.3	10.3	10.3	± 3.1
RETIC	%	2.43	2.44	2.44	± 1.22	5.65	5.87	5.87	± 1.47	11.70	11.80	11.80	± 2.40
RETIC	10 <sup>6</sup> /µl	0.083	0.084	0.084	± 0.042	0.194	0.203	0.203	± 0.051	0.401	0.405	0.405	± 0.081
MFI MRI *		33.5	33.5	33.5	± 8.0	45.1	44.5	44.5	± 10.0	49.9	48.9	48.9	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.42	3.45	3.45	± 0.30	3.44	3.45	3.45	± 0.30	3.43	3.43	3.43	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 095  
Rev 1

**CONTROL**

(Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>5.7</b>			± 2.3	<b>10.3</b>			± 3.1
RETIC	%	<b>2.44</b>			± 1.22	<b>5.67</b>			± 1.42	<b>11.50</b>			± 2.30
RETIC	10 <sup>6</sup> /µl	<b>0.083</b>			± 0.042	<b>0.193</b>			± 0.048	<b>0.390</b>			± 0.078
MFI MRI *		<b>35.9</b>			± 8.0	<b>46.4</b>			± 10.0	<b>49.8</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.41</b>			± 0.30	<b>3.40</b>			± 0.30	<b>3.39</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

\* For research use only in Canada and USA