

ABX Minotrol Retic



LOT RX 094
Rev 1

CONTROL

(Exp.) 2014-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual													
RETIC	%	1.6	1.6	1.6	± 0.8	5.8	5.8	5.8	± 2.3	11.0	11.0	11.0	± 3.3
RETIC	%	2.07	2.21	2.21	± 1.11	5.50	5.64	5.64	± 1.41	11.40	11.60	11.60	± 2.30
RETIC	10 ⁶ /µl	0.070	0.075	0.075	± 0.038	0.183	0.187	0.187	± 0.047	0.383	0.389	0.389	± 0.078
MFI MRI *		29.2	30.3	30.3	± 8.0	42.3	42.8	42.8	± 10.0	47.8	48.5	48.5	± 12.0
GR RBC	10 ⁶ /µl	3.39	3.38	3.38	± 0.30	3.32	3.31	3.31	± 0.30	3.36	3.35	3.35	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 094

CONTROL

 (Exp.) 2014-11-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.6			± 0.8	5.8			± 2.3	11.0			± 3.3
RETIC	%	2.15			± 1.08	5.45			± 1.36	11.10			± 2.20
RETIC	10 ⁶ /µl	0.073			± 0.037	0.182			± 0.046	0.373			± 0.075
MFI MRI *		28.4			± 8.0	40.1			± 10.0	46.8			± 12.0
GR RBC	10 ⁶ /µl	3.41			± 0.30	3.34			± 0.30	3.36			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083 -A