

ABX Minotrol Retic



LOT RX 075
Rev 1

CONTROL

(Exp.) 2015-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.6	1.6	1.6	± 0.8	5.8	5.8	5.8	± 2.3	10.7	10.7	10.7	± 3.2
RETIC	%	1.62	1.84	1.84	± 0.92	5.35	5.61	5.61	± 1.40	11.40	11.60	11.60	± 2.30
RETIC	10 ⁶ /µl	0.058	0.066	0.066	± 0.033	0.192	0.202	0.202	± 0.051	0.397	0.404	0.404	± 0.081
MFI MRI *		28.6	28.5	28.5	± 8.0	45.4	45.3	45.3	± 10.0	49.6	51.9	51.9	± 12.0
GR RBC	10 ⁶ /µl	3.56	3.57	3.57	± 0.30	3.59	3.60	3.60	± 0.30	3.48	3.48	3.48	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 075
Rev 1

CONTROL

(Exp.) 2015-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.6			± 0.8	5.8			± 2.3	10.7			± 3.2
RETIC	%	1.78			± 0.89	5.45			± 1.36	11.30			± 2.30
RETIC	10 ⁶ /µl	0.063			± 0.032	0.196			± 0.049	0.393			± 0.079
MFI MRI *		25.9			± 8.0	44.3			± 10.0	49.5			± 12.0
GR RBC	10 ⁶ /µl	3.56			± 0.30	3.60			± 0.30	3.48			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA