

ABX Minotrol Retic



LOT RX 115
Rev 1

CONTROL

(Exp.) 2016-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	6.0	6.0	6.0	± 2.4	11.4	11.4	11.4	± 3.4
RETIC	%	1.86	2.01	2.01	± 1.01	5.39	5.50	5.50	± 1.38	11.50	11.80	11.80	± 2.40
RETIC	10 ⁶ /µl	0.063	0.069	0.069	± 0.035	0.184	0.189	0.189	± 0.047	0.391	0.402	0.402	± 0.080
MFI MRI *		28.9	27.4	27.4	± 8.0	43.2	43.9	43.9	± 10.0	47.0	47.9	47.9	± 12.0
GR RBC	10 ⁶ /µl	3.39	3.42	3.42	± 0.30	3.42	3.43	3.43	± 0.30	3.40	3.41	3.41	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 115
Rev 1

CONTROL

(Exp.) 2016-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	6.0			± 2.4	11.4			± 3.4
RETIC	%	2.01			± 1.01	5.40			± 1.35	11.70			± 2.40
RETIC	10 ⁶ /µl	0.063			± 0.032	0.185			± 0.046	0.397			± 0.079
MFI MRI *		26.5			± 8.0	42.7			± 10.0	47.9			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.42			± 0.30	3.39			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A

* For research use only in Canada and USA