

ABX Minotrol Retic



LOT RX 411
Rev 1

CONTROL

(Exp.) 2018-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.8	6.0	6.0	6.0	± 2.4	10.8	10.8	10.8	± 3.2
RETIC	%	2.15	2.27	2.27	± 1.14	5.34	5.65	5.65	± 1.41	10.90	11.60	11.60	± 2.40
RETIC	10 ⁶ /µl	0.076	0.078	0.078	± 0.039	0.190	0.194	0.194	± 0.049	0.389	0.400	0.400	± 0.080
MFI MRI *		34.4	33.9	33.9	± 8.0	45.2	45.8	45.8	± 10.0	49.4	51.0	51.0	± 12.0
GR RBC	10 ⁶ /µl	3.54	3.43	3.43	± 0.30	3.56	3.44	3.44	± 0.30	3.57	3.45	3.45	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 411
Rev 1

CONTROL

(Exp.) 2018-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.8	6.0			± 2.4	10.8			± 3.2
RETIC	%	2.17			± 1.09	5.47			± 1.37	11.20			± 2.24
RETIC	10 ⁶ /µl	0.074			± 0.037	0.185			± 0.046	0.380			± 0.076
MFI MRI *		33.3			± 8.0	45.1			± 10.0	50.1			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.38			± 0.30	3.39			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA