

# ABX Minotrol Retic



**LOT** RX 056  
Rev 1

**CONTROL**

(Exp.) 2016-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.6	1.6	1.6	± 0.8	5.5	5.5	5.5	± 2.2	10.2	10.2	10.2	± 3.1
RETIC	%	2.10	2.11	2.11	± 1.06	5.66	5.66	5.66	± 1.42	11.70	11.60	11.60	± 2.40
RETIC	10 <sup>6</sup> /µl	0.073	0.072	0.072	± 0.037	0.195	0.195	0.195	± 0.049	0.401	0.394	0.394	± 0.082
MFI MRI *		30.6	31.0	31.0	± 8.0	44.7	44.8	44.8	± 10.0	49.2	50.6	50.6	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.47	3.43	3.43	± 0.30	3.45	3.45	3.45	± 0.30	3.43	3.40	3.40	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.34 FRONT / RECTO 9930083 -A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 056  
Rev 1

**CONTROL**

(Exp.) 2016-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 0.8	<b>5.5</b>			± 2.2	<b>10.2</b>			± 3.1
RETIC	%	<b>2.00</b>			± 1.00	<b>5.47</b>			± 1.37	<b>11.30</b>			± 2.26
RETIC	10 <sup>6</sup> /µl	<b>0.069</b>			± 0.035	<b>0.187</b>			± 0.047	<b>0.385</b>			± 0.077
MFI MRI *		<b>28.9</b>			± 8.0	<b>43.1</b>			± 10.0	<b>49.1</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.41</b>			± 0.30	<b>3.41</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.34 BACK / VERSO 9930083-A