

ABX Minotrol Retic



LOT RX 407
Rev 1

CONTROL

(Exp.) 2017-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual RETIC	%	1.4	1.4	1.4	± 0.7	5.9	5.9	5.9	± 2.4	11.2	11.2	11.2	± 3.4
RETIC	%	2.30	2.24	2.24	± 1.15	5.49	5.73	5.73	± 1.43	10.80	11.30	11.30	± 2.30
RETIC	10 ⁶ /µl	0.078	0.077	0.077	± 0.039	0.187	0.195	0.195	± 0.049	0.373	0.391	0.391	± 0.078
MFI MRI *		27.7	30.1	30.1	± 8.0	39.5	40.9	40.9	± 10.0	43.6	43.7	43.7	± 12.0
GR RBC	10 ⁶ /µl	3.41	3.42	3.42	± 0.30	3.40	3.41	3.41	± 0.30	3.45	3.46	3.46	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 407
Rev 1

CONTROL

(Exp.) 2017-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	5.9			± 2.4	11.2			± 3.4
RETIC	%	2.28			± 1.14	5.45			± 1.36	10.90			± 2.20
RETIC	10 ⁶ /µl	0.078			± 0.039	0.186			± 0.047	0.377			± 0.075
MFI MRI *		28.8			± 8.0	39.9			± 10.0	44.5			± 12.0
GR RBC	10 ⁶ /µl	3.42			± 0.30	3.41			± 0.30	3.46			± 0.30
PIC RANGE		8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

* For research use only in Canada and USA