

# ABX Minotrol Retic



<b>LOT</b> RX 408		<b>CONTROL</b>							(Exp.) 2018-01-05 (YYYY - MM - DD)				
Rev 2		<b>ABX</b>											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA			
		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus		120 RETIC	DX120	DX Nexus	
Méthode Manuelle / Manual													
RETIC	%	1.4	1.4	1.4	± 0.7	6.3	6.3	6.3	± 2.5	10.9	10.9	10.9	± 3.3
RETIC	%	2.52	2.60	2.60	± 1.30	5.71	5.75	5.75	± 1.44	11.50	11.80	11.80	± 2.40
RETIC	10 <sup>6</sup> /µl	0.087	0.091	0.091	± 0.046	0.198	0.200	0.200	± 0.050	0.398	0.408	0.408	± 0.082
MFI MRI *		36.9	36.9	36.9	± 8.0	43.9	43.9	43.9	± 10.0	51.9	51.9	51.9	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.45	3.49	3.49	± 0.30	3.47	3.47	3.47	± 0.30	3.46	3.46	3.46	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

# ABX Minotrol Retic



<b>LOT</b> RX 408		<b>CONTROL</b>			(Exp.) 2018-01-05 (YYYY - MM - DD)								
Rev 2		<b>ABX</b>											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		XLR				XLR				XLR			
Méthode Manuelle / Manual													
RETIC	%	<b>1.4</b>			± 0.7	<b>6.3</b>			± 2.5	<b>10.9</b>			± 3.3
RETIC	%	<b>2.24</b>			± 1.12	<b>5.50</b>			± 1.38	<b>11.60</b>			± 2.32
RETIC	10 <sup>6</sup> /µl	<b>0.077</b>			± 0.039	<b>0.189</b>			± 0.047	<b>0.396</b>			± 0.079
MFI MRI *		<b>35.1</b>			± 8.0	<b>41.1</b>			± 10.0	<b>45.8</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.45</b>			± 0.30	<b>3.44</b>			± 0.30	<b>3.41</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A