

# ABX Minotrol Retic



**LOT** RX 409  
Rev 1

**CONTROL**

(Exp.) 2018-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		1				2				3			
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus	120 RETIC	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	1.5	± 0.9	5.5	5.5	5.5	± 2.2	11.4	11.4	11.4	± 3.4
RETIC	%	2.18	2.17	2.17	± 1.09	5.35	5.65	5.65	± 1.42	10.90	11.60	11.60	± 2.40
RETIC	10 <sup>6</sup> /µl	0.076	0.075	0.075	± 0.038	0.190	0.199	0.199	± 0.050	0.390	0.412	0.412	± 0.082
MFI MRI *		29.1	26.8	26.8	± 8.0	45.4	43.8	43.8	± 10.0	51.2	51.5	51.5	± 12.0
GR RBC	10 <sup>6</sup> /µl	3.48	3.45	3.45	± 0.30	3.56	3.53	3.53	± 0.30	3.58	3.55	3.55	± 0.30
PIC RANGE		27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0	27.0	27.0	27.0	± 5.0

Ref: TEMP-0382 Rev.36 FRONT / RECTO 1219930083-A

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 409  
Rev 1

**CONTROL**

(Exp.) 2018-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA		
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.9	<b>5.5</b>			± 2.2	<b>11.4</b>			± 3.4
RETIC	%	<b>1.92</b>			± 0.96	<b>5.29</b>			± 1.32	<b>11.20</b>			± 2.30
RETIC	10 <sup>6</sup> /µl	<b>0.066</b>			± 0.033	<b>0.186</b>			± 0.047	<b>0.396</b>			± 0.079
MFI MRI *		<b>27.3</b>			± 8.0	<b>45.1</b>			± 10.0	<b>52.5</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.45</b>			± 0.30	<b>3.52</b>			± 0.30	<b>3.54</b>			± 0.30
PIC RANGE		<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.36 BACK / VERSO 1219930083-A

\* For research use only in Canada and USA