

ABX Minotrol Retic



LOT RX 430
Rev 1

CONTROL

(Exp.) 2021-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
Méthode Manuelle / Manual RETIC	%		DX120 1.4	DX Nexus 1.4	± 0.7		DX120 6.1	DX Nexus 6.1	± 2.4		DX120 11.1	DX Nexus 11.1	± 3.3
RETIC	%		1.66	1.66	± 0.83		5.38	5.38	± 1.35		11.10	11.10	± 2.20
RETIC	10 ⁶ /µl		0.057	0.057	± 0.029		0.184	0.184	± 0.046		0.379	0.379	± 0.076
MFI MRI *	%		31.4	31.4	± 8.0		46.0	46.0	± 10.0		50.6	50.6	± 12.0
GR RBC	10 ⁶ /µl		3.46	3.46	± 0.30		3.42	3.42	± 0.30		3.41	3.41	± 0.30
PIC RANGE	-		27.0	27.0	± 5.0		27.0	27.0	± 5.0		27.0	27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.38 FRONT / RECTO 1219930083-A

ABX Minotrol Retic



LOT RX 430
Rev 1

CONTROL

(Exp.) 2021-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA	
XLR			XLR			XLR			XLR				
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	6.1			± 2.4	11.1			± 3.3
RETIC	%	1.64			± 0.82	5.55			± 1.39	11.56			± 2.31
RETIC	10 ⁶ /µl	0.054			± 0.027	0.181			± 0.045	0.378			± 0.076
MFI MRI *	%	32.0			± 8.0	45.2			± 10.0	48.7			± 12.0
GR RBC	10 ⁶ /µl	3.31			± 0.30	3.26			± 0.30	3.27			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.38 BACK / VERSO 1219930083-A

* For research use only in Canada and USA