

# ABX Minotrol Retic



**LOT** RX 404  
Rev 2

**CONTROL**

(Exp.) 2017-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		YUMIZEN		1		YUMIZEN		2		YUMIZEN		3	
H2500			H2500			H2500							
Méthode Manuelle / Manual RETIC	%	<b>1.7</b>			± 0.9	<b>5.5</b>			± 2.2	<b>12.0</b>			± 4.2
RETIC	%	<b>2.08</b>			± 1.04	<b>5.84</b>			± 1.36	<b>11.61</b>			± 2.20
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.074</b>			± 0.037	<b>0.202</b>			± 0.047	<b>0.399</b>			± 0.077
	10 <sup>9</sup> /l	<b>74</b>			± 37	<b>202</b>			± 47	<b>399</b>			± 77
MFI MRI *		<b>28.6</b>			± 8.0	<b>42.7</b>			± 10.0	<b>46.2</b>			± 12.0
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.54</b>			± 0.30	<b>3.46</b>			± 0.30	<b>3.44</b>			± 0.30
PIC RANGE		<b>14</b>			± 4	<b>14</b>			± 4	<b>14</b>			± 4

Ref: TEMP-0382 Rev.36 1300032515-A

\* For research use only in Canada and USA