

# ABX Minotrol Retic



**LOT** RX 428  
Rev 1

**CONTROL**

(Exp.) 2021-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 1.0	<b>5.6</b>			± 2.2	<b>11.3</b>			± 3.4
RETIC	%	<b>1.62</b>			± 0.81	<b>5.27</b>			± 1.32	<b>11.10</b>			± 2.20
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.055</b>			± 0.028	<b>0.180</b>			± 0.045	<b>0.382</b>			± 0.076
	10 <sup>9</sup> /l	<b>55</b>			± 28	<b>180</b>			± 45	<b>382</b>			± 76
MFI MRI *	%	<b>26.7</b>			± 8.0	<b>41.3</b>			± 10.0	<b>46.9</b>			± 12.0
	-	<b>0.267</b>			± 0.080	<b>0.413</b>			± 0.100	<b>0.469</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.39</b>			± 0.30	<b>3.42</b>			± 0.30	<b>3.44</b>			± 0.30
PIC RANGE	-	<b>13</b>			± 4	<b>13</b>			± 4	<b>13</b>			± 4

1300080026  
Ref: TEMP-0382 Rev.38

\* For research use only in Canada and USA