

# ABX Minotrol Retic



**LOT** RX 427  
Rev 1

**CONTROL**

(Exp.) 2021-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.9	<b>6.0</b>			± 2.7	<b>11.4</b>			± 3.4
RETIC	%	<b>1.80</b>			± 0.90	<b>5.44</b>			± 1.36	<b>11.17</b>			± 2.24
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.062</b>			± 0.031	<b>0.188</b>			± 0.047	<b>0.384</b>			± 0.077
	10 <sup>9</sup> /l	<b>62</b>			± 31	<b>188</b>			± 47	<b>384</b>			± 77
MFI MRI *	%	<b>29.7</b>			± 8.0	<b>41.3</b>			± 10.0	<b>45.4</b>			± 12.0
	-	<b>0.297</b>			± 0.080	<b>0.413</b>			± 0.100	<b>0.454</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.47</b>			± 0.30	<b>3.45</b>			± 0.30	<b>3.44</b>			± 0.30
PIC RANGE	-	<b>13</b>			± 4	<b>13</b>			± 4	<b>13</b>			± 4

1300080026  
Ref: TEMP-0382 Rev.38

\* For research use only in Canada and USA