

ABX Minotrol Retic



LOT RX 414

CONTROL

(Exp.) 2019-01-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		YUMIZEN		1		YUMIZEN		2		YUMIZEN		3	
		H2500				H2500				H2500			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.8	5.3			± 2.6	10.6			± 3.2
RETIC	%	1.48			± 0.74	5.07			± 1.27	11.00			± 2.20
RETIC	10 ⁶ /mm ³	0.052			± 0.026	0.179			± 0.045	0.386			± 0.077
	10 ⁹ /l	52			± 26	179			± 45	386			± 77
MFI MRI *		23.7			± 8.0	39.2			± 10.0	44.0			± 12.0
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.49			± 0.30	3.54			± 0.30	3.51			± 0.30
PIC RANGE		13			± 4	13			± 4	13			± 4

1300032515-A
Ref: TEMP-0382 Rev.36

* For research use only in Canada and USA