

# ABX Minotrol Retic



**LOT** RX 423  
Rev 1

**CONTROL**

(Exp.) 2020-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.8</b>			± 1.0	<b>5.4</b>			± 2.2	<b>10.6</b>			± 3.5
RETIC	%	<b>1.96</b>			± 0.98	<b>5.47</b>			± 1.37	<b>11.62</b>			± 2.32
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.070</b>			± 0.035	<b>0.195</b>			± 0.049	<b>0.415</b>			± 0.083
	10 <sup>9</sup> /l	<b>70</b>			± 35	<b>195</b>			± 49	<b>415</b>			± 83
MFI MRI *	%	<b>27.5</b>			± 8.0	<b>40.3</b>			± 10.0	<b>43.8</b>			± 12.0
	-	<b>0.275</b>			± 0.080	<b>0.403</b>			± 0.100	<b>0.438</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.55</b>			± 0.30	<b>3.56</b>			± 0.30	<b>3.57</b>			± 0.30
PIC RANGE	-	<b>13</b>			± 4	<b>13</b>			± 4	<b>13</b>			± 4

\* For research use only in Canada and USA

1300080026  
Ref: TEMP-0382 Rev.37