


ABX Minotrol Retic

LOT RX 408		CONTROL						 (Exp.) 2018-01-05 (YYYY - MM - DD)					
Rev 2		ABX											
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN H2500				YUMIZEN H2500				YUMIZEN H2500			
Méthode Manuelle / Manual													
RETIC	%	1.4			± 0.7	6.3			± 2.5	10.9			± 3.3
RETIC	%	2.28			± 1.14	5.71			± 1.43	12.10			± 2.42
RETIC	10 ⁶ /mm ³	0.080			± 0.040	0.199			± 0.050	0.420			± 0.084
	10 ⁹ /l	80			± 40	199			± 50	420			± 84
MFI MRI *		34.6			± 8.0	39.8			± 10.0	44.3			± 12.0
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.50			± 0.30	3.48			± 0.30	3.47			± 0.30
PIC RANGE		13			± 4	13			± 4	13			± 4

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.36 1300032515-A