

ABX Minotrol Retic



LOT RX 422
Rev 2

CONTROL

(Exp.) 2020-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	1.3			± 0.7	6.0			± 2.4	11.5			± 3.5
RETIC	%	1.64			± 0.82	5.70			± 1.43	11.62			± 2.32
RETIC	10 ⁶ /mm ³	0.058			± 0.029	0.202			± 0.051	0.411			± 0.082
	10 ⁹ /l	58			± 29	202			± 51	411			± 82
MFI MRI *	%	30.8			± 8.0	42.6			± 10.0	45.8			± 12.0
	-	0.308			± 0.080	0.426			± 0.100	0.458			± 0.120
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.55			± 0.30	3.55			± 0.30	3.54			± 0.30
PIC RANGE	-	13			± 4	13			± 4	13			± 4

1300080026
Ref: TEMP-0382 Rev.37

* For research use only in Canada and USA