

# ABX Minotrol Retic



**LOT** RX 419  
Rev 1

**CONTROL**

(Exp.) 2019-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>4.8</b>			± 2.1	<b>10.5</b>			± 3.2
RETIC	%	<b>1.57</b>			± 0.79	<b>5.18</b>			± 1.30	<b>10.98</b>			± 2.20
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.054</b>			± 0.027	<b>0.179</b>			± 0.045	<b>0.386</b>			± 0.077
	10 <sup>9</sup> /l	<b>54</b>			± 27	<b>179</b>			± 45	<b>386</b>			± 77
MFI MRI *		<b>26.4</b>			± 8.0	<b>40.1</b>			± 10.0	<b>43.6</b>			± 12.0
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.47</b>			± 0.30	<b>3.45</b>			± 0.30	<b>3.52</b>			± 0.30
PIC RANGE		<b>13</b>			± 4	<b>13</b>			± 4	<b>13</b>			± 4

Ref: TEMP-0382 Rev.36 1300032515-A

\* For research use only in Canada and USA