

ABX Minotrol Retic



LOT RX 430
Rev 1

CONTROL

(Exp.) 2021-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
		H2500				H2500				H2500			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	6.1			± 2.4	11.1			± 3.3
RETIC	%	1.67			± 0.84	5.76			± 1.44	11.81			± 2.36
RETIC	10 ⁶ /mm ³	0.058			± 0.029	0.198			± 0.050	0.404			± 0.081
	10 ⁹ /l	58			± 29	198			± 50	404			± 81
MFI MRI *	%	32.2			± 8.0	45.2			± 10.0	48.7			± 12.0
	-	0.322			± 0.080	0.452			± 0.100	0.487			± 0.120
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.46			± 0.30	3.43			± 0.30	3.42			± 0.30
PIC RANGE	-	13			± 4	13			± 4	13			± 4

1300080026
Ref: TEMP-0382 Rev.38

* For research use only in Canada and USA