

# ABX Minotrol Retic



**LOT** RX 434  
Rev 1

**CONTROL**

(Exp.) 2022-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.3</b>			± 0.7	<b>5.9</b>			± 2.4	<b>11.4</b>			± 3.4
RETIC	%	<b>1.68</b>			± 0.84	<b>5.33</b>			± 1.33	<b>11.44</b>			± 2.29
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.060</b>			± 0.030	<b>0.188</b>			± 0.047	<b>0.407</b>			± 0.081
	10 <sup>9</sup> /l	<b>60</b>			± 30	<b>188</b>			± 47	<b>407</b>			± 81
MFI MRI *	%	<b>32.8</b>			± 8.0	<b>46.4</b>			± 10.0	<b>49.0</b>			± 12.0
	-	<b>0.328</b>			± 0.080	<b>0.464</b>			± 0.100	<b>0.490</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.56</b>			± 0.30	<b>3.53</b>			± 0.30	<b>3.56</b>			± 0.30
PIC RANGE	-	<b>13</b>			± 8	<b>13</b>			± 8	<b>13</b>			± 8

1300080026  
Ref: TEMP-0382 Rev.40

\* For research use only in Canada and USA