


LOT PX 422
Rev 2


CONTROL

 (Exp.) 2020-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio															
		CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					
		80 XL80	XLR				80 XL80	XLR				80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5				± 0.4	7.5	7.5				± 1.0	17.8	17.8		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.35	2.35			± 0.16	4.67	4.67			± 0.20	5.25	5.25			± 0.25	
HB HGB	g/dl	6.6	6.6			± 0.4	14.0	14.0			± 0.5	16.7	16.7			± 0.6	
	g/l	66	66			± 4	140	140			± 5	167	167			± 6	
	mmol/l	4.10	4.10			± 0.25	8.69	8.69			± 0.31	10.37	10.37			± 0.37	
HT HCT	%	19.3	19.3			± 1.5	39.2	39.2			± 2.0	46.2	46.2			± 2.5	
	l/l	0.193	0.193			± 0.015	0.392	0.392			± 0.020	0.462	0.462			± 0.025	
VGM MCV	µm ³ ; fl	82	82			± 5	84	84			± 5	88	88			± 5	
TGMH MCH	pg	28.1	28.1			± 2.0	30.0	30.0			± 2.0	31.8	31.8			± 2.5	
	fmol	1.74	1.74			± 0.12	1.86	1.86			± 0.12	1.98	1.98			± 0.16	
CCMH MCHC	g/dl	34.3	34.3			± 3.0	35.7	35.7			± 3.0	36.1	36.1			± 3.0	
	g/l	343	343			± 30	357	357			± 30	361	361			± 30	
	mmol/l	21.27	21.27			± 1.86	22.16	22.16			± 1.86	22.45	22.45			± 1.86	
IDR RDW	%	14.0	13.0			± 4.0	14.0	12.5			± 4.0	13.5	12.0			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	63	63			± 20	241	241			± 30	487	487			± 50	
VPM MPV	µm ³ ; fl	10.4	10.4			± 2.0	9.6	9.6			± 2.0	9.8	9.8			± 2.0	
NEUT	#	1.39	1.39			± 0.35	4.22	4.24			± 0.90	12.44	12.34			± 1.90	
	%	55.7	55.5			± 10.0	56.2	56.5			± 10.0	69.9	69.3			± 10.0	
LYMPHO	#	0.77	0.78			± 0.33	2.46	2.41			± 0.70	2.83	2.83			± 1.50	
	%	30.7	31.0			± 12.0	32.8	32.1			± 8.0	15.9	15.9			± 8.0	
MONO	#	0.09	0.10			± 0.09	0.27	0.32			± 0.27	0.77	0.87			± 0.77	
	%	3.4	3.9			± 3.4	3.6	4.2			± 3.6	4.3	4.9			± 4.3	
EOS	#	0.17	0.16			± 0.16	0.29	0.28			± 0.28	0.98	0.96			± 0.96	
	%	6.9	6.3			± 6.3	3.9	3.7			± 3.7	5.5	5.4			± 5.4	
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.78	0.80			± 0.78	
	%	3.3	3.3			± 3.3	3.5	3.5			± 3.5	4.4	4.5			± 4.4	

LOT PX 422
Rev 2

CONTROL

 (Exp.) 2020-05-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.4	7.4	7.6			± 1.0	17.9	17.8	18.1			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.36	2.34			± 0.16	4.65	4.65	4.63			± 0.20	5.21	5.20	5.23			± 0.25	
	g/dl	6.6	6.5	6.6			± 0.4	13.9	13.8	13.9			± 0.5	16.7	16.5	16.6			± 0.6	
HB HGB	g/l	66	65	66			± 4	139	138	139			± 5	167	165	166			± 6	
	mmol/l	4.10	4.04	4.10			± 0.25	8.63	8.57	8.63			± 0.31	10.37	10.25	10.31			± 0.37	
HT HCT	%	19.1	18.9	18.4			± 1.5	38.6	38.1	37.7			± 2.0	45.8	45.2	44.5			± 2.5	
	l/l	0.191	0.189	0.184			± 0.015	0.386	0.381	0.377			± 0.020	0.458	0.452	0.445			± 0.025	
VGM MCV	µm ³ ; fl	80	80	78.5			± 5	83	82	81.5			± 5	88	87	85.0			± 5	
TGMH MCH	pg	27.6	27.5	28.2			± 2.0	29.9	29.7	30.0			± 2.0	32.1	31.7	31.7			± 2.5	
	fmol	1.71	1.71	1.75			± 0.12	1.86	1.84	1.86			± 0.12	1.99	1.97	1.97			± 0.16	
	g/dl	34.5	34.4	35.9			± 3.0	36.0	36.2	36.8			± 3.0	36.4	36.5	37.3			± 3.0	
CCMH MCHC	g/l	345	344	359			± 30	360	362	368			± 30	364	365	373			± 30	
	mmol/l	21.44	21.38	22.31			± 1.86	22.37	22.48	22.88			± 1.86	22.62	22.65	23.19			± 1.86	
IDR RDW	%	13.5	13.5	12.5			± 4.0	13.0	13.0	11.5			± 4.0	13.0	12.5	11.0			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	67	63	65			± 20	242	240	234			± 30	480	485	475			± 50	
VPM MPV	µm ³ ; fl	10.2	10.2	9.9			± 2.0	9.5	9.4	9.5			± 2.0	9.6	9.7	9.5			± 2.0	
NEUT	#	1.36	1.35	1.41			± 0.35	4.16	4.09	4.33			± 0.90	12.57	12.32	12.91			± 1.90	
	%	56.5	56.3	56.3			± 10.0	56.2	55.3	57.0			± 10.0	70.2	69.2	71.3			± 10.0	
LYMPHO	#	0.73	0.72	0.75			± 0.33	2.46	2.43	2.44			± 0.70	2.86	2.78	2.77			± 1.50	
	%	30.4	30.2	30.1			± 12.0	33.2	32.9	32.1			± 8.0	16.0	15.6	15.3			± 8.0	
MONO	#	0.08	0.09	0.10			± 0.08	0.24	0.34	0.30			± 0.24	0.75	1.00	0.78			± 0.75	
	%	3.3	3.7	3.8			± 3.3	3.3	4.6	3.9			± 3.3	4.2	5.6	4.3			± 4.2	
EOS	#	0.16	0.16	0.16			± 0.16	0.28	0.29	0.27			± 0.27	0.93	0.94	0.87			± 0.87	
	%	6.5	6.6	6.3			± 6.3	3.8	3.9	3.5			± 3.5	5.2	5.3	4.8			± 4.8	
BASO	#	0.08	0.08	0.09			± 0.08	0.26	0.24	0.27			± 0.24	0.79	0.77	0.78			± 0.77	
	%	3.3	3.2	3.5			± 3.2	3.5	3.3	3.5			± 3.3	4.4	4.3	4.3			± 4.3	

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43