


LOT PX 433
Rev 1


CONTROL

 (Exp.) 2022-03-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.3	7.3			± 1.0	17.0	17.0			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.32	2.32			± 0.16	4.50	4.50			± 0.20	5.11	5.11			± 0.25			
HB HGB	g/dl	6.7	6.7			± 0.4	13.8	13.8			± 0.5	16.7	16.7			± 0.6			
	g/l	67	67			± 4	138	138			± 5	167	167			± 6			
	mmol/l	4.16	4.16			± 0.25	8.57	8.57			± 0.31	10.37	10.37			± 0.37			
HT HCT	%	19.7	19.7			± 1.5	38.3	38.3			± 2.0	46.0	46.0			± 2.5			
	l/l	0.197	0.197			± 0.015	0.383	0.383			± 0.020	0.460	0.460			± 0.025			
VGM MCV	µm ³ fl	85	85			± 5	85	85			± 5	90	90			± 5			
TGMH MCH	pg	28.9	28.9			± 2.0	30.7	30.7			± 2.0	32.7	32.7			± 2.5			
	fmol	1.79	1.79			± 0.12	1.90	1.90			± 0.12	2.03	2.03			± 0.16			
CCMH MCHC	g/dl	34.0	34.0			± 3.0	36.1	36.1			± 3.0	36.3	36.3			± 3.0			
	g/l	340	340			± 30	361	361			± 30	363	363			± 30			
	mmol/l	21.10	21.10			± 1.86	22.40	22.40			± 1.86	22.55	22.55			± 1.86			
IDR RDW	%	13.0	13.0			± 5.0	12.5	12.5			± 5.0	12.0	12.0			± 5.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69			± 20	250	250			± 30	487	487			± 50			
VPM MPV	µm ³ fl	10.5	10.5			± 2.0	10.5	10.5			± 2.0	10.4	10.4			± 2.0			
NEUT	#	1.43	1.43			± 0.35	4.30	4.30			± 0.90	12.36	12.36			± 1.90			
	%	57.3	57.3			± 10.0	58.9	58.9			± 10.0	72.7	72.7			± 10.0			
LYMPHO	#	0.74	0.74			± 0.33	2.34	2.34			± 0.70	2.70	2.70			± 1.50			
	%	29.6	29.6			± 12.0	32.1	32.1			± 8.0	15.9	15.9			± 8.0			
MONO	#	0.05	0.05			± 0.05	0.12	0.12			± 0.12	0.41	0.41			± 0.41			
	%	1.9	1.9			± 1.9	1.7	1.7			± 1.7	2.4	2.4			± 2.4			
EOS	#	0.20	0.20			± 0.20	0.28	0.28			± 0.28	0.78	0.78			± 0.78			
	%	7.8	7.8			± 7.8	3.8	3.8			± 3.8	4.6	4.6			± 4.6			
BASO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.75	0.75			± 0.75			
	%	3.4	3.4			± 3.4	3.5	3.5			± 3.5	4.4	4.4			± 4.4			

LOT PX 433
Rev 1

CONTROL

 (Exp.) 2022-03-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.4	2.5		± 0.4	7.3	7.2	7.2		± 1.0	17.4	17.2	17.1		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.37	2.34		± 0.16	4.55	4.49	4.51		± 0.20	5.11	5.03	5.10		± 0.25			
	g/dl	6.8	6.7	6.8		± 0.4	13.9	13.7	13.7		± 0.5	16.8	16.4	16.6		± 0.6			
HB HGB	g/l	68	67	68		± 4	139	137	137		± 5	168	164	166		± 6			
	mmol/l	4.22	4.16	4.22		± 0.25	8.63	8.51	8.51		± 0.31	10.43	10.18	10.31		± 0.37			
HT HCT	%	19.4	19.0	19.0		± 1.5	38.2	37.3	37.7		± 2.0	46.0	44.8	45.4		± 2.5			
	l/l	0.194	0.190	0.190		± 0.015	0.382	0.373	0.377		± 0.020	0.460	0.448	0.454		± 0.025			
VGM MCV	µm ³ ·fl	81	80	81.0		± 5	84	83	83.5		± 5	90	89	89.0		± 5			
TGMH MCH	pg	28.5	28.3	29.1		± 2.0	30.5	30.5	30.4		± 2.0	32.9	32.6	32.5		± 2.5			
	fmol	1.77	1.76	1.80		± 0.12	1.90	1.89	1.89		± 0.12	2.04	2.02	2.02		± 0.16			
	g/dl	35.1	35.3	35.9		± 3.0	36.4	36.8	36.4		± 3.0	36.5	36.6	36.6		± 3.0			
CCMH MCHC	g/l	351	353	359		± 30	364	368	364		± 30	365	366	366		± 30			
	mmol/l	21.81	21.94	22.28		± 1.86	22.58	22.83	22.59		± 1.86	22.68	22.75	22.71		± 1.86			
IDR RDW	%	13.5	13.0	12.5		± 4.0	12.5	12.0	12.5		± 4.0	12.0	12.0	12.5		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	73		± 20	255	252	252		± 30	486	493	490		± 50			
VPM MPV	µm ³ ·fl	10.0	10.0	9.5		± 2.0	9.8	9.8	9.3		± 2.0	9.2	9.5	8.8		± 2.0			
NEUT	#	1.43	1.36	1.45		± 0.35	4.18	4.15	4.25		± 0.90	12.46	12.33	12.74		± 1.90			
	%	57.0	56.6	58.0		± 10.0	57.3	57.7	59.0		± 10.0	71.6	71.7	74.5		± 10.0			
LYMPHO	#	0.76	0.71	0.74		± 0.33	2.42	2.32	2.30		± 0.70	2.92	2.70	2.65		± 1.50			
	%	30.4	29.7	29.5		± 12.0	33.2	32.2	32.0		± 8.0	16.8	15.7	15.5		± 8.0			
MONO	#	0.05	0.07	0.06		± 0.05	0.14	0.19	0.14		± 0.14	0.44	0.58	0.34		± 0.34			
	%	2.1	2.8	2.5		± 2.1	1.9	2.7	2.0		± 1.9	2.5	3.4	2.0		± 2.0			
EOS	#	0.18	0.18	0.16		± 0.16	0.28	0.28	0.25		± 0.25	0.82	0.83	0.68		± 0.68			
	%	7.1	7.5	6.5		± 6.5	3.8	3.9	3.5		± 3.5	4.7	4.8	4.0		± 4.0			
BASO	#	0.09	0.08	0.09		± 0.08	0.28	0.25	0.25		± 0.25	0.77	0.76	0.68		± 0.68			
	%	3.4	3.4	3.5		± 3.4	3.8	3.5	3.5		± 3.5	4.4	4.4	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.45