


LOT PX 404
Rev 1


CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRP	60 60C+ ES60	80 XL80	MS60	XLR	MSCRP	60 60C+ ES60	80 XL80	MS60	XLR	MSCRP				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.6	± 0.4	7.3	7.3	7.4	7.3	7.5	± 1.0	17.4	17.1	17.4	17.1	17.7	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.34	2.30	2.35	2.30	2.35	± 0.16	4.58	4.60	4.60	4.60	4.62	± 0.20	5.15	5.19	5.20	5.19	5.22	± 0.25	
HB	HGB	g/dl	6.6	6.6	6.6	6.6	6.6	± 0.4	13.5	13.6	13.6	13.6	13.5	± 0.5	16.4	16.5	16.5	16.5	16.4	± 0.6	
		g/l	66	66	66	66	66	± 4	135	136	136	136	135	± 5	164	165	165	165	164	± 6	
		mmol/l	4.10	4.10	4.10	4.10	4.10	± 0.25	8.38	8.45	8.45	8.45	8.38	± 0.31	10.18	10.25	10.25	10.25	10.18	± 0.37	
HT	HCT	%	18.7	19.1	18.8	19.1	18.7	± 1.5	37.1	37.7	36.8	37.3	37.4	± 2.0	44.8	45.2	44.7	44.6	44.9	± 2.5	
		l/l	0.187	0.191	0.188	0.191	0.187	± 0.015	0.371	0.377	0.368	0.373	0.374	± 0.020	0.448	0.452	0.447	0.446	0.449	± 0.025	
VGM	MCV	µm ³ ; fl	80	83	80	83	79.5	± 5	81	82	80	81	81.0	± 5	87	87	86	86	86.0	± 5	
TGMH	MCH	pg	28.2	28.7	28.1	28.7	28.1	± 2.0	29.5	29.6	29.6	29.6	29.2	± 2.0	31.8	31.8	31.7	31.8	31.4	± 2.5	
		fmol	1.75	1.78	1.74	1.78	1.74	± 0.12	1.83	1.84	1.84	1.84	1.81	± 0.12	1.98	1.97	1.97	1.97	1.95	± 0.16	
CCMH	MCHC	g/dl	35.3	34.6	35.1	34.6	35.3	± 3.0	36.4	36.1	37.0	36.5	36.1	± 3.0	36.6	36.5	36.9	37.0	36.5	± 3.0	
		g/l	353	346	351	346	353	± 30	364	361	370	365	361	± 30	366	365	369	370	365	± 30	
		mmol/l	21.89	21.47	21.80	21.47	21.94	± 1.86	22.60	22.39	22.95	22.67	22.40	± 1.86	22.73	22.69	22.91	22.96	22.69	± 1.86	
IDR	RDW	%	12.8	13.0	12.7	13.0	12.3	± 4.0	12.5	13.0	12.6	13.0	12.5	± 4.0	12.6	13.0	12.3	13.0	12.8	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	72	73	72	73	± 20	250	247	253	247	243	± 30	495	495	520	495	490	± 50	
VPM	MPV	µm ³ ; fl	8.8	9.0	8.9	9.0	8.3	± 2.0	9.5	9.8	9.6	9.8	9.1	± 2.0	9.1	9.3	9.1	9.3	8.7	± 2.0	
NEUT	#	#	1.39	1.37	1.41	1.37	1.46	± 0.35	4.07	4.07	4.08	4.07	4.19	± 0.90	12.21	11.83	12.23	11.83	12.53	± 1.90	
		%	55.5	54.9	56.4	54.9	56.0	± 10.0	55.8	55.7	55.2	55.7	55.8	± 10.0	70.2	69.2	70.3	69.2	70.8	± 10.0	
LYMPHO	#	#	0.79	0.80	0.79	0.80	0.84	± 0.33	2.39	2.39	2.52	2.39	2.55	± 0.70	3.03	2.98	3.10	2.98	3.20	± 1.50	
		%	31.5	31.9	31.5	31.9	32.2	± 12.0	32.8	32.7	34.0	32.7	34.0	± 8.0	17.4	17.4	17.8	17.4	18.1	± 8.0	
MONO	#	#	0.10	0.11	0.09	0.11	0.10	± 0.09	0.30	0.31	0.26	0.31	0.27	± 0.26	0.70	0.80	0.61	0.80	0.60	± 0.60	
		%	4.0	4.3	3.4	4.3	3.7	± 3.4	4.1	4.3	3.5	4.3	3.6	± 3.5	4.0	4.7	3.5	4.7	3.4	± 3.4	
EOS	#	#	0.14	0.14	0.14	0.14	0.12	± 0.12	0.27	0.28	0.28	0.28	0.25	± 0.25	0.73	0.75	0.71	0.75	0.64	± 0.64	
		%	5.6	5.7	5.4	5.7	4.7	± 4.7	3.7	3.8	3.8	3.8	3.3	± 3.3	4.2	4.4	4.1	4.4	3.6	± 3.6	
BASO	#	#	0.09	0.08	0.08	0.08	0.09	± 0.08	0.26	0.26	0.26	0.26	0.25	± 0.25	0.73	0.74	0.75	0.74	0.73	± 0.73	
		%	3.4	3.2	3.3	3.2	3.4	± 3.2	3.6	3.5	3.5	3.5	3.3	± 3.3	4.2	4.3	4.3	4.3	4.1	± 4.1	

LOT PX 404
Rev 1

CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.7	7.7	7.7		± 1.0	18.5	18.5	18.5		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35	2.35		± 0.16	4.63	4.63	4.63		± 0.20	5.28	5.28	5.28		± 0.25			
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.5	13.5	13.5		± 0.5	16.3	16.3	16.3		± 0.6			
	g/l	67	67	67		± 4	135	135	135		± 5	163	163	163		± 6			
	mmol/l	4.16	4.16	4.16		± 0.25	8.38	8.38	8.38		± 0.31	10.12	10.12	10.12		± 0.37			
HT HCT	%	19.5	19.5	19.5		± 1.5	38.0	38.0	38.0		± 2.0	46.5	46.5	46.5		± 2.5			
	l/l	0.195	0.195	0.195		± 0.015	0.380	0.380	0.380		± 0.020	0.465	0.465	0.465		± 0.025			
VGM MCV	µm ³ ; fl	83	83	83		± 5	82	82	82		± 5	88	88	88		± 5			
TGMH MCH	pg	28.5	28.5	28.5		± 2.0	29.2	29.2	29.2		± 2.0	30.9	30.9	30.9		± 2.5			
	fmol	1.77	1.77	1.77		± 0.12	1.81	1.81	1.81		± 0.12	1.92	1.92	1.92		± 0.16			
CCMH MCHC	g/dl	34.4	34.4	34.4		± 3.0	35.6	35.6	35.6		± 3.0	35.1	35.1	35.1		± 3.0			
	g/l	344	344	344		± 30	356	356	356		± 30	351	351	351		± 30			
	mmol/l	21.33	21.33	21.33		± 1.86	22.08	22.08	22.08		± 1.86	21.79	21.79	21.79		± 1.86			
IDR RDW	%	14.8	14.8	14.8		± 4.0	16.2	16.2	16.2		± 4.0	15.4	15.4	15.4		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	252	252	252		± 30	495	495	495		± 50			
VPM MPV	µm ³ ; fl	9.1	9.1	9.1		± 2.0	9.7	9.7	9.7		± 2.0	9.1	9.1	9.1		± 2.0			
NEUT	#	1.43	1.49	1.49		± 0.35	4.24	4.37	4.37		± 0.90	12.90	13.20	13.20		± 1.90			
	%	54.9	57.3	57.3		± 10.0	55.1	56.7	56.7		± 10.0	69.6	71.0	71.0		± 10.0			
LYMPHO	#	0.82	0.75	0.75		± 0.33	2.45	2.29	2.29		± 0.70	3.15	2.90	2.90		± 1.50			
	%	31.6	28.7	28.7		± 12.0	31.8	29.7	29.7		± 8.0	17.0	15.7	15.7		± 8.0			
MONO	#	0.14	0.14	0.14		± 0.14	0.47	0.49	0.49		± 0.47	1.17	1.15	1.15		± 1.15			
	%	5.3	5.4	5.4		± 5.3	6.1	6.3	6.3		± 6.1	6.3	6.2	6.2		± 6.2			
EOS	#	0.15	0.16	0.16		± 0.15	0.31	0.33	0.33		± 0.31	0.85	0.85	0.85		± 0.85			
	%	5.7	6.1	6.1		± 5.7	4.0	4.3	4.3		± 4.0	4.6	4.6	4.6		± 4.6			
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.46	0.46	0.46		± 0.46			
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5			

1300032435-A
FRONT / RECTO
Ref: TEMP-0821 Rev.40