


LOT PX 408
Rev 1


CONTROL

 (Exp.) 2018-01-05
(YYYY - MM - DD)

		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		80 XL80	XLR				80 XL80	XLR				80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7			± 0.4	7.9	7.9			± 1.0	18.5	18.5			± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.22	2.22			± 0.16	4.42	4.42			± 0.20	5.03	5.03			± 0.25	
HB HGB	g/dl	6.4	6.4			± 0.4	12.7	12.7			± 0.5	15.9	15.9			± 0.6	
	g/l	64	64			± 4	127	127			± 5	159	159			± 6	
	mmol/l	3.97	3.97			± 0.25	7.89	7.89			± 0.31	9.87	9.87			± 0.37	
HT HCT	%	18.4	18.4			± 1.5	35.4	34.9			± 2.0	43.8	43.3			± 2.5	
	l/l	0.184	0.184			± 0.015	0.354	0.349			± 0.020	0.438	0.433			± 0.025	
VGM MCV	µm ³ ; fl	83	83			± 5	80	79			± 5	87	86			± 5	
TGMH MCH	pg	28.8	28.8			± 2.0	28.7	28.7			± 2.0	31.6	31.6			± 2.5	
	fmol	1.79	1.79			± 0.12	1.78	1.78			± 0.12	1.96	1.96			± 0.16	
CCMH MCHC	g/dl	34.7	34.7			± 3.0	35.9	36.4			± 3.0	36.3	36.8			± 3.0	
	g/l	347	347			± 30	359	364			± 30	363	368			± 30	
	mmol/l	21.57	21.57			± 1.86	22.30	22.59			± 1.86	22.56	22.83			± 1.86	
IDR RDW	%	13.5	13.5			± 4.0	13.0	13.0			± 4.0	13.0	13.0			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70			± 20	245	245			± 30	495	495			± 50	
VPM MPV	µm ³ ; fl	10.2	10.2			± 2.0	10.0	10.0			± 2.0	9.0	9.0			± 2.0	
NEUT	#	1.46	1.46			± 0.35	4.40	4.40			± 0.90	12.82	12.82			± 1.90	
	%	53.9	53.9			± 10.0	55.7	55.7			± 10.0	69.3	69.3			± 10.0	
LYMPHO	#	0.88	0.88			± 0.33	2.55	2.55			± 0.70	3.03	3.03			± 1.50	
	%	32.6	32.6			± 12.0	32.3	32.3			± 8.0	16.4	16.4			± 8.0	
MONO	#	0.14	0.14			± 0.14	0.40	0.40			± 0.40	1.00	1.00			± 1.00	
	%	5.2	5.2			± 5.2	5.0	5.0			± 5.0	5.4	5.4			± 5.4	
EOS	#	0.14	0.14			± 0.14	0.28	0.28			± 0.28	0.83	0.83			± 0.83	
	%	5.3	5.3			± 5.3	3.6	3.6			± 3.6	4.5	4.5			± 4.5	
BASO	#	0.08	0.08			± 0.08	0.27	0.27			± 0.27	0.81	0.81			± 0.81	
	%	3.0	3.0			± 3.0	3.4	3.4			± 3.4	4.4	4.4			± 4.4	

LOT PX 408
Rev 1

CONTROL

 (Exp.) 2018-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.9	7.9	7.9		± 1.0	18.8	18.5	18.8		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.31	2.28	2.28		± 0.16	4.44	4.44	4.45		± 0.20	5.03	5.04	5.01		± 0.25	
		g/dl	6.5	6.4	6.5		± 0.4	12.7	12.7	12.7		± 0.5	15.8	15.8	15.6		± 0.6	
HB	HGB	g/l	65	64	65		± 4	127	127	127		± 5	158	158	156		± 6	
		mmol/l	4.04	3.97	4.04		± 0.25	7.89	7.89	7.89		± 0.31	9.81	9.81	9.69		± 0.37	
HT	HCT	%	18.5	18.2	18.4		± 1.5	35.5	35.1	35.6		± 2.0	43.8	43.3	43.1		± 2.5	
		l/l	0.185	0.182	0.184		± 0.015	0.355	0.351	0.356		± 0.020	0.438	0.433	0.431		± 0.025	
VGM	MCV	µm ³ ; fl	80	80	80.5		± 5	80	79	80.0		± 5	87	86	86.0		± 5	
TGMH	MCH	pg	28.1	28.1	28.5		± 2.0	28.6	28.6	28.5		± 2.0	31.4	31.3	31.1		± 2.5	
		fmol	1.75	1.74	1.77		± 0.12	1.78	1.78	1.77		± 0.12	1.95	1.95	1.93		± 0.16	
		g/dl	35.2	35.1	35.4		± 3.0	35.8	36.2	35.7		± 3.0	36.1	36.5	36.2		± 3.0	
CCMH	MCHC	g/l	352	351	354		± 30	358	362	357		± 30	361	365	362		± 30	
		mmol/l	21.84	21.79	21.99		± 1.86	22.20	22.48	22.15		± 1.86	22.42	22.64	22.48		± 1.86	
IDR	RDW	%	13.5	13.5	13.5		± 4.0	13.0	12.8	12.5		± 4.0	13.0	12.2	12.2		± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	76	72		± 20	249	247	240		± 30	490	500	475		± 50	
VPM	MPV	µm ³ ; fl	10.0	10.0	9.2		± 2.0	9.6	9.6	9.1		± 2.0	8.8	8.8	8.4		± 2.0	
		#	1.42	1.37	1.46		± 0.35	4.46	4.41	4.45		± 0.90	13.20	12.88	13.33		± 1.90	
		%	54.5	52.7	56.0		± 10.0	56.5	55.8	56.3		± 10.0	70.2	69.6	70.9		± 10.0	
		#	0.84	0.85	0.81		± 0.33	2.53	2.45	2.50		± 0.70	3.03	2.83	2.80		± 1.50	
		%	32.3	32.5	31.0		± 12.0	32.0	31.0	31.6		± 8.0	16.1	15.3	14.9		± 8.0	
		#	0.13	0.15	0.14		± 0.13	0.36	0.46	0.42		± 0.36	0.92	1.17	1.13		± 0.92	
		%	4.9	5.9	5.5		± 4.9	4.6	5.8	5.3		± 4.6	4.9	6.3	6.0		± 4.9	
		#	0.14	0.15	0.10		± 0.10	0.28	0.32	0.26		± 0.26	0.85	0.83	0.75		± 0.75	
		%	5.3	5.9	4.0		± 4.0	3.6	4.0	3.3		± 3.3	4.5	4.5	4.0		± 4.0	
		#	0.08	0.08	0.09		± 0.08	0.26	0.27	0.28		± 0.26	0.81	0.80	0.79		± 0.79	
		%	3.0	3.0	3.5		± 3.0	3.3	3.4	3.5		± 3.3	4.3	4.3	4.2		± 4.2	

1300040596
FRONT / RECTO
Ref: TEMP-0821 Rev.41