


LOT PX 417
Rev 3


CONTROL

 (Exp.) 2019-07-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5				± 0.4	7.3	7.3				± 1.0	18.0	18.0				± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.29	2.29				± 0.16	4.59	4.59				± 0.20	5.19	5.19				± 0.25	
HB HGB	g/dl	6.6	6.6				± 0.4	13.9	13.9				± 0.5	16.6	16.6				± 0.6	
	g/l	66	66				± 4	139	139				± 5	166	166				± 6	
	mmol/l	4.10	4.10				± 0.25	8.63	8.63				± 0.31	10.31	10.31				± 0.37	
HT HCT	%	19.0	19.0				± 1.5	39.0	39.0				± 2.0	45.7	45.7				± 2.5	
	l/l	0.190	0.190				± 0.015	0.390	0.390				± 0.020	0.457	0.457				± 0.025	
VGM MCV	µm ³ fl	83	83				± 5	85	85				± 5	88	88				± 5	
TGMH MCH	pg	28.8	28.8				± 2.0	30.3	30.3				± 2.0	32.0	32.0				± 2.5	
	fmol	1.79	1.79				± 0.12	1.88	1.88				± 0.12	1.99	1.99				± 0.16	
CCMH MCHC	g/dl	34.7	34.7				± 3.0	35.6	35.6				± 3.0	36.3	36.3				± 3.0	
	g/l	347	347				± 30	356	356				± 30	363	363				± 30	
	mmol/l	21.56	21.56				± 1.86	22.12	22.12				± 1.86	22.57	22.57				± 1.86	
IDR RDW	%	13.0	11.5				± 4.0	13.5	12.5				± 4.0	13.0	12.0				± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67				± 20	245	245				± 30	495	495				± 50	
VPM MPV	µm ³ fl	10.0	10.0				± 2.0	10.0	10.3				± 2.0	9.6	9.9				± 2.0	
NEUT	#	1.43	1.43				± 0.35	4.02	4.02				± 0.90	12.44	12.44				± 1.90	
	%	57.1	57.1				± 10.0	55.1	55.1				± 10.0	69.1	69.1				± 10.0	
LYMPHO	#	0.75	0.75				± 0.33	2.37	2.37				± 0.70	2.79	2.79				± 1.50	
	%	30.1	30.1				± 12.0	32.4	32.4				± 8.0	15.5	15.5				± 8.0	
MONO	#	0.11	0.11				± 0.11	0.30	0.30				± 0.30	0.88	0.88				± 0.88	
	%	4.3	4.3				± 4.3	4.1	4.1				± 4.1	4.9	4.9				± 4.9	
EOS	#	0.13	0.13				± 0.13	0.35	0.35				± 0.35	1.08	1.08				± 1.08	
	%	5.2	5.2				± 5.2	4.8	4.8				± 4.8	6.0	6.0				± 6.0	
BASO	#	0.08	0.08				± 0.08	0.26	0.26				± 0.26	0.81	0.81				± 0.81	
	%	3.3	3.3				± 3.3	3.6	3.6				± 3.6	4.5	4.5				± 4.5	

LOT PX 417
Rev 3

CONTROL

 (Exp.) 2019-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.3	7.2	7.4			± 1.0	18.3	18.0	18.4			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.31	2.32			± 0.16	4.59	4.57	4.60			± 0.20	5.15	5.16	5.19			± 0.25
		g/dl	6.6	6.6	6.6			± 0.4	13.9	13.8	13.8			± 0.5	16.5	16.4	16.4			± 0.6
HB	HGB	g/l	66	66	66			± 4	139	138	138			± 5	165	164	164			± 6
		mmol/l	4.10	4.10	4.10			± 0.25	8.63	8.57	8.57			± 0.31	10.25	10.18	10.18			± 0.37
HT	HCT	%	18.8	18.5	18.6			± 1.5	38.6	37.9	38.0			± 2.0	45.3	45.4	44.9			± 2.5
		l/l	0.188	0.185	0.186			± 0.015	0.386	0.379	0.380			± 0.020	0.453	0.454	0.449			± 0.025
VGM	MCV	µm ³ ; fl	80	80	80.0			± 5	84	83	82.6			± 5	88	88	86.5			± 5
TGMH	MCH	pg	28.1	28.6	28.4			± 2.0	30.3	30.2	30.0			± 2.0	32.0	31.8	31.6			± 2.5
		fmol	1.74	1.77	1.77			± 0.12	1.88	1.88	1.86			± 0.12	1.99	1.97	1.96			± 0.16
		g/dl	35.1	35.7	35.6			± 3.0	36.1	36.4	36.3			± 3.0	36.4	36.1	36.5			± 3.0
CCMH	MCHC	g/l	351	357	356			± 30	361	364	363			± 30	364	361	365			± 30
		mmol/l	21.80	22.18	22.08			± 1.86	22.39	22.59	22.55			± 1.86	22.61	22.43	22.69			± 1.86
IDR	RDW	%	12.5	13.0	12.5			± 4.0	13.0	12.5	12.0			± 4.0	12.8	12.5	12.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	66	70			± 20	245	245	240			± 30	485	495	475			± 50
VPM	MPV	µm ³ ; fl	9.7	9.8	9.2			± 2.0	9.8	9.7	9.3			± 2.0	9.5	9.6	8.7			± 2.0
		#	1.35	1.37	1.44			± 0.35	4.09	4.02	4.13			± 0.90	12.83	12.60	12.90			± 1.90
		%	56.4	57.0	57.6			± 10.0	56.0	55.9	55.8			± 10.0	70.1	70.0	70.1			± 10.0
		#	0.73	0.71	0.76			± 0.33	2.40	2.34	2.48			± 0.70	2.76	2.65	2.94			± 1.50
		%	30.4	29.5	30.5			± 12.0	32.9	32.5	33.5			± 8.0	15.1	14.7	16.0			± 8.0
		#	0.11	0.12	0.10			± 0.10	0.30	0.35	0.30			± 0.30	0.93	0.99	0.83			± 0.83
		%	4.5	4.8	4.0			± 4.0	4.1	4.8	4.0			± 4.0	5.1	5.5	4.5			± 4.5
		#	0.13	0.13	0.12			± 0.12	0.25	0.24	0.24			± 0.24	0.95	0.97	0.92			± 0.92
		%	5.4	5.4	4.6			± 4.6	3.4	3.4	3.2			± 3.2	5.2	5.4	5.0			± 5.0
		#	0.08	0.08	0.08			± 0.08	0.26	0.24	0.26			± 0.24	0.82	0.79	0.81			± 0.79
		%	3.3	3.3	3.3			± 3.3	3.6	3.4	3.5			± 3.4	4.5	4.4	4.4			± 4.4

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43