


LOT PX 426
Rev 2


CONTROL

 (Exp.) 2021-01-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8				± 0.4	7.8	7.8				± 1.0	18.0	18.0				± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.23	2.23				± 0.16	4.52	4.52				± 0.20	5.00	5.00				± 0.25	
HB HGB	g/dl	6.3	6.3				± 0.4	13.6	13.6				± 0.5	16.2	16.2				± 0.6	
	g/l	63	63				± 4	136	136				± 5	162	162				± 6	
	mmol/l	3.91	3.91				± 0.25	8.45	8.45				± 0.31	10.06	10.06				± 0.37	
HT HCT	%	18.3	18.3				± 1.5	37.5	37.5				± 2.0	44.5	44.5				± 2.5	
	l/l	0.183	0.183				± 0.015	0.375	0.375				± 0.020	0.445	0.445				± 0.025	
VGM MCV	µm ³ ; fl	82	82				± 5	83	83				± 5	89	89				± 5	
TGMH MCH	pg	28.3	28.3				± 2.0	30.1	30.1				± 2.0	32.4	32.4				± 2.5	
	fmol	1.75	1.75				± 0.12	1.87	1.87				± 0.12	2.01	2.01				± 0.16	
CCMH MCHC	g/dl	34.5	34.5				± 3.0	36.3	36.3				± 3.0	36.4	36.4				± 3.0	
	g/l	345	345				± 30	363	363				± 30	364	364				± 30	
	mmol/l	21.40	21.40				± 1.86	22.51	22.51				± 1.86	22.61	22.61				± 1.86	
IDR RDW	%	14.0	13.0				± 4.0	13.3	12.5				± 4.0	13.0	12.0				± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66				± 20	245	245				± 30	490	490				± 50	
VPM MPV	µm ³ ; fl	9.9	9.9				± 2.0	8.9	8.9				± 2.0	8.9	8.9				± 2.0	
NEUT	#	1.47	1.47				± 0.35	4.05	4.05				± 0.90	12.02	12.02				± 1.90	
	%	52.4	52.4				± 10.0	51.9	51.9				± 10.0	66.8	66.8				± 10.0	
LYMPHO	#	0.80	0.80				± 0.33	2.35	2.35				± 0.70	2.84	2.84				± 1.50	
	%	28.6	28.6				± 12.0	30.1	30.1				± 8.0	15.8	15.8				± 8.0	
MONO	#	0.23	0.23				± 0.23	0.73	0.73				± 0.73	1.46	1.46				± 1.46	
	%	8.1	8.1				± 8.1	9.3	9.3				± 9.3	8.1	8.1				± 8.1	
EOS	#	0.21	0.21				± 0.21	0.38	0.38				± 0.38	0.86	0.86				± 0.86	
	%	7.5	7.5				± 7.5	4.9	4.9				± 4.9	4.8	4.8				± 4.8	
BASO	#	0.10	0.10				± 0.10	0.30	0.30				± 0.30	0.81	0.81				± 0.81	
	%	3.4	3.4				± 3.4	3.8	3.8				± 3.8	4.5	4.5				± 4.5	

LOT PX 426
Rev 2

CONTROL

 (Exp.) 2021-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			60 60C+ ES60	MS60		MSCR	
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.7			± 0.4	7.8	7.9	7.8			± 1.0	18.5	18.6	18.5			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.28	2.26	2.26			± 0.16	4.50	4.49	4.50			± 0.20	5.00	4.96	5.00			± 0.25
		g/dl	6.4	6.3	6.4			± 0.4	13.6	13.4	13.5			± 0.5	16.3	16.0	16.2			± 0.6
HB	HGB	g/l	64	63	64			± 4	136	134	135			± 5	163	160	162			± 6
		mmol/l	3.97	3.91	3.97			± 0.25	8.45	8.32	8.38			± 0.31	10.12	9.94	10.06			± 0.37
HT	HCT	%	18.2	17.9	17.7			± 1.5	37.4	37.3	37.4			± 2.0	45.0	44.1	44.0			± 2.5
		l/l	0.182	0.179	0.177			± 0.015	0.374	0.373	0.374			± 0.020	0.450	0.441	0.440			± 0.025
VGM	MCV	µm ³ ; fl	80	79	78.5			± 5	83	83	83.0			± 5	90	89	88.0			± 5
TGMH	MCH	pg	28.1	27.9	28.3			± 2.0	30.2	29.8	30.0			± 2.0	32.6	32.3	32.4			± 2.5
		fmol	1.74	1.73	1.76			± 0.12	1.88	1.85	1.86			± 0.12	2.02	2.00	2.01			± 0.16
		g/dl	35.1	35.3	36.1			± 3.0	36.4	36.0	36.1			± 3.0	36.2	36.2	36.8			± 3.0
CCMH	MCHC	g/l	351	353	361			± 30	364	360	361			± 30	362	362	368			± 30
		mmol/l	21.79	21.91	22.40			± 1.86	22.61	22.33	22.45			± 1.86	22.49	22.51	22.86			± 1.86
IDR	RDW	%	13.5	13.0	13.2			± 4.0	12.5	12.0	12.0			± 4.0	12.0	12.0	11.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	67	65			± 20	247	247	244			± 30	487	490	486			± 50
VPM	MPV	µm ³ ; fl	9.7	9.9	9.2			± 2.0	8.6	8.6	8.3			± 2.0	8.6	8.6	8.3			± 2.0
		#	1.48	1.44	1.50			± 0.35	4.17	4.01	4.35			± 0.90	12.56	12.03	12.67			± 1.90
		%	52.7	51.4	55.5			± 10.0	53.5	50.7	55.8			± 10.0	67.9	64.7	68.5			± 10.0
		#	0.79	0.80	0.75			± 0.33	2.31	2.31	2.18			± 0.70	2.87	2.85	2.68			± 1.50
		%	28.3	28.7	27.7			± 12.0	29.6	29.3	28.0			± 8.0	15.5	15.3	14.5			± 8.0
		#	0.22	0.27	0.18			± 0.18	0.65	0.92	0.62			± 0.62	1.33	1.97	1.41			± 1.33
		%	7.7	9.5	6.5			± 6.5	8.3	11.6	8.0			± 8.0	7.2	10.6	7.6			± 7.2
		#	0.22	0.20	0.18			± 0.18	0.37	0.36	0.35			± 0.35	0.89	0.89	0.93			± 0.89
		%	7.9	7.0	6.8			± 6.8	4.8	4.6	4.5			± 4.5	4.8	4.8	5.0			± 4.8
		#	0.10	0.10	0.09			± 0.09	0.30	0.30	0.29			± 0.29	0.85	0.86	0.81			± 0.81
		%	3.4	3.4	3.5			± 3.4	3.8	3.8	3.7			± 3.7	4.6	4.6	4.4			± 4.4

Ref: TEMP-0821 Rev.44 FRONT / RECTO 1300040596