

**LOT** PX 409  
Rev 1

**CONTROL**

 (Exp.) **2018-03-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE		
		CONTROL				L	CONTROL				N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA						
		80 XL80	XLR			80 XL80	XLR			80 XL80	XLR			80 XL80	XLR			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.5</b>	<b>2.5</b>			± 0.4	<b>7.4</b>	<b>7.4</b>			± 1.0	<b>17.7</b>	<b>17.7</b>			± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.27</b>	<b>2.27</b>			± 0.16	<b>4.55</b>	<b>4.55</b>			± 0.20	<b>5.09</b>	<b>5.09</b>			± 0.25		
HB HGB	g/dl	<b>6.7</b>	<b>6.7</b>			± 0.4	<b>13.4</b>	<b>13.4</b>			± 0.5	<b>16.4</b>	<b>16.4</b>			± 0.6		
	g/l	<b>67</b>	<b>67</b>			± 4	<b>134</b>	<b>134</b>			± 5	<b>164</b>	<b>164</b>			± 6		
HT HCT	mmol/l	<b>4.16</b>	<b>4.16</b>			± 0.25	<b>8.32</b>	<b>8.32</b>			± 0.31	<b>10.18</b>	<b>10.18</b>			± 0.37		
	%	<b>18.8</b>	<b>18.8</b>			± 1.5	<b>36.4</b>	<b>35.9</b>			± 2.0	<b>44.3</b>	<b>43.8</b>			± 2.5		
VGM MCV	l/l	<b>0.188</b>	<b>0.188</b>			± 0.015	<b>0.364</b>	<b>0.359</b>			± 0.020	<b>0.443</b>	<b>0.438</b>			± 0.025		
	µm <sup>3</sup> ; fl	<b>83</b>	<b>83</b>			± 5	<b>80</b>	<b>79</b>			± 5	<b>87</b>	<b>86</b>			± 5		
TGMH MCH	pg	<b>29.5</b>	<b>29.5</b>			± 2.0	<b>29.5</b>	<b>29.5</b>			± 2.0	<b>32.2</b>	<b>32.2</b>			± 2.5		
	fmol	<b>1.83</b>	<b>1.83</b>			± 0.12	<b>1.83</b>	<b>1.83</b>			± 0.12	<b>2.00</b>	<b>2.00</b>			± 0.16		
CCMH MCHC	g/dl	<b>35.6</b>	<b>35.6</b>			± 3.0	<b>36.8</b>	<b>37.3</b>			± 3.0	<b>37.0</b>	<b>37.5</b>			± 3.0		
	g/l	<b>356</b>	<b>356</b>			± 30	<b>368</b>	<b>373</b>			± 30	<b>370</b>	<b>375</b>			± 30		
	mmol/l	<b>22.08</b>	<b>22.08</b>			± 1.86	<b>22.86</b>	<b>23.15</b>			± 1.86	<b>23.00</b>	<b>23.27</b>			± 1.86		
IDR RDW	%	<b>13.0</b>	<b>12.5</b>			± 4.0	<b>13.0</b>	<b>12.5</b>			± 4.0	<b>13.0</b>	<b>12.0</b>			± 4.0		
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>65</b>	<b>65</b>			± 20	<b>250</b>	<b>246</b>			± 30	<b>500</b>	<b>495</b>			± 50		
VPM MPV	µm <sup>3</sup> ; fl	<b>9.4</b>	<b>9.4</b>			± 2.0	<b>9.4</b>	<b>9.4</b>			± 2.0	<b>8.7</b>	<b>8.7</b>			± 2.0		
NEUT	#	<b>1.41</b>	<b>1.41</b>			± 0.35	<b>4.11</b>	<b>4.11</b>			± 0.90	<b>12.12</b>	<b>12.12</b>			± 1.90		
	%	<b>56.2</b>	<b>56.2</b>			± 10.0	<b>55.5</b>	<b>55.5</b>			± 10.0	<b>68.5</b>	<b>68.5</b>			± 10.0		
LYMPHO	#	<b>0.70</b>	<b>0.70</b>			± 0.33	<b>2.23</b>	<b>2.23</b>			± 0.70	<b>2.92</b>	<b>2.92</b>			± 1.50		
	%	<b>27.8</b>	<b>27.8</b>			± 12.0	<b>30.1</b>	<b>30.1</b>			± 8.0	<b>16.5</b>	<b>16.5</b>			± 8.0		
MONO	#	<b>0.17</b>	<b>0.17</b>			± 0.17	<b>0.52</b>	<b>0.52</b>			± 0.52	<b>1.03</b>	<b>1.03</b>			± 1.03		
	%	<b>6.6</b>	<b>6.6</b>			± 6.6	<b>7.0</b>	<b>7.0</b>			± 7.0	<b>5.8</b>	<b>5.8</b>			± 5.8		
EOS	#	<b>0.15</b>	<b>0.15</b>			± 0.15	<b>0.28</b>	<b>0.28</b>			± 0.28	<b>0.85</b>	<b>0.85</b>			± 0.85		
	%	<b>6.1</b>	<b>6.1</b>			± 6.1	<b>3.8</b>	<b>3.8</b>			± 3.8	<b>4.8</b>	<b>4.8</b>			± 4.8		
BASO	#	<b>0.08</b>	<b>0.08</b>			± 0.08	<b>0.27</b>	<b>0.27</b>			± 0.27	<b>0.78</b>	<b>0.78</b>			± 0.78		
	%	<b>3.3</b>	<b>3.3</b>			± 3.3	<b>3.6</b>	<b>3.6</b>			± 3.6	<b>4.4</b>	<b>4.4</b>			± 4.4		

# ABX Difftrol



**LOT** PX 409  
Rev 1

**CONTROL**

(Exp.) 2018-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.6		± 0.4	7.5	7.4	7.5		± 1.0	18.3	17.9	18.0		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.33	2.34		± 0.16	4.54	4.58	4.56		± 0.20	5.07	5.12	5.10		± 0.25	
HB	HGB	g/dl	6.7	6.7	6.7		± 0.4	13.4	13.4	13.5		± 0.5	16.3	16.4	16.3		± 0.6	
		g/l	67	67	67		± 4	134	134	135		± 5	163	164	163		± 6	
		mmol/l	4.16	4.16	4.16		± 0.25	8.32	8.32	8.38		± 0.31	10.12	10.18	10.12		± 0.37	
HT	HCT	%	18.7	18.6	18.7		± 1.5	36.3	36.2	36.0		± 2.0	44.1	44.0	43.6		± 2.5	
		l/l	0.187	0.186	0.187		± 0.015	0.363	0.362	0.360		± 0.020	0.441	0.440	0.436		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	80	80	80.0		± 5	80	79	79.0		± 5	87	86	85.5		± 5	
TGMH	MCH	pg	28.6	28.8	28.6		± 2.0	29.5	29.3	29.6		± 2.0	32.1	32.0	32.0		± 2.5	
		fmol	1.78	1.79	1.78		± 0.12	1.83	1.82	1.84		± 0.12	2.00	1.99	1.98		± 0.16	
CCMH	MCHC	g/dl	35.8	35.9	35.8		± 3.0	36.9	37.0	37.5		± 3.0	37.0	37.2	37.4		± 3.0	
		g/l	358	359	358		± 30	369	370	375		± 30	370	372	374		± 30	
		mmol/l	22.23	22.32	22.23		± 1.86	22.91	23.00	23.27		± 1.86	22.95	23.13	23.21		± 1.86	
IDR	RDW	%	13.5	13.0	12.5		± 4.0	13.7	13.0	12.5		± 4.0	13.0	12.0	11.8		± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68	68	70		± 20	255	255	250		± 30	500	510	495		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.2	9.4	8.7		± 2.0	9.2	9.4	8.8		± 2.0	8.6	8.7	8.2		± 2.0	
NEUT		#	1.39	1.37	1.46		± 0.35	4.19	4.08	4.24		± 0.90	12.74	12.35	12.71		± 1.90	
		%	55.7	54.7	56.0		± 10.0	55.9	55.2	56.5		± 10.0	69.6	69.0	70.6		± 10.0	
LYMPHO		#	0.72	0.71	0.77		± 0.33	2.29	2.21	2.29		± 0.70	2.91	2.79	2.88		± 1.50	
		%	28.7	28.5	29.5		± 12.0	30.5	29.9	30.5		± 8.0	15.9	15.6	16.0		± 8.0	
MONO		#	0.16	0.19	0.16		± 0.16	0.48	0.56	0.45		± 0.45	1.02	1.11	0.85		± 0.85	
		%	6.3	7.5	6.0		± 6.0	6.4	7.6	6.0		± 6.0	5.6	6.2	4.7		± 4.7	
EOS		#	0.15	0.15	0.13		± 0.13	0.27	0.27	0.26		± 0.26	0.88	0.88	0.81		± 0.81	
		%	6.0	6.0	5.0		± 5.0	3.6	3.7	3.5		± 3.5	4.8	4.9	4.5		± 4.5	
BASO		#	0.08	0.08	0.09		± 0.08	0.27	0.27	0.26		± 0.26	0.75	0.77	0.76		± 0.75	
		%	3.3	3.3	3.5		± 3.3	3.6	3.6	3.5		± 3.5	4.1	4.3	4.2		± 4.1	

Ref: TEMP-0821 Rev.42 FRONT / RECTO 1300040596