

**HORIBA ABX SAS**  
 Parc Euromédecine  
 34184 Montpellier cedex 4 - France  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** PX 112  
 Rev 1

**CONTROL**

(Exp.) 2013-01-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio													TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H		
			PENTRA				PENTRA				PENTRA					
			60 60C+ ES60	80 XL80	MS60		60 60C+ ES60	80 XL80	MS60		60 60C+ ES60	80 XL80	MS60		TOLERANCES TOLERANCE	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4		± 0.4	7.4	7.3	7.2		± 1.0	17.6	17.5	17.2		± 2.2
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.33	2.37		± 0.12	4.63	4.62	4.65		± 0.15	5.16	5.18	5.18		± 0.20
HB HGB	g/dl	6.5	6.5	6.5		± 0.4	13.6	13.6	13.6		± 0.5	16.3	16.3	16.3		± 0.6
	g/l	65	65	65		± 4	136	136	136		± 5	163	163	163		± 6
	mmol/l	4.04	4.04	4.04		± 0.25	8.45	8.45	8.45		± 0.31	10.12	10.12	10.12		± 0.37
HT HCT	%	18.5	19.1	18.2		± 1.5	37.5	37.9	36.7		± 2.0	45.4	45.6	44.0		± 2.5
	l/l	0.185	0.191	0.182		± 0.015	0.375	0.379	0.367		± 0.020	0.454	0.456	0.440		± 0.025
VGM MCV	µm <sup>3</sup> ; fl	78	82	77		± 5.0	81	82	79		± 5.0	88	88	85		± 5.0
TGMH MCH	pg	27.4	27.9	27.4		± 2.0	29.4	29.4	29.2		± 2.0	31.6	31.5	31.5		± 2.5
	fmol	1.70	1.73	1.70		± 0.12	1.82	1.83	1.82		± 0.12	1.96	1.95	1.95		± 0.16
CCMH MCHC	g/dl	35.2	34.0	35.6		± 3.0	36.3	35.9	37.0		± 3.0	35.9	35.8	37.0		± 3.0
	g/l	352	340	356		± 30	363	359	370		± 30	359	358	370		± 30
	mmol/l	21.84	21.13	22.12		± 1.86	22.52	22.29	22.99		± 1.86	22.29	22.21	22.99		± 1.86
IDR RDW	%	12.8	13.2	13.1		± 4.0	12.4	12.5	12.3		± 4.0	12.3	12.7	12.2		± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	65	68	67		± 20	250	256	252		± 30	500	510	510		± 50
VPM MPV	µm <sup>3</sup> ; fl	9.3	9.6	9.5		± 2.0	9.1	9.3	9.1		± 2.0	9.1	9.2	9.0		± 2.0
NEUT	#	1.33	1.33	1.31		± 0.30	4.03	3.98	3.85		± 0.80	12.23	12.08	11.87		± 1.90
	%	55.5	55.5	54.5		± 10.0	54.5	54.5	53.5		± 10.0	69.5	69.0	69.0		± 10.0
LYMPHO	#	0.78	0.77	0.79		± 0.20	2.44	2.41	2.48		± 0.70	2.99	2.98	3.01		± 1.50
	%	32.5	32.0	33.0		± 8.0	33.0	33.0	34.5		± 8.0	17.0	17.0	17.5		± 8.0
MONO	#	0.08	0.08	0.08		± 0.08	0.33	0.33	0.33		± 0.33	0.70	0.70	0.70		± 0.70
	%	3.5	3.5	3.5		± 3.5	4.5	4.5	4.5		± 4.5	4.0	4.0	4.0		± 4.0
EOS	#	0.14	0.14	0.14		± 0.14	0.33	0.33	0.31		± 0.31	0.88	0.96	0.86		± 0.73
	%	5.5	6.0	5.5		± 5.5	4.5	4.5	4.0		± 4.0	5.0	5.5	5.0		± 4.0
BASO	#	0.08	0.08	0.08		± 0.08	0.26	0.26	0.25		± 0.23	0.79	0.79	0.77		± 0.55
	%	3.0	3.0	3.5		± 3.0	3.5	3.5	3.5		± 3.0	4.5	4.5	4.5		± 3.0

Ref: TEMP-0821 Rev.18 BACK / VERSO 9930061-A

**HORIBA ABX SAS**  
 Parc Euromédecine  
 34184 Montpellier cedex 4 - France  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** PX 112  
 Rev 1

**CONTROL**

(Exp.) 2013-01-05  
 (YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX 120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX 120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX 120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.7	7.7	7.7		± 1.0	18.2	18.2	18.2		± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.41	2.41		± 0.12	4.68	4.68	4.68		± 0.15	5.31	5.31	5.31		± 0.20			
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	13.6	13.6	13.6		± 0.5	16.3	16.3	16.3		± 0.6			
	g/l	66	66	66		± 4	136	136	136		± 5	163	163	163		± 6			
	mmol/l	4.10	4.10	4.10		± 0.25	8.45	8.45	8.45		± 0.31	10.12	10.12	10.12		± 0.37			
HT HCT	%	19.5	19.5	19.5		± 1.5	38.4	38.4	38.4		± 2.0	46.7	46.7	46.7		± 2.5			
	l/l	0.195	0.195	0.195		± 0.015	0.384	0.384	0.384		± 0.020	0.467	0.467	0.467		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	81	81	81		± 5.0	82	82	82		± 5.0	88	88	88		± 5.0			
TGMH MCH	pg	27.4	27.4	27.4		± 2.0	29.1	29.1	29.1		± 2.0	30.7	30.7	30.7		± 2.5			
	fmol	1.70	1.70	1.70		± 0.12	1.80	1.80	1.80		± 0.12	1.91	1.91	1.91		± 0.16			
CCMH MCHC	g/dl	33.8	33.8	33.8		± 3.0	35.4	35.4	35.4		± 3.0	34.9	34.9	34.9		± 3.0			
	g/l	338	338	338		± 30	354	354	354		± 30	349	349	349		± 30			
	mmol/l	21.00	21.00	21.00		± 1.86	22.01	22.01	22.01		± 1.86	21.66	21.66	21.66		± 1.86			
IDR RDW	%	15.7	15.7	15.7		± 4.0	16.2	16.2	16.2		± 4.0	15.0	15.0	15.0		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	255	255	255		± 30	510	510	510		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.5	9.5	9.5		± 2.0	9.1	9.1	9.1		± 2.0	9.0	9.0	9.0		± 2.0			
NEUT	#	1.38	1.43	1.43		± 0.30	4.24	4.31	4.31		± 0.80	12.80	12.90	12.90		± 1.90			
	%	55.0	57.0	57.0		± 10.0	55.0	56.0	56.0		± 10.0	70.0	70.5	70.5		± 10.0			
LYMPHO	#	0.81	0.75	0.75		± 0.20	2.54	2.39	2.39		± 0.70	3.09	2.82	2.82		± 1.50			
	%	32.5	30.0	30.0		± 8.0	33.0	31.0	31.0		± 8.0	17.0	15.5	15.5		± 8.0			
MONO	#	0.10	0.10	0.10		± 0.08	0.39	0.46	0.46		± 0.33	1.09	1.18	1.18		± 0.70			
	%	4.0	4.0	4.0		± 3.5	5.0	6.0	6.0		± 4.5	6.0	6.5	6.5		± 4.0			
EOS	#	0.14	0.14	0.14		± 0.14	0.31	0.31	0.31		± 0.31	0.73	0.82	0.82		± 0.73			
	%	5.5	5.5	5.5		± 5.5	4.0	4.0	4.0		± 4.0	4.0	4.5	4.5		± 4.0			
BASO	#	0.08	0.09	0.09		± 0.08	0.23	0.23	0.23		± 0.23	0.55	0.55	0.55		± 0.55			
	%	3.0	3.5	3.5		± 3.0	3.0	3.0	3.0		± 3.0	3.0	3.0	3.0		± 3.0			

9930066-A  
 BACK / VERSO  
 Ref: TEMP-0821 Rev.18