

ABX Difftrol



LOT PX 072
Rev 1

CONTROL

(Exp.) 2012-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			60 60C+ ES60		80 XL80	MS60	PENTRA			60 60C+ ES60	80 XL80	MS60	PENTRA				
			60 60C+ ES60	80 XL80	MS60					60 60C+ ES60	80 XL80					MS60		60 60C+ ES60	80 XL80	MS60
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.3	7.3	7.4			± 1.0	17.5	17.2	17.3			± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38	2.35	2.40			± 0.12	4.54	4.55	4.57			± 0.15	5.08	5.12	5.10			± 0.20
HB	HGB	g/dl	6.6	6.6	6.7			± 0.4	13.5	13.5	13.6			± 0.5	16.4	16.4	16.4			± 0.6
		g/l	66	66	67			± 4	135	135	136			± 5	164	164	164			± 6
		mmol/l	4.10	4.10	4.19			± 0.25	8.38	8.38	8.45			± 0.31	10.18	10.18	10.18			± 0.37
HT	HCT	%	18.8	19.5	18.7			± 1.5	37.2	38.2	37.0			± 2.0	45.2	45.6	43.9			± 2.5
		l/l	0.188	0.195	0.187			± 0.015	0.372	0.382	0.370			± 0.020	0.452	0.456	0.439			± 0.025
VGM	MCV	µm ³ ; fl	79	83	78			± 4.0	82	84	81			± 4.0	89	89	86			± 4.0
TGMH	MCH	pg	27.7	28.1	28.1			± 2.0	29.7	29.7	29.8			± 2.0	32.3	32.0	32.2			± 2.5
		fmol	1.72	1.74	1.74			± 0.12	1.85	1.84	1.85			± 0.12	2.00	1.99	2.00			± 0.16
CCMH	MCHC	g/dl	35.1	33.8	36.0			± 3.0	36.3	35.3	36.7			± 3.0	36.3	36.0	37.4			± 3.0
		g/l	351	338	360			± 30	363	353	367			± 30	363	360	374			± 30
		mmol/l	21.80	21.01	22.36			± 1.86	22.52	21.93	22.82			± 1.86	22.53	22.35	23.22			± 1.86
IDR	RDW	%	13.1	13.4	12.6			± 4.0	12.6	13.0	12.4			± 4.0	12.5	13.1	12.1			± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	70			± 20	255	256	260			± 30	500	505	510			± 50
VPM	MPV	µm ³ ; fl	9.0	9.1	9.2			± 2.0	8.3	8.4	8.3			± 2.0	8.3	8.4	8.3			± 2.0
NEUT		#	1.36	1.34	1.38			± 0.30	3.96	3.91	3.97			± 0.80	11.69	11.09	11.68			± 1.90
		%	56.5	56.0	55.0			± 10.0	54.2	53.5	53.7			± 10.0	66.8	64.5	67.5			± 10.0
LYMPHO		#	0.77	0.76	0.84			± 0.30	2.48	2.52	2.59			± 0.70	3.15	3.27	3.20			± 1.50
		%	32.0	31.5	33.5			± 8.0	34.0	34.5	35.0			± 8.0	18.0	19.0	18.5			± 8.0
MONO		#	0.10	0.10	0.10			± 0.10	0.30	0.30	0.30			± 0.30	0.88	0.88	0.88			± 0.88
		%	4.0	4.0	4.0			± 4.0	4.0	4.0	4.0			± 4.0	5.0	5.0	5.0			± 5.0
EOS		#	0.13	0.13	0.13			± 0.13	0.33	0.33	0.31			± 0.31	1.05	1.20	0.82			± 0.82
		%	5.0	5.5	5.0			± 5.0	4.5	4.5	4.0			± 4.0	6.0	7.0	4.5			± 4.5
BASO		#	0.06	0.07	0.06			± 0.06	0.24	0.26	0.24			± 0.20	0.74	0.77	0.78			± 0.46
		%	2.5	3.0	2.5			± 2.5	3.3	3.5	3.3			± 2.5	4.2	4.5	4.5			± 2.5

Ref: TEMP-0821 Rev.18 BACK / VERSO 9930061-A

LOT PX 072
Rev 1

CONTROL

(Exp.) 2012-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE			
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	7.8	7.8	7.8			± 1.0	18.2	18.2	18.2			± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.40	2.40	2.40			± 0.12	4.60	4.60	4.60			± 0.15	5.20	5.20	5.20			± 0.20
HB	HGB	g/dl	6.7	6.7	6.7			± 0.4	13.5	13.5	13.5			± 0.5	16.2	16.2	16.2			± 0.6
		g/l	67	67	67			± 4	135	135	135			± 5	162	162	162			± 6
		mmol/l	4.16	4.16	4.16			± 0.25	8.38	8.38	8.38			± 0.31	10.06	10.06	10.06			± 0.37
HT	HCT	%	19.4	19.4	19.4			± 1.5	38.2	38.2	38.2			± 2.0	45.8	45.8	45.8			± 2.5
		l/l	0.194	0.194	0.194			± 0.015	0.382	0.382	0.382			± 0.020	0.458	0.458	0.458			± 0.025
VGM	MCV	µm ³ ; fl	81	81	81			± 4.0	83	83	83			± 4.0	88	88	88			± 4.0
TGMH	MCH	pg	27.9	27.9	27.9			± 2.0	29.3	29.3	29.3			± 2.0	31.2	31.2	31.2			± 2.5
		fmol	1.73	1.73	1.73			± 0.12	1.82	1.82	1.82			± 0.12	1.93	1.93	1.93			± 0.16
CCMH	MCHC	g/dl	34.5	34.5	34.5			± 3.0	35.4	35.4	35.4			± 3.0	35.4	35.4	35.4			± 3.0
		g/l	345	345	345			± 30	354	354	354			± 30	354	354	354			± 30
		mmol/l	21.40	21.40	21.40			± 1.86	21.96	21.96	21.96			± 1.86	21.98	21.98	21.98			± 1.86
IDR	RDW	%	15.4	15.4	15.4			± 4.0	15.7	15.7	15.7			± 4.0	15.2	15.2	15.2			± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70			± 20	255	255	255			± 30	496	496	496			± 50
VPM	MPV	µm ³ ; fl	9.2	9.2	9.2			± 2.0	8.4	8.4	8.4			± 2.0	8.4	8.4	8.4			± 2.0
NEUT		#	1.41	1.40	1.40			± 0.30	4.25	4.37	4.37			± 0.80	12.40	12.60	12.60			± 1.90
		%	56.5	56.0	56.0			± 10.0	54.5	56.0	56.0			± 10.0	68.0	69.0	69.0			± 10.0
LYMPHO		#	0.80	0.78	0.78			± 0.30	2.65	2.46	2.46			± 0.70	3.55	3.09	3.09			± 1.50
		%	32.0	31.0	31.0			± 8.0	34.0	31.5	31.5			± 8.0	19.5	17.0	17.0			± 8.0
MONO		#	0.10	0.14	0.14			± 0.10	0.39	0.43	0.43			± 0.30	1.00	1.27	1.27			± 0.88
		%	4.0	5.5	5.5			± 4.0	5.0	5.5	5.5			± 4.0	5.5	7.0	7.0			± 5.0
EOS		#	0.13	0.13	0.13			± 0.13	0.31	0.31	0.31			± 0.31	0.82	0.82	0.82			± 0.82
		%	5.0	5.0	5.0			± 5.0	4.0	4.0	4.0			± 4.0	4.5	4.5	4.5			± 4.5
BASO		#	0.06	0.06	0.06			± 0.06	0.20	0.23	0.23			± 0.20	0.46	0.46	0.46			± 0.46
		%	2.5	2.5	2.5			± 2.5	2.5	3.0	3.0			± 2.5	2.5	2.5	2.5			± 2.5