

# ABX Difftrol



**LOT** PX 052  
Rev 1

**CONTROL**

(Exp.) 2012-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio													TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H		
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE		
			60 60C+ ES60	80 XL80	MS60		60 60C+ ES60	80 XL80	MS60		60 60C+ ES60	80 XL80	MS60			
GB WBC	$10^9/mm^3 ; 10^9/l$	2.3	2.3	2.3		± 0.4	7.3	7.3	7.3		± 1.0	17.8	17.6	17.5		± 2.2
GR RBC	$10^9/mm^3 ; 10^{12}/l$	2.38	2.33	2.33		± 0.12	4.48	4.48	4.45		± 0.15	5.06	5.06	5.05		± 0.20
HB HGB	g/dl	6.8	6.8	6.7		± 0.4	13.4	13.5	13.4		± 0.5	16.2	16.2	16.2		± 0.6
	g/l	68	68	67		± 4	134	135	134		± 5	162	162	162		± 6
	mmol/l	4.22	4.22	4.16		± 0.25	8.32	8.38	8.32		± 0.31	10.06	10.06	10.06		± 0.37
HT HCT	%	19.0	19.6	18.4		± 1.5	36.7	37.2	35.6		± 2.0	44.5	44.5	43.4		± 2.5
	l/l	0.190	0.196	0.184		± 0.015	0.367	0.372	0.356		± 0.020	0.445	0.445	0.434		± 0.025
VGM MCV	$\mu m^3/fl$	80	84	79		± 4.0	82	83	80		± 4.0	88	88	86		± 4.0
TGMH MCH	pg	28.6	29.2	28.8		± 2.0	29.9	30.1	30.1		± 2.0	32.0	32.0	32.1		± 2.5
	fmol	1.77	1.81	1.79		± 0.12	1.86	1.87	1.87		± 0.12	1.99	1.99	1.99		± 0.16
CCMH MCHC	g/dl	35.7	34.7	36.4		± 3.0	36.5	36.3	37.6		± 3.0	36.4	36.4	37.3		± 3.0
	g/l	357	347	364		± 30	365	363	376		± 30	364	364	373		± 30
	mmol/l	22.18	21.58	22.60		± 1.86	22.65	22.55	23.37		± 1.86	22.59	22.59	23.16		± 1.86
IDR RDW	%	13.5	14.0	12.6		± 4.0	12.2	13.5	12.0		± 4.0	12.0	13.7	12.0		± 4.0
PLAQ. PLTS	$10^9/mm^3 ; 10^9/l$	70	70	68		± 20	252	255	245		± 30	500	505	500		± 50
VPM MPV	$\mu m^3/fl$	8.7	8.9	8.9		± 2.0	8.2	8.3	8.2		± 2.0	8.2	8.3	8.2		± 2.0
NEUT	#	1.27	1.35	1.27		± 0.30	3.96	3.96	3.94		± 0.80	12.32	12.06	12.11		± 1.80
	%	55.0	58.5	55.3		± 10.0	54.3	54.2	54.0		± 10.0	69.2	68.5	69.2		± 10.0
LYMPHO	#	0.72	0.67	0.72		± 0.20	2.52	2.50	2.52		± 0.60	3.03	2.90	3.03		± 1.50
	%	31.5	29.0	31.5		± 8.0	34.5	34.2	34.5		± 8.0	17.0	16.5	17.3		± 8.0
MONO	#	0.10	0.08	0.10		± 0.08	0.31	0.22	0.31		± 0.22	0.85	0.62	0.80		± 0.62
	%	4.5	3.5	4.5		± 3.5	4.2	3.0	4.2		± 3.0	4.8	3.5	4.5		± 3.5
EOS	#	0.13	0.13	0.13		± 0.12	0.26	0.35	0.28		± 0.26	0.80	1.23	0.80		± 0.80
	%	5.8	5.8	5.5		± 5.0	3.5	4.8	3.8		± 3.5	4.5	7.0	4.5		± 4.5
BASO	#	0.07	0.07	0.07		± 0.07	0.26	0.28	0.26		± 0.26	0.80	0.80	0.80		± 0.80
	%	3.2	3.2	3.2		± 3.2	3.5	3.8	3.5		± 3.5	4.5	4.5	4.5		± 4.0

9930061-A  
BACK / VERSO  
Ref. TEMP-0821 Rev.17

# ABX Difftrol



**LOT**

**PX 052**

**CONTROL**

(Exp.) **2012-07-05**  
(YYYY - MM - DD)

Rev 1

		<b>ABX Lysebio</b>														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4		± 0.4	7.6	7.6	7.6		± 1.0	18.7	18.7	18.7		± 2.2
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.38	2.38		± 0.12	4.52	4.52	4.52		± 0.15	5.18	5.18	5.18		± 0.20
HB HGB	g/dl	6.8	6.8	6.8		± 0.4	13.4	13.4	13.4		± 0.5	16.2	16.2	16.2		± 0.6
	g/l	68	68	68		± 4	134	134	134		± 5	162	162	162		± 6
	mmol/l	4.22	4.22	4.22		± 0.25	8.32	8.32	8.32		± 0.31	10.06	10.06	10.06		± 0.37
HT HCT	%	20.0	20.0	20.0		± 1.5	38.0	38.0	38.0		± 2.0	46.1	46.1	46.1		± 2.5
	l/l	0.200	0.200	0.200		± 0.015	0.380	0.380	0.380		± 0.020	0.461	0.461	0.461		± 0.025
VGM MCV	µm <sup>3</sup> ; fl	84	84	84		± 4.0	84	84	84		± 4.0	89	89	89		± 4.0
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	29.6	29.6	29.6		± 2.0	31.3	31.3	31.3		± 2.5
	fmol	1.77	1.77	1.77		± 0.12	1.84	1.84	1.84		± 0.12	1.94	1.94	1.94		± 0.16
CCMH MCHC	g/dl	34.0	34.0	34.0		± 3.0	35.3	35.3	35.3		± 3.0	35.1	35.1	35.1		± 3.0
	g/l	340	340	340		± 30	353	353	353		± 30	351	351	351		± 30
	mmol/l	21.12	21.12	21.12		± 1.86	21.92	21.92	21.92		± 1.86	21.82	21.82	21.82		± 1.86
IDR RDW	%	15.6	15.6	15.6		± 4.0	16.0	16.0	16.0		± 4.0	15.5	15.5	15.5		± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	253	253	253		± 30	500	500	500		± 50
VPM MPV	µm <sup>3</sup> ; fl	9.0	9.0	9.0		± 2.0	8.2	8.2	8.2		± 2.0	8.2	8.2	8.2		± 2.0
NEUT	#	1.36	1.43	1.43		± 0.30	4.18	4.29	4.29		± 0.80	13.50	13.40	13.40		± 1.90
	%	56.5	59.5	59.5		± 10.0	55.0	56.5	56.5		± 10.0	72.0	71.8	71.8		± 10.0
LYMPHO	#	0.74	0.65	0.65		± 0.20	2.66	2.32	2.32		± 0.70	3.18	2.84	2.84		± 1.50
	%	31.0	27.0	27.0		± 8.0	35.0	30.5	30.5		± 8.0	17.0	15.2	15.2		± 8.0
MONO	#	0.12	0.14	0.14		± 0.12	0.34	0.46	0.46		± 0.34	0.94	1.22	1.22		± 0.94
	%	5.0	6.0	6.0		± 5.0	4.5	6.0	6.0		± 4.5	5.0	6.5	6.5		± 5.0
EOS	#	0.12	0.12	0.12		± 0.12	0.23	0.27	0.27		± 0.23	0.71	0.75	0.75		± 0.71
	%	5.0	5.0	5.0		± 5.0	3.0	3.5	3.5		± 3.0	3.8	4.0	4.0		± 3.8
BASO	#	0.06	0.06	0.06		± 0.06	0.19	0.27	0.27		± 0.19	0.41	0.47	0.47		± 0.41
	%	2.5	2.5	2.5		± 2.5	2.5	3.5	3.5		± 2.5	2.2	2.5	2.5		± 2.2

Ref: TEMP-0821 Rev.17 BACK / VERSO 9930066-A