

LOT PX 071

CONTROL

(Exp.) 2011-09-05
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.6	2.6	2.5		± 0.4	7.4	7.4	7.8	7.8		7.4	± 1.0	17.8	17.6	18.4		
GR	RBC	10 ¹² /mm ³ ; 10 ¹² /l	2.34	2.28	2.35	2.35	2.32	± 0.12	4.67	4.67	4.74	4.74	4.63	± 0.15	5.20	5.20	5.30	5.30	5.17	± 0.20	
HB	HGB	g/dl	5.8	5.8	5.9	5.9	5.7	± 0.4	13.6	13.5	13.6	13.6	13.5	± 0.5	16.8	16.8	16.6	16.6	16.8	± 0.6	
		g/l	58	58	59	59	57	± 4	136	135	136	136	135	± 5	168	168	166	166	168	± 6	
		mmol/l	3.60	3.60	3.66	3.66	3.54	± 0.25	8.45	8.38	8.45	8.45	8.38	± 0.31	10.43	10.43	10.31	10.31	10.43	± 0.37	
HT	HCT	%	17.6	17.8	18.6	18.6	17.2	± 1.5	39.7	40.2	41.2	41.2	38.9	± 2.0	48.4	48.4	49.3	49.3	47.6	± 2.5	
		l/l	0.176	0.178	0.186	0.186	0.172	± 0.015	0.397	0.402	0.412	0.412	0.389	± 0.020	0.484	0.484	0.493	0.493	0.476	± 0.025	
VGM	MCV	µm ³ ; fl	75	78	79	79	74	± 5.0	85	86	87	87	84	± 4.0	93	93	93	93	92	± 4.0	
TGMH	MCH	pg	24.8	25.4	25.1	25.1	24.6	± 2.0	29.1	28.9	28.7	28.7	29.2	± 2.0	32.3	32.3	31.3	31.3	32.5	± 2.5	
		fmol	1.54	1.58	1.56	1.56	1.53	± 0.12	1.81	1.80	1.78	1.78	1.81	± 0.12	2.01	2.01	1.95	1.95	2.02	± 0.16	
CCMH	MCHC	g/dl	33.0	32.6	31.8	31.8	33.2	± 3.0	34.3	33.6	33.0	33.0	34.7	± 3.0	34.7	34.7	33.7	33.7	35.3	± 3.0	
		g/l	330	326	318	318	332	± 30	343	336	330	330	347	± 30	347	347	337	337	353	± 30	
		mmol/l	20.52	20.25	19.74	19.74	20.62	± 1.86	21.28	20.87	20.48	20.48	21.56	± 1.86	21.57	21.57	20.91	20.91	21.93	± 1.86	
IDR	RDW	%	13.5	12.9	15.2	15.2	13.3	± 4.0	12.2	12.3	14.5	14.5	12.2	± 4.0	11.6	11.7	13.6	13.6	11.2	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	68	70	70	70	± 20	260	255	260	260	260	± 30	495	495	495	495	506	± 50	
VPM	MPV	µm ³ ; fl	9.4	9.4	9.4	9.4	9.3	± 2.0	8.1	8.2	8.2	8.2	8.0	± 2.0	8.0	8.0	8.3	8.3	8.0	± 2.0	
NEUT	#		1.50	1.49	1.55	1.58	1.48	± 0.30	4.66	4.59	4.89	4.98	4.61	± 0.80	13.47	13.09	14.20	14.30	13.17	± 1.90	
		%	60.0	59.5	59.7	60.8	59.0	± 10.0	63.0	62.0	62.7	63.8	62.3	± 10.0	75.7	74.4	77.2	77.7	75.7	± 10.0	
LYMPHO	#		0.72	0.73	0.75	0.72	0.74	± 0.30	2.09	2.17	2.30	2.22	2.15	± 0.70	2.44	2.55	2.67	2.39	2.44	± 1.50	
		%	28.7	29.2	29.0	27.5	29.4	± 8.0	28.3	29.3	29.5	28.4	29.0	± 8.0	13.7	14.5	14.5	13.0	14.0	± 8.0	
MONO	#		0.05	0.05	0.05	0.05	0.05	± 0.05	0.14	0.14	0.14	0.14	0.14	± 0.14	0.33	0.33	0.33	0.33	0.33	± 0.33	
		%	1.8	1.8	1.8	1.8	1.8	± 1.8	1.8	1.8	1.8	1.8	1.8	± 1.8	1.8	1.8	1.8	1.8	1.8	± 1.8	
EOS	#		0.17	0.17	0.17	0.17	0.17	± 0.17	0.27	0.27	0.27	0.27	0.27	± 0.27	0.83	0.88	0.83	0.98	0.83	± 0.83	
		%	6.5	6.5	6.5	6.5	6.5	± 6.5	3.5	3.5	3.5	3.5	3.5	± 3.5	4.5	5.0	4.5	5.3	4.5	± 4.5	
BASO	#		0.08	0.08	0.08	0.09	0.08	± 0.08	0.25	0.25	0.20	0.20	0.25	± 0.20	0.77	0.76	0.37	0.40	0.70	± 0.37	
		%	3.0	3.0	3.0	3.4	3.3	± 3.0	3.4	3.4	2.5	2.5	3.4	± 2.5	4.3	4.3	2.0	2.2	4.0	± 2.0	

Ref: FORM-0821 Rev.13 BACK / VERSO 9930052-B