

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT		PX 033		CONTROL		(Exp.) 2013-05-05										
		Rev 2				(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				L	CONTROL				N	CONTROL				H
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE
60 60C+ ES60	80 XL80	MS60		60 60C+ ES60	80 XL80		MS60		60 60C+ ES60	80 XL80		MS60				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5		± 0.4	7.4	7.4		± 1.0	17.7	17.5	17.6		± 2.2	
GR RBC	10 ¹² /mm ³ ; 10 ¹² /l	2.45	2.38	2.44		± 0.12	4.62	4.60	4.63		± 0.15	5.07	5.09	5.09		± 0.20
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.7	13.6	13.7		± 0.5	15.8	15.9	15.9		± 0.6
	g/l	67	67	67		± 4	137	136	137		± 5	158	159	159		± 6
	mmol/l	4.16	4.16	4.16		± 0.25	8.51	8.45	8.51		± 0.31	9.81	9.87	9.87		± 0.37
HT HCT	%	19.4	19.3	19.0		± 1.5	37.9	38.2	37.5		± 2.0	44.6	44.8	43.3		± 2.5
	lit	0.194	0.193	0.190		± 0.015	0.379	0.382	0.375		± 0.020	0.446	0.448	0.433		± 0.025
VGM MCV	µm ³ ; fl	79	81	78		± 5	82	83	81		± 5	88	88	85		± 5
TGMH MCH	pg	27.3	28.2	27.5		± 2.0	29.7	29.6	29.6		± 2.0	31.2	31.2	31.2		± 2.5
	fmol	1.70	1.75	1.71		± 0.12	1.84	1.84	1.84		± 0.12	1.94	1.94	1.94		± 0.16
	g/dl	34.6	34.8	35.2		± 3.0	36.2	35.6	36.5		± 3.0	35.4	35.5	36.8		± 3.0
CCMH MCHC	g/l	346	348	352		± 30	362	356	365		± 30	354	355	368		± 30
	mmol/l	21.50	21.58	21.86		± 1.86	22.46	22.12	22.69		± 1.86	21.99	22.04	22.82		± 1.86
	%	13.5	13.2	12.6		± 4.0	13.5	13.8	12.5		± 4.0	13.0	13.7	11.8		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	255	255	260		± 30	505	505	512		± 50
VPM MPV	µm ³ ; fl	8.9	8.9	8.8		± 2.0	9.1	9.2	9.0		± 2.0	8.6	8.8	8.6		± 2.0
NEUT	#	1.37	1.36	1.39		± 0.30	4.04	4.03	3.96		± 0.80	12.16	11.73	12.00		± 1.80
	%	57.0	56.5	55.5		± 10.0	54.6	54.5	53.5		± 10.0	68.7	67.0	68.2		± 10.0
LYMPHO	#	0.71	0.72	0.79		± 0.20	2.44	2.44	2.52		± 0.60	3.10	3.41	3.20		± 1.50
	%	29.5	30.0	31.5		± 8.0	33.0	33.0	34.0		± 8.0	17.5	19.5	18.2		± 8.0
MONO	#	0.12	0.12	0.12		± 0.12	0.39	0.39	0.39		± 0.39	0.80	0.79	0.84		± 0.75
	%	5.0	5.0	5.0		± 5.0	5.0	5.0	5.0		± 5.0	4.5	4.5	4.8		± 4.0
EOS	#	0.13	0.13	0.13		± 0.13	0.31	0.33	0.31		± 0.31	0.89	0.88	0.84		± 0.74
	%	5.5	5.5	5.0		± 5.0	4.0	4.5	4.0		± 4.0	5.0	5.0	4.8		± 4.0
BASO	#	0.07	0.07	0.08		± 0.06	0.25	0.23	0.26		± 0.23	0.76	0.70	0.70		± 0.46
	%	3.0	3.0	3.0		± 2.5	3.4	3.0	3.5		± 3.0	4.3	4.0	4.0		± 2.5

LOT		PX 033		CONTROL		(Exp.) 2013-05-05										
		Rev 2				(YYYY-MM-DD)										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														
		CONTROL				L	CONTROL				N	CONTROL				H
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE
120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.7	7.7	7.7		± 1.0	18.5	18.5	18.5		± 2.2
GR RBC	10 ¹² /mm ³ ; 10 ¹² /l	2.44	2.44	2.44		± 0.12	4.64	4.64	4.64		± 0.15	5.17	5.17	5.17		± 0.20
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.7	13.7	13.7		± 0.5	15.9	15.9	15.9		± 0.6
	g/l	67	67	67		± 4	137	137	137		± 5	159	159	159		± 6
	mmol/l	4.16	4.16	4.16		± 0.25	8.51	8.51	8.51		± 0.31	9.87	9.87	9.87		± 0.37
HT HCT	%	19.8	19.8	19.8		± 1.5	39.0	39.0	39.0		± 2.0	45.5	45.5	45.5		± 2.5
	lit	0.198	0.198	0.198		± 0.015	0.390	0.390	0.390		± 0.020	0.455	0.455	0.455		± 0.025
VGM MCV	µm ³ ; fl	81	81	81		± 5	84	84	84		± 5	88	88	88		± 5
TGMH MCH	pg	27.5	27.5	27.5		± 2.0	29.5	29.5	29.5		± 2.0	30.8	30.8	30.8		± 2.5
	fmol	1.71	1.71	1.71		± 0.12	1.83	1.83	1.83		± 0.12	1.91	1.91	1.91		± 0.16
	g/dl	33.9	33.9	33.9		± 3.0	35.1	35.1	35.1		± 3.0	34.9	34.9	34.9		± 3.0
CCMH MCHC	g/l	339	339	339		± 30	351	351	351		± 30	349	349	349		± 30
	mmol/l	21.05	21.05	21.05		± 1.86	21.83	21.83	21.83		± 1.86	21.70	21.70	21.70		± 1.86
	%	15.8	15.8	15.8		± 4.0	16.0	16.0	16.0		± 4.0	14.5	14.5	14.5		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	260	260	260		± 30	505	505	505		± 50
VPM MPV	µm ³ ; fl	8.9	8.9	8.9		± 2.0	9.1	9.1	9.1		± 2.0	8.6	8.6	8.6		± 2.0
NEUT	#	1.37	1.39	1.39		± 0.30	4.24	4.39	4.39		± 0.80	12.80	12.80	12.80		± 1.80
	%	57.0	58.0	58.0		± 10.0	55.0	57.0	57.0		± 10.0	69.0	69.0	69.0		± 10.0
LYMPHO	#	0.73	0.65	0.65		± 0.20	2.54	2.27	2.27		± 0.60	3.42	3.15	3.15		± 1.50
	%	30.5	27.0	27.0		± 8.0	33.0	29.5	29.5		± 8.0	18.5	17.0	17.0		± 8.0
MONO	#	0.12	0.16	0.16		± 0.12	0.39	0.46	0.46		± 0.39	1.11	1.30	1.30		± 0.75
	%	5.0	6.5	6.5		± 5.0	5.0	6.0	6.0		± 5.0	6.0	7.0	7.0		± 4.0
EOS	#	0.13	0.13	0.13		± 0.13	0.31	0.31	0.31		± 0.31	0.74	0.74	0.74		± 0.74
	%	5.0	5.5	5.5		± 5.0	4.0	4.0	4.0		± 4.0	4.0	4.0	4.0		± 4.0
BASO	#	0.06	0.07	0.07		± 0.06	0.23	0.27	0.27		± 0.23	0.46	0.56	0.56		± 0.46
	%	2.5	3.0	3.0		± 2.5	3.0	3.5	3.5		± 3.0	2.5	3.0	3.0		± 2.5

Ref: TEMP-0821 Rev.30 BACK / VERSO 9930073-A