

LOT PX 073

CONTROL

(Exp.) **2013-09-05**
 (YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE			
		60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5	2.4		± 0.4	7.3	7.3	7.3		7.3	± 1.0	17.8	17.5			17.9	17.5
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.37	2.39	2.37	± 0.12	4.58	4.58	4.58	4.58	± 0.15	5.10	5.11	5.08	5.11	± 0.20			
HB HGB	g/dl	6.8	6.8	6.8	6.8	± 0.4	13.3	13.3	13.3	13.3	± 0.5	16.1	16.1	16.1	16.1	± 0.6			
	g/l	68	68	68	68	± 4	133	133	133	133	± 5	161	161	161	161	± 6			
	mmol/l	4.22	4.22	4.22	4.22	± 0.25	8.26	8.26	8.26	8.26	± 0.31	10.00	10.00	10.00	10.00	± 0.37			
HT HCT	%	19.1	19.7	18.6	19.7	± 1.5	36.2	36.6	35.7	36.6	± 2.0	44.4	44.5	43.2	44.5	± 2.5			
	l/l	0.191	0.197	0.186	0.197	± 0.015	0.362	0.366	0.357	0.366	± 0.020	0.444	0.445	0.432	0.445	± 0.025			
VGM MCV	µm ³ ; fl	79	83	78	83	± 5	79	80	78	80	± 5	87	87	85	87	± 5			
TGMH MCH	pg	28.1	28.7	28.5	28.7	± 2.0	29.0	29.0	29.0	29.0	± 2.0	31.6	31.5	31.7	31.5	± 2.5			
	fmol	1.74	1.78	1.77	1.78	± 0.12	1.80	1.80	1.80	1.80	± 0.12	1.96	1.96	1.97	1.96	± 0.16			
CCMH MCHC	g/dl	35.6	34.6	36.5	34.6	± 3.0	36.8	36.3	37.2	36.3	± 3.0	36.3	36.2	37.3	36.2	± 3.0			
	g/l	356	346	365	346	± 30	368	363	372	363	± 30	363	362	373	362	± 30			
	mmol/l	22.09	21.47	22.65	21.47	± 1.86	22.83	22.54	23.12	22.54	± 1.86	22.53	22.49	23.15	22.49	± 1.86			
IDR RDW	%	12.7	13.7	12.5	13.7	± 4.0	12.5	13.6	12.5	13.6	± 4.0	12.3	13.4	11.7	13.4	± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	72	70	72	± 20	250	245	245	245	± 30	510	505	510	505	± 50			
VPM MPV	µm ³ ; fl	9.2	9.4	9.6	9.4	± 2.0	9.4	9.6	9.6	9.6	± 2.0	9.3	9.6	9.5	9.6	± 2.0			
NEUT	#	1.36	1.36	1.38	1.36	± 0.30	4.23	4.25	4.20	4.25	± 0.80	12.82	12.60	12.89	12.60	± 1.90			
	%	56.7	56.7	55.0	56.7	± 10.0	57.9	58.2	57.5	58.2	± 10.0	72.0	72.0	72.0	72.0	± 10.0			
LYMPHO	#	0.73	0.72	0.79	0.72	± 0.25	2.28	2.21	2.26	2.21	± 0.65	2.58	2.54	2.69	2.54	± 1.50			
	%	30.5	30.0	31.7	30.0	± 8.0	31.3	30.3	31.0	30.3	± 8.0	14.5	14.5	15.0	14.5	± 8.0			
MONO	#	0.08	0.08	0.08	0.08	± 0.08	0.30	0.30	0.33	0.30	± 0.30	0.71	0.62	0.62	0.62	± 0.62			
	%	3.5	3.3	3.3	3.3	± 3.3	4.0	4.0	4.5	4.0	± 4.0	4.0	3.5	3.5	3.5	± 3.5			
EOS	#	0.14	0.16	0.16	0.16	± 0.14	0.27	0.29	0.27	0.29	± 0.27	0.89	0.96	0.90	0.96	± 0.83			
	%	6.0	6.5	6.5	6.5	± 5.5	3.5	4.0	3.5	4.0	± 3.5	5.0	5.5	5.0	5.5	± 4.5			
BASO	#	0.08	0.08	0.09	0.08	± 0.08	0.24	0.26	0.26	0.26	± 0.23	0.80	0.79	0.81	0.79	± 0.46			
	%	3.3	3.5	3.5	3.5	± 3.0	3.3	3.5	3.5	3.5	± 3.0	4.5	4.5	4.5	4.5	± 2.5			

LOT PX 073

CONTROL

(Exp.) 2013-09-05
 (YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	$10^9/mm^3; 10^9/l$	2.5	2.5	2.5	± 0.4	7.8	7.8	7.8	± 1.0	18.5	18.5	18.5	± 2.2					
GR RBC	$10^6/mm^3; 10^{12}/l$	2.42	2.42	2.42	± 0.12	4.62	4.62	4.62	± 0.15	5.20	5.20	5.20	± 0.20					
HB HGB	g/dl	6.9	6.9	6.9	± 0.4	13.2	13.2	13.2	± 0.5	16.0	16.0	16.0	± 0.6					
	g/l	69	69	69	± 4	132	132	132	± 5	160	160	160	± 6					
	mmol/l	4.28	4.28	4.28	± 0.25	8.20	8.20	8.20	± 0.31	9.94	9.94	9.94	± 0.37					
HT HCT	%	19.8	19.8	19.8	± 1.5	37.4	37.4	37.4	± 2.0	45.2	45.2	45.2	± 2.5					
	l/l	0.198	0.198	0.198	± 0.015	0.374	0.374	0.374	± 0.020	0.452	0.452	0.452	± 0.025					
VGM MCV	$\mu m^3; fl$	82	82	82	± 5	81	81	81	± 5	87	87	87	± 5					
TGMH MCH	pg	28.5	28.5	28.5	± 2.0	28.6	28.6	28.6	± 2.0	30.8	30.8	30.8	± 2.5					
	fmol	1.77	1.77	1.77	± 0.12	1.77	1.77	1.77	± 0.12	1.91	1.91	1.91	± 0.16					
CCMH MCHC	g/dl	34.8	34.8	34.8	± 3.0	35.3	35.3	35.3	± 3.0	35.4	35.4	35.4	± 3.0					
	g/l	348	348	348	± 30	353	353	353	± 30	354	354	354	± 30					
	mmol/l	21.59	21.59	21.59	± 1.86	21.90	21.90	21.90	± 1.86	21.96	21.96	21.96	± 1.86					
IDR RDW	%	16.2	16.2	16.2	± 4.0	16.8	16.8	16.8	± 4.0	16.0	16.0	16.0	± 4.0					
PLAQ. PLTS	$10^3/mm^3; 10^9/l$	78	78	78	± 20	250	250	250	± 30	510	510	510	± 50					
VPM MPV	$\mu m^3; fl$	9.4	9.4	9.4	± 2.0	9.5	9.5	9.5	± 2.0	9.5	9.5	9.5	± 2.0					
NEUT	#	1.41	1.48	1.48	± 0.30	4.33	4.67	4.67	± 0.80	13.20	13.30	13.30	± 1.90					
	%	56.3	59.0	59.0	± 10.0	55.5	59.9	59.9	± 10.0	71.5	72.1	72.1	± 10.0					
LYMPHO	#	0.76	0.69	0.69	± 0.25	2.48	2.14	2.14	± 0.65	2.72	2.60	2.60	± 1.50					
	%	30.2	27.6	27.6	± 8.0	31.8	27.5	27.5	± 8.0	14.7	14.0	14.0	± 8.0					
MONO	#	0.13	0.10	0.10	± 0.08	0.48	0.42	0.42	± 0.30	1.26	1.18	1.18	± 0.62					
	%	5.0	3.9	3.9	± 3.3	6.2	5.4	5.4	± 4.0	6.8	6.4	6.4	± 3.5					
EOS	#	0.14	0.16	0.16	± 0.14	0.27	0.31	0.31	± 0.27	0.83	0.93	0.93	± 0.83					
	%	5.5	6.5	6.5	± 5.5	3.5	4.0	4.0	± 3.5	4.5	5.0	5.0	± 4.5					
BASO	#	0.08	0.08	0.08	± 0.08	0.23	0.25	0.25	± 0.23	0.46	0.46	0.46	± 0.46					
	%	3.0	3.0	3.0	± 3.0	3.0	3.2	3.2	± 3.0	2.5	2.5	2.5	± 2.5					

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B