


LOT PX 035
Rev 1


CONTROL

 (Exp.) 2015-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.5	± 0.4	7.6	7.6	7.6	7.6	7.6	± 1.0	18.1	17.7	18.0	17.7	17.8	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.39	2.33	2.37	2.33	2.33	± 0.16	4.66	4.66	4.69	4.66	4.60	± 0.20	5.10	5.10	5.08	5.10	5.04	± 0.25	
HB	HGB	g/dl	6.8	6.8	6.8	6.8	6.7	± 0.4	13.7	13.8	13.7	13.8	13.6	± 0.5	16.2	16.3	16.2	16.3	16.1	± 0.6	
		g/l	68	68	68	68	67	± 4	137	138	137	138	136	± 5	162	163	162	163	161	± 6	
		mmol/l	4.22	4.22	4.22	4.22	4.16	± 0.25	8.51	8.57	8.51	8.57	8.45	± 0.31	10.06	10.12	10.06	10.12	10.00	± 0.37	
HT	HCT	%	19.1	19.3	19.0	19.3	18.8	± 1.5	37.7	38.2	37.5	38.2	37.2	± 2.0	44.9	44.9	44.2	44.9	44.4	± 2.5	
		l/l	0.191	0.193	0.190	0.193	0.188	± 0.015	0.377	0.382	0.375	0.382	0.372	± 0.020	0.449	0.449	0.442	0.449	0.444	± 0.025	
VGM	MCV	µm ³ ; fl	80	83	80	83	80.7	± 5	81	82	80	82	80.8	± 5	88	88	87	88	88.0	± 5	
TGMH	MCH	pg	28.5	29.2	28.7	29.2	28.8	± 2.0	29.4	29.6	29.2	29.6	29.6	± 2.0	31.8	32.0	31.9	32.0	31.9	± 2.5	
		fmol	1.77	1.81	1.78	1.81	1.79	± 0.12	1.83	1.84	1.81	1.84	1.84	± 0.12	1.97	1.98	1.98	1.98	1.98	± 0.16	
CCMH	MCHC	g/dl	35.6	35.2	35.9	35.2	35.6	± 3.0	36.3	36.1	36.5	36.1	36.6	± 3.0	36.1	36.3	36.7	36.3	36.3	± 3.0	
		g/l	356	352	359	352	356	± 30	363	361	365	361	366	± 30	361	363	367	363	363	± 30	
		mmol/l	22.09	21.84	22.27	21.84	22.13	± 1.86	22.54	22.43	22.68	22.43	22.72	± 1.86	22.42	22.55	22.76	22.55	22.54	± 1.86	
IDR	RDW	%	12.6	14.3	12.3	14.3	11.5	± 4.0	12.3	13.5	12.2	13.5	10.8	± 4.0	11.8	13.3	11.7	13.3	10.6	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	70	68	70	74	± 20	255	255	255	255	255	± 30	495	505	505	505	500	± 50	
VPM	MPV	µm ³ ; fl	8.3	8.4	8.4	8.4	7.9	± 2.0	9.1	9.4	9.2	9.4	8.6	± 2.0	8.2	8.4	8.2	8.4	7.7	± 2.0	
NEUT	#	#	1.46	1.49	1.50	1.49	1.56	± 0.35	4.37	4.45	4.48	4.45	4.60	± 0.90	12.94	12.76	13.14	12.76	13.23	± 1.90	
		%	58.5	59.5	60.0	59.5	62.2	± 10.0	57.5	58.5	59.0	58.5	60.5	± 10.0	71.5	72.1	73.0	72.1	74.3	± 10.0	
LYMPHO	#	#	0.71	0.70	0.69	0.70	0.66	± 0.31	2.39	2.37	2.33	2.37	2.24	± 0.70	2.62	2.66	2.61	2.66	2.49	± 1.50	
		%	28.5	28.0	27.5	28.0	26.3	± 12.0	31.5	31.2	30.7	31.2	29.5	± 8.0	14.5	15.0	14.5	15.0	14.0	± 8.0	
MONO	#	#	0.09	0.08	0.08	0.08	0.08	± 0.08	0.30	0.23	0.23	0.23	0.23	± 0.23	0.81	0.62	0.53	0.62	0.53	± 0.53	
		%	3.5	3.0	3.0	3.0	3.0	± 3.0	4.0	3.0	3.0	3.0	3.0	± 3.0	4.5	3.5	3.0	3.5	3.0	± 3.0	
EOS	#	#	0.15	0.15	0.15	0.15	0.13	± 0.13	0.27	0.29	0.29	0.29	0.27	± 0.27	0.91	0.89	0.90	0.89	0.80	± 0.80	
		%	6.0	6.0	6.0	6.0	5.0	± 5.0	3.5	3.8	3.8	3.8	3.5	± 3.5	5.0	5.0	5.0	5.0	4.5	± 4.5	
BASO	#	#	0.09	0.09	0.09	0.09	0.09	± 0.09	0.27	0.27	0.27	0.27	0.27	± 0.27	0.81	0.78	0.81	0.78	0.75	± 0.75	
		%	3.5	3.5	3.5	3.5	3.5	± 3.5	3.5	3.5	3.5	3.5	3.5	± 3.5	4.5	4.4	4.5	4.4	4.2	± 4.2	

LOT PX 035
Rev 1

CONTROL

 (Exp.) 2015-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			120	DX120	DX NEXUS	120	DX120	DX NEXUS	120	DX120	DX NEXUS	120	DX120	DX NEXUS						
120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS									
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	7.8	7.8	7.8			± 1.0	18.6	18.6	18.6			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	2.40			± 0.16	4.67	4.67	4.67			± 0.20	5.13	5.13	5.13			± 0.25
		g/dl	6.9	6.9	6.9			± 0.4	13.7	13.7	13.7			± 0.5	16.1	16.1	16.1			± 0.6
HB	HGB	g/l	69	69	69			± 4	137	137	137			± 5	161	161	161			± 6
		mmol/l	4.28	4.28	4.28			± 0.25	8.51	8.51	8.51			± 0.31	10.00	10.00	10.00			± 0.37
HT	HCT	%	19.9	19.9	19.9			± 1.5	38.3	38.3	38.3			± 2.0	45.7	45.7	45.7			± 2.5
		l/l	0.199	0.199	0.199			± 0.015	0.383	0.383	0.383			± 0.020	0.457	0.457	0.457			± 0.025
VGM	MCV	µm ³ ; fl	83	83	83			± 5	82	82	82			± 5	89	89	89			± 5
TGMH	MCH	pg	28.8	28.8	28.8			± 2.0	29.3	29.3	29.3			± 2.0	31.4	31.4	31.4			± 2.5
		fmol	1.79	1.79	1.79			± 0.12	1.82	1.82	1.82			± 0.12	1.95	1.95	1.95			± 0.16
		g/dl	34.6	34.6	34.6			± 3.0	35.8	35.8	35.8			± 3.0	35.3	35.3	35.3			± 3.0
CCMH	MCHC	g/l	346	346	346			± 30	358	358	358			± 30	353	353	353			± 30
		mmol/l	21.51	21.51	21.51			± 1.86	22.22	22.22	22.22			± 1.86	21.90	21.90	21.90			± 1.86
IDR	RDW	%	15.0	15.0	15.0			± 4.0	15.5	15.5	15.5			± 4.0	13.8	13.8	13.8			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74	74			± 20	260	260	260			± 30	500	500	500			± 50
VPM	MPV	µm ³ ; fl	8.3	8.3	8.3			± 2.0	9.0	9.0	9.0			± 2.0	8.1	8.1	8.1			± 2.0
		#	1.47	1.53	1.53			± 0.35	4.52	4.66	4.66			± 0.90	13.40	13.60	13.60			± 1.90
		%	58.7	61.2	61.2			± 10.0	57.9	59.8	59.8			± 10.0	72.0	73.3	73.3			± 10.0
		#	0.72	0.65	0.65			± 0.31	2.39	2.21	2.21			± 0.70	2.72	2.46	2.46			± 1.50
		%	28.7	25.8	25.8			± 12.0	30.6	28.3	28.3			± 8.0	14.6	13.2	13.2			± 8.0
		#	0.12	0.12	0.12			± 0.12	0.41	0.42	0.42			± 0.41	1.15	1.15	1.15			± 1.15
		%	4.7	4.8	4.8			± 4.7	5.2	5.4	5.4			± 5.2	6.2	6.2	6.2			± 6.2
		#	0.14	0.14	0.14			± 0.14	0.26	0.27	0.27			± 0.26	0.87	0.89	0.89			± 0.87
		%	5.4	5.7	5.7			± 5.4	3.3	3.5	3.5			± 3.3	4.7	4.8	4.8			± 4.7
		#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.47	0.47	0.47			± 0.47
		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

9930091-A
Ref: TEMP-0821 Rev.35