


LOT PX 115
Rev 1


CONTROL

 (Exp.) 2016-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.4	2.5	2.4	± 0.4	7.8	7.8	7.8	7.8	7.8	± 1.0	17.8	17.3	17.5	17.3	17.6	± 2.2		
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.42	2.37	2.42	2.37	2.40	± 0.16	4.52	4.51	4.53	4.51	4.47	± 0.20	5.08	5.11	5.12	5.11	5.06	± 0.25		
HB	HGB	g/dl	6.9	6.9	6.9	6.9	6.9	± 0.4	13.6	13.6	13.6	13.6	13.5	± 0.5	16.2	16.3	16.3	16.3	16.2	± 0.6		
		g/l	69	69	69	69	69	± 4	136	136	136	136	135	± 5	162	163	163	163	162	± 6		
		mmol/l	4.28	4.28	4.28	4.28	4.28	± 0.25	8.45	8.45	8.45	8.45	8.38	± 0.31	10.06	10.12	10.12	10.12	10.06	± 0.37		
HT	HCT	%	19.4	19.4	19.4	19.4	19.2	± 1.5	37.1	37.0	36.7	37.0	36.7	± 2.0	44.7	44.5	44.0	44.5	44.0	± 2.5		
		l/l	0.194	0.194	0.194	0.194	0.192	± 0.015	0.371	0.370	0.367	0.370	0.367	± 0.020	0.447	0.445	0.440	0.445	0.440	± 0.025		
VGM	MCV	µm ³ ; fl	80	82	80	82	80.0	± 5	82	82	81	82	82.0	± 5	88	87	86	87	87.0	± 5		
TGMH	MCH	pg	28.5	29.1	28.5	29.1	28.8	± 2.0	30.1	30.2	30.0	30.2	30.2	± 2.0	31.9	31.9	31.8	31.9	32.0	± 2.5		
		fmol	1.77	1.81	1.77	1.81	1.79	± 0.12	1.87	1.87	1.86	1.87	1.88	± 0.12	1.98	1.98	1.98	1.98	1.99	± 0.16		
CCMH	MCHC	g/dl	35.6	35.5	35.6	35.5	35.9	± 3.0	36.7	36.8	37.1	36.8	36.8	± 3.0	36.2	36.7	37.0	36.7	36.8	± 3.0		
		g/l	356	355	356	355	359	± 30	367	368	371	368	368	± 30	362	367	370	367	368	± 30		
		mmol/l	22.13	22.05	22.13	22.05	22.32	± 1.86	22.79	22.84	23.02	22.84	22.87	± 1.86	22.50	22.77	22.99	22.77	22.85	± 1.86		
IDR	RDW	%	13.4	13.0	12.0	13.0	11.8	± 4.0	12.9	12.5	12.2	12.5	12.0	± 4.0	12.4	12.0	11.5	12.0	11.5	± 4.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	78	75	75	± 20	257	260	265	260	260	± 30	520	520	540	520	525	± 50		
VPM	MPV	µm ³ ; fl	9.4	9.5	9.3	9.5	9.0	± 2.0	9.2	9.4	9.0	9.4	8.7	± 2.0	8.6	8.9	8.8	8.9	8.2	± 2.0		
NEUT	#		1.30	1.36	1.32	1.36	1.36	± 0.35	4.33	4.31	4.33	4.31	4.37	± 0.90	12.46	12.02	12.34	12.02	12.72	± 1.90		
		%	54.1	54.3	54.8	54.3	56.6	± 10.0	55.5	55.2	55.5	55.2	56.0	± 10.0	70.0	69.5	70.5	69.5	72.3	± 10.0		
LYMPHO	#		0.75	0.80	0.76	0.80	0.74	± 0.33	2.54	2.51	2.56	2.51	2.51	± 0.70	2.78	2.77	2.73	2.77	2.64	± 1.50		
		%	31.3	32.0	31.8	32.0	31.0	± 12.0	32.5	32.2	32.8	32.2	32.2	± 8.0	15.6	16.0	15.6	16.0	15.0	± 8.0		
MONO	#		0.12	0.09	0.08	0.09	0.08	± 0.08	0.35	0.30	0.31	0.30	0.27	± 0.27	0.89	0.78	0.79	0.78	0.62	± 0.62		
		%	4.8	3.5	3.2	3.5	3.2	± 3.2	4.5	3.8	4.0	3.8	3.5	± 3.5	5.0	4.5	4.5	4.5	3.5	± 3.5		
EOS	#		0.16	0.17	0.16	0.17	0.14	± 0.14	0.31	0.39	0.33	0.39	0.37	± 0.31	0.89	0.95	0.88	0.95	0.88	± 0.88		
		%	6.5	6.8	6.8	6.8	5.8	± 5.8	4.0	5.0	4.2	5.0	4.8	± 4.0	5.0	5.5	5.0	5.5	5.0	± 5.0		
BASO	#		0.08	0.09	0.08	0.09	0.08	± 0.08	0.27	0.30	0.27	0.30	0.27	± 0.27	0.78	0.78	0.77	0.78	0.74	± 0.74		
		%	3.3	3.4	3.4	3.4	3.4	± 3.3	3.5	3.8	3.5	3.8	3.5	± 3.5	4.4	4.5	4.4	4.5	4.2	± 4.2		

LOT PX 115
Rev 1

CONTROL

 (Exp.) 2016-01-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	8.0	8.0	8.0			± 1.0	18.5	18.5	18.5			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.43	2.43			± 0.16	4.53	4.53	4.53			± 0.20	5.16	5.16	5.16			± 0.25	
	g/dl	7.0	7.0	7.0			± 0.4	13.6	13.6	13.6			± 0.5	16.1	16.1	16.1			± 0.6	
HB HGB	g/l	70	70	70			± 4	136	136	136			± 5	161	161	161			± 6	
	mmol/l	4.35	4.35	4.35			± 0.25	8.45	8.45	8.45			± 0.31	10.00	10.00	10.00			± 0.37	
HT HCT	%	20.2	20.2	20.2			± 1.5	37.6	37.6	37.6			± 2.0	45.4	45.4	45.4			± 2.5	
	l/l	0.202	0.202	0.202			± 0.015	0.376	0.376	0.376			± 0.020	0.454	0.454	0.454			± 0.025	
VGM MCV	µm ³ ; fl	83	83	83			± 5	83	83	83			± 5	88	88	88			± 5	
TGMH MCH	pg	28.8	28.8	28.8			± 2.0	30.0	30.0	30.0			± 2.0	31.2	31.2	31.2			± 2.5	
	fmol	1.79	1.79	1.79			± 0.12	1.86	1.86	1.86			± 0.12	1.94	1.94	1.94			± 0.16	
	g/dl	34.7	34.7	34.7			± 3.0	36.2	36.2	36.2			± 3.0	35.5	35.5	35.5			± 3.0	
CCMH MCHC	g/l	347	347	347			± 30	362	362	362			± 30	355	355	355			± 30	
	mmol/l	21.55	21.55	21.55			± 1.86	22.46	22.46	22.46			± 1.86	22.02	22.02	22.02			± 1.86	
IDR RDW	%	14.7	14.7	14.7			± 4.0	15.7	15.7	15.7			± 4.0	14.4	14.4	14.4			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	80	80			± 20	268	268	268			± 30	520	520	520			± 50	
VPM MPV	µm ³ ; fl	9.3	9.3	9.3			± 2.0	9.0	9.0	9.0			± 2.0	8.5	8.5	8.5			± 2.0	
NEUT	#	1.35	1.42	1.42			± 0.35	4.43	4.58	4.58			± 0.90	13.00	13.20	13.20			± 1.90	
	%	54.1	56.7	56.7			± 10.0	55.4	57.2	57.2			± 10.0	70.0	71.4	71.4			± 10.0	
LYMPHO	#	0.78	0.70	0.70			± 0.33	2.50	2.31	2.31			± 0.70	2.72	2.46	2.46			± 1.50	
	%	31.0	27.9	27.9			± 12.0	31.2	28.9	28.9			± 8.0	14.7	13.3	13.3			± 8.0	
MONO	#	0.15	0.15	0.15			± 0.15	0.59	0.62	0.62			± 0.59	1.57	1.55	1.55			± 1.55	
	%	6.0	6.1	6.1			± 6.0	7.4	7.7	7.7			± 7.4	8.5	8.4	8.4			± 8.4	
EOS	#	0.16	0.17	0.17			± 0.16	0.24	0.26	0.26			± 0.24	0.80	0.81	0.81			± 0.80	
	%	6.4	6.8	6.8			± 6.4	3.0	3.2	3.2			± 3.0	4.3	4.4	4.4			± 4.3	
BASO	#	0.06	0.06	0.06			± 0.06	0.24	0.24	0.24			± 0.24	0.46	0.46	0.46			± 0.46	
	%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5	

9930096-A FRONT / RECTO Ref: TEMP-0821 Rev.36