

LOT PX 094
 Rev 1

CONTROL

(Exp.) 2014-11-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebilo																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		60 60C+ ES60	80 XL80	MS60	XL80	MSCRIP		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4	2.4	2.5		± 0.4	7.4	7.5	7.5	7.5		7.4	± 1.0	17.9	17.7	17.7		
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.49	2.45	2.46	2.45	2.42	± 0.16	4.63	4.62	4.65	4.62	4.56	± 0.20	5.10	5.10	5.10	5.10	5.00	± 0.25	
HB	HGB	g/dl	7.1	7.1	7.1	7.1	7.0	± 0.4	13.7	13.7	13.7	13.7	13.5	± 0.5	16.5	16.6	16.5	16.6	16.3	± 0.6	
		g/l	71	71	71	71	70	± 4	137	137	137	137	135	± 5	165	166	165	166	163	± 6	
		mmol/l	4.41	4.41	4.41	4.41	4.35	± 0.25	8.51	8.51	8.51	8.51	8.38	± 0.31	10.25	10.31	10.25	10.31	10.12	± 0.37	
HT	HCT	%	19.9	20.3	19.7	20.3	19.6	± 1.5	38.0	38.3	37.2	38.3	37.9	± 2.0	45.4	45.4	44.4	45.4	44.6	± 2.5	
		l/l	0.199	0.203	0.197	0.203	0.196	± 0.015	0.380	0.383	0.372	0.383	0.379	± 0.020	0.454	0.454	0.444	0.454	0.446	± 0.025	
VGM	MCV	µm ³ ·fl	80	83	80	83	81.1	± 5	82	83	80	83	83.1	± 5	89	89	87	89	89.3	± 5	
TGMH	MCH	pg	28.5	29.0	28.9	29.0	28.9	± 2.0	29.6	29.7	29.5	29.7	29.6	± 2.0	32.4	32.5	32.4	32.5	32.6	± 2.5	
		fmol	1.77	1.80	1.79	1.80	1.80	± 0.12	1.84	1.84	1.83	1.84	1.84	± 0.12	2.01	2.02	2.01	2.02	2.02	± 0.16	
CCMH	MCHC	g/dl	35.6	34.9	36.1	34.9	35.7	± 3.0	36.1	35.7	36.8	35.7	35.6	± 3.0	36.4	36.6	37.2	36.6	36.5	± 3.0	
		g/l	356	349	361	349	357	± 30	361	357	368	357	356	± 30	364	366	372	366	365	± 30	
		mmol/l	22.13	21.68	22.40	21.68	22.15	± 1.86	22.41	22.19	22.87	22.19	22.12	± 1.86	22.57	22.71	23.09	22.71	22.67	± 1.86	
IDR	RDW	%	12.6	13.6	12.7	13.6	10.7	± 4.0	12.5	14.0	12.6	14.0	10.7	± 4.0	12.0	13.6	12.0	13.6	10.7	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	72	70	72	74	± 2.0	252	253	254	253	256	± 3.0	505	510	515	510	511	± 5.0	
VPM	MPV	µm ³ ·fl	9.7	10.0	9.8	10.0	8.9	± 2.0	9.6	9.9	9.8	9.9	8.9	± 2.0	9.5	9.8	9.6	9.8	8.7	± 2.0	
NEUT	#	%	1.33	1.37	1.37	1.37	1.49	± 0.35	4.14	4.24	4.25	4.24	4.28	± 0.90	12.94	12.83	12.97	12.83	12.97	± 1.90	
		%	55.5	57.2	57.0	57.2	59.4	± 10.0	56.0	56.5	56.7	56.5	57.8	± 10.0	72.3	72.5	73.3	72.5	74.1	± 10.0	
LYMPHO	#	%	0.77	0.73	0.73	0.73	0.71	± 0.30	2.44	2.48	2.48	2.48	2.32	± 0.70	2.77	2.74	2.66	2.74	2.49	± 1.50	
		%	32.0	30.5	30.5	30.5	28.3	± 12.0	33.0	33.0	33.0	33.0	31.4	± 8.0	15.5	15.5	15.0	15.5	14.2	± 8.0	
MONO	#	%	0.08	0.07	0.07	0.07	0.09	± 0.07	0.27	0.23	0.23	0.23	0.27	± 0.23	0.63	0.62	0.53	0.62	0.61	± 0.53	
		%	3.5	3.0	3.0	3.0	3.5	± 3.0	3.6	3.0	3.0	3.0	3.6	± 3.0	3.5	3.5	3.0	3.5	3.5	± 3.0	
EOS	#	%	0.13	0.14	0.14	0.14	0.14	± 0.13	0.28	0.30	0.29	0.30	0.28	± 0.28	0.75	0.71	0.74	0.71	0.70	± 0.70	
		%	5.5	6.0	6.0	6.0	5.5	± 5.5	3.8	4.0	3.8	4.0	3.8	± 3.8	4.2	4.0	4.2	4.0	4.0	± 4.0	
BASO	#	%	0.08	0.08	0.08	0.08	0.08	± 0.08	0.27	0.26	0.26	0.26	0.25	± 0.25	0.81	0.80	0.80	0.80	0.74	± 0.74	
		%	3.5	3.3	3.5	3.3	3.3	± 3.3	3.6	3.5	3.5	3.5	3.4	± 3.4	4.5	4.5	4.5	4.5	4.2	± 4.2	

LOT PX 094
 Rev 1

CONTROL

(Exp.) 2014-11-05
 (YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120	DX120	DX NEXUS		120		DX120	DX NEXUS		120	DX120		DX NEXUS		120	DX120	DX NEXUS	
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8		± 1.0	18.4	18.4	18.4		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.48	2.48		± 0.16	4.64	4.64	4.64		± 0.20	5.17	5.17	5.17		± 0.25			
HB HGB	g/dl	7.2	7.2	7.2		± 0.4	13.6	13.6	13.6		± 0.5	16.4	16.4	16.4		± 0.6			
	g/l	72	72	72		± 4	136	136	136		± 5	164	164	164		± 6			
	mmol/l	4.47	4.47	4.47		± 0.25	8.45	8.45	8.45		± 0.31	10.18	10.18	10.18		± 0.37			
HT HCT	%	20.6	20.6	20.6		± 1.5	38.5	38.5	38.5		± 2.0	46.0	46.0	46.0		± 2.5			
	l/l	0.206	0.206	0.206		± 0.015	0.385	0.385	0.385		± 0.020	0.460	0.460	0.460		± 0.025			
VGM MCV	µm ³ ·fl	83	83	83		± 5	83	83	83		± 5	89	89	89		± 5			
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.3	29.3	29.3		± 2.0	31.7	31.7	31.7		± 2.5			
	fmol	1.80	1.80	1.80		± 0.12	1.82	1.82	1.82		± 0.12	1.97	1.97	1.97		± 0.16			
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	35.3	35.3	35.3		± 3.0	35.6	35.6	35.6		± 3.0			
	g/l	350	350	350		± 30	353	353	353		± 30	356	356	356		± 30			
	mmol/l	21.72	21.72	21.72		± 1.86	21.93	21.93	21.93		± 1.86	22.13	22.13	22.13		± 1.86			
IDR RDW	%	14.8	14.8	14.8		± 4.0	15.8	15.8	15.8		± 4.0	14.7	14.7	14.7		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	78	78		± 20	262	262	262		± 30	520	520	520		± 50			
VPM MPV	µm ³ ·fl	9.5	9.5	9.5		± 2.0	9.5	9.5	9.5		± 2.0	9.2	9.2	9.2		± 2.0			
NEUT	#	1.45	1.51	1.51		± 0.35	4.41	4.56	4.56		± 0.90	13.40	13.70	13.70		± 1.90			
	%	58.0	60.2	60.2		± 10.0	56.6	58.4	58.4		± 10.0	72.8	74.3	74.3		± 10.0			
LYMPHO	#	0.74	0.68	0.68		± 0.30	2.46	2.30	2.30		± 0.70	2.74	2.54	2.54		± 1.50			
	%	29.6	27.0	27.0		± 12.0	31.6	29.5	29.5		± 8.0	14.9	13.8	13.8		± 8.0			
MONO	#	0.10	0.11	0.11		± 0.10	0.41	0.43	0.43		± 0.41	1.07	1.05	1.05		± 1.05			
	%	4.1	4.2	4.2		± 4.1	5.3	5.5	5.5		± 5.3	5.8	5.7	5.7		± 5.7			
EOS	#	0.13	0.14	0.14		± 0.13	0.26	0.28	0.28		± 0.26	0.68	0.68	0.68		± 0.68			
	%	5.3	5.6	5.6		± 5.3	3.3	3.6	3.6		± 3.3	3.7	3.7	3.7		± 3.7			
BASO	#	0.08	0.08	0.08		± 0.08	0.25	0.23	0.23		± 0.23	0.52	0.46	0.46		± 0.46			
	%	3.0	3.0	3.0		± 3.0	3.2	3.0	3.0		± 3.0	2.8	2.5	2.5		± 2.5			

9930091-A
 Ref. TEMP-0821 Rev.35