

LOT **PX 054**
 Rev 2

CONTROL

(Exp.) **2014-07-05**
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE	
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.7	2.6	2.7		± 0.4	7.3	7.4	7.4		7.4	± 1.0	17.8	17.7			17.9
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.32	2.36	2.32	± 0.16	4.59	4.55	4.54	4.55	± 0.20	5.15	5.15	5.16	5.15	± 0.25		
HB	HGB	g/dl	6.7	6.7	6.6	6.7	± 0.4	13.7	13.7	13.7	13.7	± 0.5	16.6	16.6	16.5	16.6	± 0.6		
		g/l	67	67	66	67	± 4	137	137	137	137	± 5	166	166	165	166	± 6		
		mmol/l	4.16	4.16	4.10	4.16	± 0.25	8.51	8.51	8.51	8.51	± 0.31	10.31	10.31	10.25	10.31	± 0.37		
HT	HCT	%	19.0	19.3	18.6	19.3	± 1.5	38.1	38.2	36.8	38.2	± 2.0	45.3	45.3	44.4	45.3	± 2.5		
		l/l	0.190	0.193	0.186	0.193	± 0.015	0.381	0.382	0.368	0.382	± 0.020	0.453	0.453	0.444	0.453	± 0.025		
VGM	MCV	µm ³ ; fl	80	83	79	83	± 5	83	84	81	84	± 5	88	88	86	88	± 5		
TGMH	MCH	pg	28.2	28.9	28.0	28.9	± 2.0	29.8	30.1	30.2	30.1	± 2.0	32.2	32.2	32.0	32.2	± 2.5		
		fmol	1.75	1.79	1.74	1.79	± 0.12	1.85	1.87	1.87	1.87	± 0.12	2.00	2.00	1.99	2.00	± 0.16		
CCMH	MCHC	g/dl	35.2	34.8	35.4	34.8	± 3.0	36.0	35.8	37.3	35.8	± 3.0	36.6	36.6	37.2	36.6	± 3.0		
		g/l	352	348	354	348	± 30	360	358	373	358	± 30	366	366	372	366	± 30		
		mmol/l	21.85	21.61	21.98	21.61	± 1.86	22.33	22.26	23.14	22.26	± 1.86	22.75	22.75	23.09	22.75	± 1.86		
IDR	RDW	%	13.0	13.0	12.0	13.0	± 4.0	12.7	13.6	12.2	13.6	± 4.0	12.0	13.3	11.5	13.3	± 4.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	70	68	70	± 20	255	255	255	255	± 30	490	495	500	495	± 50		
VPM	MPV	µm ³ ; fl	9.0	9.1	9.2	9.1	± 2.0	8.8	9.0	9.0	9.0	± 2.0	8.5	8.6	8.5	8.6	± 2.0		
NEUT	#	#	1.49	1.58	1.46	1.58	± 0.35	4.15	4.14	4.13	4.14	± 0.90	12.82	12.53	12.89	12.53	± 1.90		
		%	57.3	58.7	56.3	58.7	± 10.0	56.8	56.0	55.8	56.0	± 10.0	72.0	70.8	72.0	70.8	± 10.0		
LYMPHO	#	#	0.68	0.70	0.69	0.70	± 0.40	2.31	2.37	2.41	2.37	± 0.70	2.60	2.74	2.60	2.74	± 1.50		
		%	26.0	25.8	26.5	25.8	± 12.0	31.7	32.0	32.5	32.0	± 8.0	14.6	15.5	14.5	15.5	± 8.0		
MONO	#	#	0.17	0.14	0.18	0.14	± 0.14	0.30	0.30	0.30	0.30	± 0.30	0.76	0.76	0.76	0.76	± 0.76		
		%	6.5	5.3	7.0	5.3	± 5.3	4.0	4.0	4.0	4.0	± 4.0	4.0	4.0	4.0	4.0	± 4.0		
EOS	#	#	0.19	0.19	0.19	0.19	± 0.19	0.29	0.31	0.31	0.31	± 0.29	0.87	0.92	0.90	0.92	± 0.87		
		%	7.0	7.0	7.0	7.0	± 7.0	4.0	4.2	4.2	4.2	± 4.0	4.9	5.2	5.0	5.2	± 4.9		
BASO	#	#	0.08	0.09	0.08	0.09	± 0.08	0.26	0.28	0.26	0.28	± 0.26	0.80	0.80	0.81	0.80	± 0.80		
		%	3.2	3.2	3.2	3.2	± 3.2	3.5	3.8	3.5	3.8	± 3.5	4.5	4.5	4.5	4.5	± 4.5		

LOT **PX 054**
 Rev 2

CONTROL

(Exp.) **2014-07-05**
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE			
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7	2.7		± 0.4	7.6	7.6	7.6		± 1.0	19.0	19.0	19.0		± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	2.40		± 0.16	4.60	4.60	4.60		± 0.20	5.26	5.26	5.26		± 0.25
HB	HGB	g/dl	6.8	6.8	6.8		± 0.4	13.7	13.7	13.7		± 0.5	16.5	16.5	16.5		± 0.6
		g/l	68	68	68		± 4	137	137	137		± 5	165	165	165		± 6
		mmol/l	4.22	4.22	4.22		± 0.25	8.51	8.51	8.51		± 0.31	10.25	10.25	10.25		± 0.37
HT	HCT	%	19.7	19.7	19.7		± 1.5	38.6	38.6	38.6		± 2.0	46.3	46.3	46.3		± 2.5
		l/l	0.197	0.197	0.197		± 0.015	0.386	0.386	0.386		± 0.020	0.463	0.463	0.463		± 0.025
VGM	MCV	µm ³ ·fl	82	82	82		± 5	84	84	84		± 5	88	88	88		± 5
TGMH	MCH	pg	28.3	28.3	28.3		± 2.0	29.8	29.8	29.8		± 2.0	31.4	31.4	31.4		± 2.5
		fmol	1.76	1.76	1.76		± 0.12	1.85	1.85	1.85		± 0.12	1.95	1.95	1.95		± 0.16
CCMH	MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.5	35.5	35.5		± 3.0	35.6	35.6	35.6		± 3.0
		g/l	346	346	346		± 30	355	355	355		± 30	356	356	356		± 30
		mmol/l	21.46	21.46	21.46		± 1.86	22.02	22.02	22.02		± 1.86	22.14	22.14	22.14		± 1.86
IDR	RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	15.0	15.0	15.0		± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	263	263	263		± 30	505	505	505		± 50
VPM	MPV	µm ³ ·fl	9.1	9.1	9.1		± 2.0	8.9	8.9	8.9		± 2.0	8.5	8.5	8.5		± 2.0
NEUT	#		1.61	1.61	1.61		± 0.35	4.31	4.54	4.54		± 0.90	13.80	14.40	14.40		± 1.90
		%	59.5	59.8	59.8		± 10.0	56.7	59.7	59.7		± 10.0	72.6	75.8	75.8		± 10.0
LYMPHO	#		0.68	0.65	0.65		± 0.40	2.36	2.23	2.23		± 0.70	2.85	2.62	2.62		± 1.50
		%	25.0	24.0	24.0		± 12.0	31.0	29.3	29.3		± 8.0	15.0	13.8	13.8		± 8.0
MONO	#		0.15	0.14	0.14		± 0.14	0.46	0.36	0.36		± 0.36	1.14	0.76	0.76		± 0.76
		%	5.5	5.0	5.0		± 5.0	6.0	4.7	4.7		± 4.7	6.0	4.0	4.0		± 4.0
EOS	#		0.19	0.22	0.22		± 0.19	0.25	0.25	0.25		± 0.25	0.74	0.74	0.74		± 0.74
		%	7.0	8.2	8.2		± 7.0	3.3	3.3	3.3		± 3.3	3.9	3.9	3.9		± 3.9
BASO	#		0.08	0.08	0.08		± 0.08	0.23	0.23	0.23		± 0.23	0.48	0.48	0.48		± 0.48
		%	3.0	3.0	3.0		± 3.0	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5

9930085-A
 Ref. TEMP-0821 Rev.34