

**LOT PX 034**

**CONTROL**

(Exp.) **2014-05-05**  
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE			
			60	80	MS60	XL80		60	80	MS60	XL80		60	80	MS60	XL80				
60C+ ES60	XL80	MS60	XL80	60C+ ES60	XL80	MS60		XL80	60C+ ES60	XL80	MS60		XL80							
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4		± 0.4	7.4	7.4	7.3	7.4		± 1.0	17.6	17.3	17.4	17.3		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.25	2.28	2.25		± 0.12	4.57	4.52	4.53	4.52		± 0.15	5.08	5.05	5.04	5.05		± 0.20
HB	HGB	g/dl	6.5	6.5	6.5	6.5		± 0.4	13.5	13.5	13.5	13.5		± 0.5	15.9	16.0	15.9	16.0		± 0.6
		g/l	65	65	65	65		± 4	135	135	135	135		± 5	159	160	159	160		± 6
		mmol/l	4.04	4.04	4.04	4.04		± 0.25	8.38	8.38	8.38	8.38		± 0.31	9.87	9.94	9.87	9.94		± 0.37
HT	HCT	%	18.5	18.7	18.0	18.7		± 1.5	37.5	37.1	36.2	37.1		± 2.0	44.7	44.4	42.8	44.4		± 2.5
		l/l	0.185	0.187	0.180	0.187		± 0.015	0.375	0.371	0.362	0.371		± 0.020	0.447	0.444	0.428	0.444		± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	80	83	79	83		± 5	82	82	80	82		± 5	88	88	85	88		± 5
TGMH	MCH	pg	28.1	28.9	28.5	28.9		± 2.0	29.5	29.9	29.8	29.9		± 2.0	31.3	31.7	31.5	31.7		± 2.5
		fmol	1.75	1.79	1.77	1.79		± 0.12	1.83	1.85	1.85	1.85		± 0.12	1.94	1.97	1.96	1.97		± 0.16
CCMH	MCHC	g/dl	35.2	34.8	36.1	34.8		± 3.0	36.0	36.4	37.3	36.4		± 3.0	35.6	36.0	37.1	36.0		± 3.0
		g/l	352	348	361	348		± 30	360	364	373	364		± 30	356	360	371	360		± 30
		mmol/l	21.84	21.61	22.41	21.61		± 1.86	22.37	22.62	23.13	22.62		± 1.86	22.09	22.36	23.05	22.36		± 1.86
IDR	RDW	%	12.4	13.5	12.8	13.5		± 4.0	12.3	12.9	12.5	12.9		± 4.0	12.0	12.9	12.1	12.9		± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	76	75	73	75		± 20	254	255	255	255		± 30	485	495	500	495		± 50
VPM	MPV	µm <sup>3</sup> ; fl	8.7	9.0	9.0	9.0		± 2.0	8.9	9.3	9.2	9.3		± 2.0	8.5	8.9	8.8	8.9		± 2.0
NEUT	#		1.30	1.33	1.30	1.33		± 0.30	4.12	4.13	4.06	4.13		± 0.80	12.32	11.97	12.23	11.97		± 1.90
		%	54.0	55.3	54.0	55.3		± 10.0	55.7	55.8	55.6	55.8		± 10.0	70.0	69.2	70.3	69.2		± 10.0
LYMPHO	#		0.73	0.71	0.73	0.71		± 0.35	2.44	2.44	2.42	2.44		± 0.65	2.82	2.91	2.78	2.91		± 1.50
		%	30.5	29.7	30.5	29.7		± 12.0	33.0	33.0	33.2	33.0		± 8.0	16.0	16.8	16.0	16.8		± 8.0
MONO	#		0.13	0.13	0.13	0.13		± 0.13	0.30	0.30	0.30	0.30		± 0.30	0.79	0.71	0.71	0.71		± 0.71
		%	5.5	5.0	5.5	5.0		± 5.0	4.0	4.0	4.0	4.0		± 4.0	4.5	4.0	4.0	4.0		± 4.0
EOS	#		0.18	0.18	0.18	0.18		± 0.18	0.29	0.29	0.29	0.29		± 0.29	0.91	0.95	0.94	0.95		± 0.91
		%	7.0	7.0	7.0	7.0		± 7.0	3.8	3.8	3.8	3.8		± 3.8	5.0	5.5	5.4	5.5		± 5.0
BASO	#		0.07	0.07	0.07	0.07		± 0.07	0.26	0.25	0.25	0.25		± 0.19	0.79	0.78	0.75	0.78		± 0.36
		%	3.0	3.0	3.0	3.0		± 2.5	3.5	3.4	3.4	3.4		± 2.5	4.5	4.5	4.3	4.5		± 2.0

**LOT** PX 034

**CONTROL**

 (Exp.) 2014-05-05  
(YYYY-MM-DD)

Rev 1

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6	2.6		± 0.4	7.6	7.6	7.6		± 1.0	18.2	18.2	18.6		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.32	2.32	2.32		± 0.12	4.58	4.58	4.58		± 0.15	5.16	5.16	5.16		± 0.20			
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	13.5	13.5	13.5		± 0.5	15.8	15.9	15.9		± 0.6			
	g/l	66	66	66		± 4	135	135	135		± 5	158	159	159		± 6			
	mmol/l	4.10	4.10	4.10		± 0.25	8.38	8.38	8.38		± 0.31	9.81	9.87	9.87		± 0.37			
HT HCT	%	19.0	19.0	19.0		± 1.5	38.0	38.0	38.0		± 2.0	45.4	45.4	45.4		± 2.5			
	l/l	0.190	0.190	0.190		± 0.015	0.380	0.380	0.380		± 0.020	0.454	0.454	0.454		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	82	82	82		± 5	83	83	83		± 5	88	88	88		± 5			
TGMH MCH	pg	28.4	28.4	28.4		± 2.0	29.5	29.5	29.5		± 2.0	30.6	30.8	30.8		± 2.5			
	fmol	1.77	1.77	1.77		± 0.12	1.83	1.83	1.83		± 0.12	1.90	1.91	1.91		± 0.16			
CCMH MCHC	g/dl	34.7	34.7	34.7		± 3.0	35.5	35.5	35.5		± 3.0	34.8	35.0	35.0		± 3.0			
	g/l	347	347	347		± 30	355	355	355		± 30	348	350	350		± 30			
	mmol/l	21.54	21.54	21.54		± 1.86	22.05	22.05	22.05		± 1.86	21.61	21.74	21.74		± 1.86			
IDR RDW	%	16.0	16.0	16.0		± 4.0	16.8	16.8	16.8		± 4.0	15.2	15.2	15.2		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	80	80	80		± 20	262	262	262		± 30	500	500	500		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.1	9.1	9.1		± 2.0	9.2	9.2	9.2		± 2.0	8.7	8.7	8.7		± 2.0			
NEUT	#	1.46	1.54	1.54		± 0.30	4.35	4.50	4.50		± 0.80	13.10	13.30	13.60		± 1.90			
	%	56.3	59.3	59.3		± 10.0	57.2	59.2	59.2		± 10.0	72.0	73.1	73.1		± 10.0			
LYMPHO	#	0.76	0.63	0.63		± 0.35	2.43	2.26	2.26		± 0.65	2.91	2.71	2.77		± 1.50			
	%	29.2	24.2	24.2		± 12.0	32.0	29.8	29.8		± 8.0	16.0	14.9	14.9		± 8.0			
MONO	#	0.13	0.13	0.13		± 0.13	0.34	0.32	0.32		± 0.30	0.91	0.80	0.82		± 0.71			
	%	5.0	5.0	5.0		± 5.0	4.5	4.2	4.2		± 4.0	5.0	4.4	4.4		± 4.0			
EOS	#	0.18	0.22	0.22		± 0.18	0.29	0.29	0.29		± 0.29	0.91	0.93	0.95		± 0.91			
	%	7.0	8.5	8.5		± 7.0	3.8	3.8	3.8		± 3.8	5.0	5.1	5.1		± 5.0			
BASO	#	0.07	0.08	0.08		± 0.07	0.19	0.23	0.23		± 0.19	0.36	0.46	0.47		± 0.36			
	%	2.5	3.0	3.0		± 2.5	2.5	3.0	3.0		± 2.5	2.0	2.5	2.5		± 2.0			

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B