

**LOT** PX 070

**CONTROL**

(Exp.) 2010-09-05  
 (YYYY - MM - DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5	2.5		± 0.4	7.4	7.4	7.8	7.8		± 1.0	17.3	17.3	18.0	18.0		± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.49	2.47	2.51	2.51		± 0.12	4.65	4.68	4.73	4.73		± 0.15	5.10	5.13	5.24	5.24		± 0.20
HB HGB	g/dl	6.6	6.6	6.6	6.6		± 0.4	14.0	14.0	14.0	14.0		± 0.5	16.5	16.6	16.5	16.5		± 0.6
	g/l	66	66	66	66		± 4	140	140	140	140		± 5	165	166	165	165		± 6
	mmol/l	4.10	4.10	4.10	4.10		± 0.25	8.69	8.69	8.69	8.69		± 0.31	10.25	10.31	10.25	10.25		± 0.37
HT HCT	%	19.4	20.0	20.3	20.3		± 1.5	40.5	40.7	41.6	41.6		± 2.0	47.9	47.7	48.7	48.7		± 2.5
	l/l	0.194	0.200	0.203	0.203		± 0.015	0.405	0.407	0.416	0.416		± 0.020	0.479	0.477	0.487	0.487		± 0.025
VGM MCV	µm <sup>3</sup> fl	78	81	81	81		± 4	87	87	88	88		± 4	94	93	93	93		± 4
TGMH MCH	pg	26.5	26.7	26.3	26.3		± 2.0	30.1	29.9	29.6	29.6		± 2.0	32.4	32.4	31.5	31.5		± 2.5
	fmol	1.65	1.66	1.63	1.63		± 0.12	1.87	1.86	1.84	1.84		± 0.12	2.01	2.01	1.96	1.96		± 0.16
CCMH MCHC	g/dl	34.0	33.0	32.5	32.5		± 3.0	34.6	34.4	33.6	33.6		± 3.0	34.4	34.8	33.9	33.9		± 3.0
	g/l	340	330	325	325		± 30	346	344	336	336		± 30	344	348	339	339		± 30
	mmol/l	21.10	20.49	20.16	20.16		± 1.86	21.49	21.35	20.89	20.89		± 1.86	21.37	21.61	21.03	21.03		± 1.86
IDR RDW	%	13.6	13.6	15.2	15.2		± 4.0	12.0	12.7	13.9	13.9		± 4.0	11.2	11.7	13.4	13.4		± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	65	67	67		± 20	245	245	248	248		± 30	475	476	475	475		± 50
VPM MPV	µm <sup>3</sup> fl	8.6	8.6	8.3	8.3		± 2.0	8.0	8.1	8.1	8.1		± 2.0	8.3	8.4	8.4	8.4		± 2.0
NEUT	#	1.43	1.40	1.46	1.46		± 0.30	4.70	4.66	4.88	4.88		± 0.80	13.06	13.08	14.00	14.00		± 1.80
	%	59.5	58.2	58.2	58.2		± 10.0	63.5	63.0	62.5	62.5		± 10.0	75.5	75.6	77.5	77.5		± 10.0
LYMPHO	#	0.68	0.70	0.75	0.75		± 0.20	2.04	2.07	2.26	2.26		± 0.60	2.16	2.16	2.34	2.34		± 1.50
	%	28.5	29.0	29.8	29.8		± 8.0	27.5	28.0	29.0	29.0		± 8.0	12.5	12.5	13.0	13.0		± 8.0
MONO	#	0.05	0.05	0.05	0.05		± 0.05	0.16	0.16	0.16	0.16		± 0.16	0.36	0.36	0.36	0.36		± 0.36
	%	2.0	2.0	2.0	2.0		± 2.0	2.0	2.0	2.0	2.0		± 2.0	2.0	2.0	2.0	2.0		± 2.0
EOS	#	0.17	0.18	0.18	0.18		± 0.16	0.27	0.27	0.27	0.27		± 0.27	0.95	0.95	0.90	0.90		± 0.90
	%	7.0	7.5	7.0	7.0		± 6.5	3.5	3.5	3.5	3.5		± 3.5	5.5	5.5	5.0	5.0		± 5.0
BASO	#	0.08	0.08	0.08	0.08		± 0.08	0.26	0.26	0.23	0.23		± 0.23	0.78	0.76	0.45	0.45		± 0.45
	%	3.0	3.3	3.0	3.0		± 3.0	3.5	3.5	3.0	3.0		± 3.0	4.5	4.4	2.5	2.5		± 2.5

Ref: FORM-0821 Rev.10 FRONT / RECTO 9930050-A