

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 014

CONTROL

(Exp.) **2014-03-05**
 (YYYY-MM-DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	
		60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3			± 0.4	7.2	7.2			± 1.0	17.6	17.2			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.30			± 0.12	4.38	4.35			± 0.15	5.20	5.20			± 0.20	
HB HGB	g/dl	6.6	6.6			± 0.4	13.1	13.1			± 0.5	16.7	16.6			± 0.6	
	g/l	66	66			± 4	131	131			± 5	167	166			± 6	
	mmol/l	4.10	4.10			± 0.25	8.14	8.14			± 0.31	10.37	10.31			± 0.37	
HT HCT	%	18.6	18.9			± 1.5	35.9	35.7			± 2.0	45.8	45.8			± 2.5	
	l/l	0.186	0.189			± 0.015	0.359	0.357			± 0.020	0.458	0.458			± 0.025	
VGM MCV	µm ³ ; fl	79	82			± 5	82	82			± 5	88	88			± 5	
TGMH MCH	pg	28.1	28.7			± 2.0	29.9	30.1			± 2.0	32.1	31.9			± 2.5	
	fmol	1.74	1.78			± 0.12	1.86	1.87			± 0.12	1.99	1.98			± 0.16	
CCMH MCHC	g/dl	35.6	35.0			± 3.0	36.5	36.7			± 3.0	36.5	36.3			± 3.0	
	g/l	356	350			± 30	365	367			± 30	365	363			± 30	
IDR RDW	mmol/l	22.08	21.73			± 1.86	22.65	22.81			± 1.86	22.66	22.53			± 1.86	
	%	12.7	13.3			± 4.0	12.0	13.0			± 4.0	11.6	13.0			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	72			± 20	258	260			± 30	510	520			± 50	
VPM MPV	µm ³ ; fl	9.2	9.4			± 2.0	8.6	8.8			± 2.0	8.9	9.1			± 2.0	
NEUT	#	1.27	1.27			± 0.25	4.14	4.18			± 0.80	12.32	12.04			± 1.90	
	%	55.0	55.3			± 10.0	57.5	58.0			± 10.0	70.0	70.0			± 10.0	
LYMPHO	#	0.70	0.70			± 0.20	2.23	2.20			± 0.60	2.90	2.84			± 1.50	
	%	30.5	30.5			± 8.0	31.0	30.5			± 8.0	16.5	16.5			± 8.0	
MONO	#	0.12	0.10			± 0.10	0.32	0.29			± 0.29	0.79	0.69			± 0.69	
	%	5.0	4.5			± 4.5	4.5	4.0			± 4.0	4.5	4.0			± 4.0	
EOS	#	0.15	0.15			± 0.12	0.25	0.29			± 0.21	0.88	0.86			± 0.73	
	%	6.5	6.5			± 5.0	3.5	4.0			± 2.8	5.0	5.0			± 4.0	
BASO	#	0.07	0.07			± 0.07	0.25	0.25			± 0.19	0.70	0.77			± 0.46	
	%	3.0	3.2			± 3.0	3.5	3.5			± 2.5	4.0	4.5			± 2.5	

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 014

CONTROL

(Exp.) **2014-03-05**
 (YYYY-MM-DD)

Rev 1

		ABX Lyse																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20	
HB HGB	g/dl						± 0.4						± 0.5						± 0.6	
	g/l						± 4						± 5						± 6	
	mmol/l						± 0.25						± 0.31						± 0.37	
HT HCT	%						± 1.5						± 2.0						± 2.5	
	l/l						± 0.015						± 0.020						± 0.025	
VGM MCV	µm ³ ; fl						± 5						± 5						± 5	
TGMH MCH	pg						± 2.0						± 2.0						± 2.5	
	fmol						± 0.12						± 0.12						± 0.16	
CCMH MCHC	g/dl						± 3.0						± 3.0						± 3.0	
	g/l						± 30						± 30						± 30	
	mmol/l						± 1.86						± 1.86						± 1.86	
IDR RDW	%						± 4.0						± 4.0						± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						± 20						± 30						± 50	
VPM MPV	µm ³ ; fl						± 2.0						± 2.0						± 2.0	
NEUT	#						± 0.25						± 0.80						± 1.90	
	%						± 10.0						± 10.0						± 10.0	
LYMPHO	#						± 0.20						± 0.60						± 1.50	
	%						± 8.0						± 8.0						± 8.0	
MONO	#						± 0.10						± 0.29						± 0.69	
	%						± 4.5						± 4.0						± 4.0	
EOS	#						± 0.12						± 0.21						± 0.73	
	%						± 5.0						± 2.8						± 4.0	
BASO	#						± 0.07						± 0.19						± 0.46	
	%						± 3.0						± 2.5						± 2.5	

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B