

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 093
 Rev 1

CONTROL

(Exp.) **2013-11-05**
 (YYYY-MM-DD)

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.2	2.3			± 0.4	7.2	7.3			± 1.0	17.2	17.1			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.38			± 0.12	4.62	4.60			± 0.15	5.12	5.12			± 0.20			
HB HGB	g/dl	6.8	6.8			± 0.4	13.5	13.6			± 0.5	16.2	16.3			± 0.6			
	g/l	68	68			± 4	135	136			± 5	162	163			± 6			
	mmol/l	4.22	4.22			± 0.25	8.38	8.45			± 0.31	10.06	10.12			± 0.37			
HT HCT	%	19.2	19.5			± 1.5	37.4	37.7			± 2.0	45.1	45.6			± 2.5			
	l/l	0.192	0.195			± 0.015	0.374	0.377			± 0.020	0.451	0.456			± 0.025			
VGM MCV	µm ³ ; fl	79	82			± 5	81	82			± 5	88	89			± 5			
TGMH MCH	pg	28.0	28.6			± 2.0	29.2	29.6			± 2.0	31.6	31.8			± 2.5			
	fmol	1.74	1.77			± 0.12	1.81	1.84			± 0.12	1.96	1.98			± 0.16			
CCMH MCHC	g/dl	35.4	34.8			± 3.0	36.1	36.1			± 3.0	36.0	35.8			± 3.0			
	g/l	354	348			± 30	361	361			± 30	360	358			± 30			
	mmol/l	22.00	21.64			± 1.86	22.40	22.39			± 1.86	22.33	22.21			± 1.86			
IDR RDW	%	13.0	13.8			± 4.0	13.0	13.7			± 4.0	12.5	13.7			± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	65			± 20	240	245			± 30	500	505			± 50			
VPM MPV	µm ³ ; fl	9.1	9.2			± 2.0	9.0	9.2			± 2.0	9.0	9.2			± 2.0			
NEUT	#	1.18	1.29			± 0.25	4.08	4.16			± 0.75	12.13	12.06			± 1.80			
	%	53.5	56.0			± 10.0	56.6	57.0			± 10.0	70.5	70.5			± 10.0			
LYMPHO	#	0.70	0.70			± 0.20	2.28	2.26			± 0.60	2.84	2.74			± 1.40			
	%	32.0	30.3			± 8.0	31.6	31.0			± 8.0	16.5	16.0			± 8.0			
MONO	#	0.11	0.09			± 0.09	0.29	0.29			± 0.29	0.69	0.69			± 0.69			
	%	5.0	4.0			± 4.0	4.0	4.0			± 4.0	4.0	4.0			± 4.0			
EOS	#	0.15	0.15			± 0.15	0.31	0.33			± 0.30	0.77	0.86			± 0.62			
	%	6.5	6.5			± 6.5	4.3	4.5			± 4.0	4.5	5.0			± 3.5			
BASO	#	0.07	0.07			± 0.07	0.25	0.26			± 0.19	0.77	0.77			± 0.35			
	%	3.0	3.2			± 3.0	3.5	3.5			± 2.5	4.5	4.5			± 2.0			

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 093

CONTROL

(Exp.) **2013-11-05**
 (YYYY-MM-DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120			120 120 RETIC		DX120 DF120			120 120 RETIC	DX120 DF120							
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0					± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15					± 0.20	
HB HGB	g/dl						± 0.4						± 0.5					± 0.6	
	g/l						± 4						± 5					± 6	
	mmol/l						± 0.25						± 0.31					± 0.37	
HT HCT	%						± 1.5						± 2.0					± 2.5	
	l/l						± 0.015						± 0.020					± 0.025	
VGM MCV	µm ³ ; fl						± 5						± 5					± 5	
TGMH MCH	pg						± 2.0						± 2.0					± 2.5	
	fmol						± 0.12						± 0.12					± 0.16	
CCMH MCHC	g/dl						± 3.0						± 3.0					± 3.0	
	g/l						± 30						± 30					± 30	
	mmol/l						± 1.86						± 1.86					± 1.86	
IDR RDW	%						± 4.0						± 4.0					± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						± 20						± 30					± 50	
VPM MPV	µm ³ ; fl						± 2.0						± 2.0					± 2.0	
NEUT	#						± 0.25						± 0.75					± 1.80	
	%						± 10.0						± 10.0					± 10.0	
LYMPHO	#						± 0.20						± 0.60					± 1.40	
	%						± 8.0						± 8.0					± 8.0	
MONO	#						± 0.09						± 0.29					± 0.69	
	%						± 4.0						± 4.0					± 4.0	
EOS	#						± 0.15						± 0.30					± 0.62	
	%						± 6.5						± 4.0					± 3.5	
BASO	#						± 0.07						± 0.19					± 0.35	
	%						± 3.0						± 2.5					± 2.0	

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B