

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT		PX 033		CONTROL																	
		Rev 2																			
		ABX Lyse																			
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE		
		PENTRA						PENTRA						PENTRA							
		60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.4	7.4			± 1.0	17.7	17.5			± 2.2				
GR	RBC	10 ¹² /mm ³ ; 10 ¹² /l	2.45	2.38			± 0.12	4.62	4.60			± 0.15	5.07	5.09			± 0.20				
HB	HGB	g/dl	6.7	6.7			± 0.4	13.6	13.6			± 0.5	15.8	15.8			± 0.6				
		g/l	67	67			± 4	136	136			± 5	158	158			± 6				
		mmol/l	4.16	4.16			± 0.25	8.45	8.45			± 0.31	9.81	9.81			± 0.37				
HT	HCT	%	19.4	19.3			± 1.5	37.9	38.2			± 2.0	44.6	44.8			± 2.5				
		l/l	0.194	0.193			± 0.015	0.379	0.382			± 0.020	0.446	0.448			± 0.025				
VGM	MCV	µm ³ ; fl	79	81			± 5	82	83			± 5	88	88			± 5				
TGMH	MCH	pg	27.3	28.2			± 2.0	29.4	29.6			± 2.0	31.2	31.0			± 2.5				
		fmol	1.70	1.75			± 0.12	1.83	1.84			± 0.12	1.94	1.93			± 0.16				
		g/dl	34.6	34.8			± 3.0	35.9	35.6			± 3.0	35.4	35.3			± 3.0				
CCMH	MCHC	g/l	346	348			± 30	359	356			± 30	354	353			± 30				
		mmol/l	21.50	21.58			± 1.86	22.29	22.12			± 1.86	21.99	21.91			± 1.86				
		%	13.5	13.2			± 4.0	13.5	13.8			± 4.0	13.0	13.7			± 4.0				
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72			± 20	255	255			± 30	505	505			± 50				
VPM	MPV	µm ³ ; fl	8.9	8.9			± 2.0	9.1	9.2			± 2.0	8.6	8.8			± 2.0				
NEUT	#		1.37	1.36			± 0.30	4.04	4.03			± 0.80	12.16	11.73			± 1.80				
		%	57.0	56.5			± 10.0	54.6	54.5			± 10.0	68.7	67.0			± 10.0				
LYMPHO	#		0.71	0.72			± 0.20	2.44	2.44			± 0.60	3.10	3.41			± 1.50				
		%	29.5	30.0			± 8.0	33.0	33.0			± 8.0	17.5	19.5			± 8.0				
MONO	#		0.12	0.12			± 0.12	0.39	0.39			± 0.39	0.80	0.79			± 0.75				
		%	5.0	5.0			± 5.0	5.0	5.0			± 5.0	4.5	4.5			± 4.0				
EOS	#		0.13	0.13			± 0.13	0.31	0.33			± 0.31	0.89	0.88			± 0.74				
		%	5.5	5.5			± 5.0	4.0	4.5			± 4.0	5.0	5.0			± 4.0				
BASO	#		0.07	0.07			± 0.06	0.25	0.23			± 0.23	0.76	0.70			± 0.46				
		%	3.0	3.0			± 2.5	3.4	3.0			± 3.0	4.3	4.0			± 2.5				

(Exp.) 2013-05-05
(YYYY-MM-DD)

LOT		PX 033		CONTROL																	
		Rev 2																			
		ABX Lyse																			
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE		
		PENTRA						PENTRA						PENTRA							
		120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.7	7.7			± 1.0	18.5	18.5			± 2.2				
GR	RBC	10 ¹² /mm ³ ; 10 ¹² /l	2.44	2.44			± 0.12	4.64	4.64			± 0.15	5.17	5.17			± 0.20				
HB	HGB	g/dl	6.8	6.8			± 0.4	13.7	13.7			± 0.5	15.9	15.9			± 0.6				
		g/l	68	68			± 4	137	137			± 5	159	159			± 6				
		mmol/l	4.22	4.22			± 0.25	8.51	8.51			± 0.31	9.87	9.87			± 0.37				
HT	HCT	%	19.8	19.8			± 1.5	39.0	39.0			± 2.0	45.5	45.5			± 2.5				
		l/l	0.198	0.198			± 0.015	0.390	0.390			± 0.020	0.455	0.455			± 0.025				
VGM	MCV	µm ³ ; fl	81	81			± 5	84	84			± 5	88	88			± 5				
TGMH	MCH	pg	27.9	27.9			± 2.0	29.5	29.5			± 2.0	30.8	30.8			± 2.5				
		fmol	1.73	1.73			± 0.12	1.83	1.83			± 0.12	1.91	1.91			± 0.16				
		g/dl	34.4	34.4			± 3.0	35.1	35.1			± 3.0	34.9	34.9			± 3.0				
CCMH	MCHC	g/l	344	344			± 30	351	351			± 30	349	349			± 30				
		mmol/l	21.37	21.37			± 1.86	21.83	21.83			± 1.86	21.70	21.70			± 1.86				
		%	15.8	15.8			± 4.0	16.0	16.0			± 4.0	14.5	14.5			± 4.0				
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72			± 20	260	260			± 30	505	505			± 50				
VPM	MPV	µm ³ ; fl	8.9	8.9			± 2.0	9.1	9.1			± 2.0	8.6	8.6			± 2.0				
NEUT	#		1.37	1.39			± 0.30	4.24	4.39			± 0.80	12.80	12.80			± 1.80				
		%	57.0	58.0			± 10.0	55.0	57.0			± 10.0	69.0	69.0			± 10.0				
LYMPHO	#		0.73	0.65			± 0.20	2.54	2.27			± 0.60	3.42	3.15			± 1.50				
		%	30.5	27.0			± 8.0	33.0	29.5			± 8.0	18.5	17.0			± 8.0				
MONO	#		0.12	0.16			± 0.12	0.39	0.46			± 0.39	1.11	1.30			± 0.75				
		%	5.0	6.5			± 5.0	5.0	6.0			± 5.0	6.0	7.0			± 4.0				
EOS	#		0.13	0.13			± 0.13	0.31	0.31			± 0.31	0.74	0.74			± 0.74				
		%	5.0	5.5			± 5.0	4.0	4.0			± 4.0	4.0	4.0			± 4.0				
BASO	#		0.06	0.07			± 0.06	0.23	0.27			± 0.23	0.46	0.56			± 0.46				
		%	2.5	3.0			± 2.5	3.0	3.5			± 3.0	2.5	3.0			± 2.5				

(Exp.) 2013-05-05
(YYYY-MM-DD)