

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 112

CONTROL

(Exp.) **2013-01-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse												TOLERANCES TOLERANCE	
				CONTROL			L	CONTROL			N	CONTROL			H		
				PENTRA				PENTRA				PENTRA					
				60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4		± 0.4	7.4	7.3		± 1.0	17.6	17.5		± 2.2			
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.37	2.33		± 0.12	4.63	4.62		± 0.15	5.16	5.18		± 0.20			
HB	HGB	g/dl	6.6	6.6		± 0.4	13.6	13.6		± 0.5	16.3	16.2		± 0.6			
		g/l	66	66		± 4	136	136		± 5	163	162		± 6			
		mmol/l	4.10	4.10		± 0.25	8.45	8.45		± 0.31	10.12	10.06		± 0.37			
HT	HCT	%	18.5	19.1		± 1.5	37.5	37.9		± 2.0	45.4	45.6		± 2.5			
		l/l	0.185	0.191		± 0.015	0.375	0.379		± 0.020	0.454	0.456		± 0.025			
VGM	MCV	µm ³ ; fl	78	82		± 5	81	82		± 5	88	88		± 5			
TGMH	MCH	pg	27.8	28.3		± 2.0	29.4	29.4		± 2.0	31.6	31.3		± 2.5			
		fmol	1.73	1.76		± 0.12	1.82	1.83		± 0.12	1.96	1.94		± 0.16			
CCMH	MCHC	g/dl	35.7	34.5		± 3.0	36.3	35.9		± 3.0	35.9	35.5		± 3.0			
		g/l	357	345		± 30	363	359		± 30	359	355		± 30			
		mmol/l	22.17	21.45		± 1.86	22.52	22.29		± 1.86	22.29	22.07		± 1.86			
IDR	RDW	%	12.8	13.2		± 4.0	12.4	12.5		± 4.0	12.3	12.7		± 4.0			
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	68		± 20	250	256		± 30	500	510		± 50			
VPM	MPV	µm ³ ; fl	9.3	9.6		± 2.0	9.1	9.3		± 2.0	9.1	9.2		± 2.0			
NEUT		#	1.33	1.33		± 0.30	4.03	3.98		± 0.80	12.23	12.08		± 1.90			
		%	55.5	55.5		± 10.0	54.5	54.5		± 10.0	69.5	69.0		± 10.0			
LYMPHO		#	0.78	0.77		± 0.20	2.44	2.41		± 0.70	2.99	2.98		± 1.50			
		%	32.5	32.0		± 8.0	33.0	33.0		± 8.0	17.0	17.0		± 8.0			
MONO		#	0.08	0.08		± 0.08	0.33	0.33		± 0.33	0.70	0.70		± 0.70			
		%	3.5	3.5		± 3.5	4.5	4.5		± 4.5	4.0	4.0		± 4.0			
EOS		#	0.14	0.14		± 0.14	0.33	0.33		± 0.31	0.88	0.96		± 0.73			
		%	5.5	6.0		± 5.5	4.5	4.5		± 4.0	5.0	5.5		± 4.0			
BASO		#	0.08	0.08		± 0.08	0.26	0.26		± 0.23	0.79	0.79		± 0.55			
		%	3.0	3.0		± 3.0	3.5	3.5		± 3.0	4.5	4.5		± 3.0			

Ref: TEMP-0821 Rev.18 FRONT / RECTO 9930061-A

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 112
 Rev 1

CONTROL

(Exp.) **2013-01-05**
 (YYYY - MM - DD)

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120	DX 120					120	DX 120					120	DX 120				
120 RETIC	DF120			120 RETIC	DF120			120 RETIC	DF120			120 RETIC	DF120						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.7	7.7			± 1.0	18.2	18.2			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41			± 0.12	4.68	4.68			± 0.15	5.31	5.31			± 0.20			
HB HGB	g/dl	6.7	6.7			± 0.4	13.6	13.6			± 0.5	16.3	16.3			± 0.6			
	g/l	67	67			± 4	136	136			± 5	163	163			± 6			
	mmol/l	4.16	4.16			± 0.25	8.45	8.45			± 0.31	10.12	10.12			± 0.37			
HT HCT	%	19.5	19.5			± 1.5	38.4	38.4			± 2.0	46.7	46.7			± 2.5			
	l/l	0.195	0.195			± 0.015	0.384	0.384			± 0.020	0.467	0.467			± 0.025			
VGM MCV	µm ³ ; fl	81	81			± 5	82	82			± 5	88	88			± 5			
TGMH MCH	pg	27.8	27.8			± 2.0	29.1	29.1			± 2.0	30.7	30.7			± 2.5			
	f/mol	1.73	1.73			± 0.12	1.80	1.80			± 0.12	1.91	1.91			± 0.16			
CCMH MCHC	g/dl	34.3	34.3			± 3.0	35.4	35.4			± 3.0	34.9	34.9			± 3.0			
	g/l	343	343			± 30	354	354			± 30	349	349			± 30			
	mmol/l	21.31	21.31			± 1.86	22.01	22.01			± 1.86	21.66	21.66			± 1.86			
IDR RDW	%	15.7	15.7			± 4.0	16.2	16.2			± 4.0	15.0	15.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70			± 20	255	255			± 30	510	510			± 50			
VPM MPV	µm ³ ; fl	9.5	9.5			± 2.0	9.1	9.1			± 2.0	9.0	9.0			± 2.0			
NEUT	#	1.38	1.43			± 0.30	4.24	4.31			± 0.80	12.80	12.90			± 1.90			
	%	55.0	57.0			± 10.0	55.0	56.0			± 10.0	70.0	70.5			± 10.0			
LYMPHO	#	0.81	0.75			± 0.20	2.54	2.39			± 0.70	3.09	2.82			± 1.50			
	%	32.5	30.0			± 8.0	33.0	31.0			± 8.0	17.0	15.5			± 8.0			
MONO	#	0.10	0.10			± 0.08	0.39	0.46			± 0.33	1.09	1.18			± 0.70			
	%	4.0	4.0			± 3.5	5.0	6.0			± 4.5	6.0	6.5			± 4.0			
EOS	#	0.14	0.14			± 0.14	0.31	0.31			± 0.31	0.73	0.82			± 0.73			
	%	5.5	5.5			± 5.5	4.0	4.0			± 4.0	4.0	4.5			± 4.0			
BASO	#	0.08	0.09			± 0.08	0.23	0.23			± 0.23	0.55	0.55			± 0.55			
	%	3.0	3.5			± 3.0	3.0	3.0			± 3.0	3.0	3.0			± 3.0			

Ref: TEMP-0821 Rev.18 FRONT / RECTO 9930066-A