

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

PX 092

CONTROL

(Exp.) **2012-11-05**
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse												TOLERANCES TOLERANCE							
				CONTROL				L		CONTROL				N				CONTROL				H	
				PENTRA						PENTRA								PENTRA					
				60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3				± 0.4	7.5	7.5			± 1.0	17.6	17.3			± 2.2					
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.48	2.42			± 0.12	4.58	4.57			± 0.15	5.17	5.18			± 0.20						
HB	HGB	g/dl	7.0	7.0			± 0.4	13.4	13.4			± 0.5	16.3	16.3			± 0.6						
		g/l	70	70			± 4	134	134			± 5	163	163			± 6						
		mmol/l	4.35	4.35			± 0.25	8.32	8.32			± 0.31	10.12	10.12			± 0.37						
HT	HCT	%	19.8	19.8			± 1.5	37.1	37.5			± 2.0	45.5	45.6			± 2.5						
		l/l	0.198	0.198			± 0.015	0.371	0.375			± 0.020	0.455	0.456			± 0.025						
VGM	MCV	µm ³ ; fl	80	82			± 5	81	82			± 5	88	88			± 5						
TGMH	MCH	pg	28.2	28.9			± 2.0	29.3	29.3			± 2.0	31.5	31.5			± 2.5						
		fmol	1.75	1.80			± 0.12	1.82	1.82			± 0.12	1.96	1.95			± 0.16						
CCMH	MCHC	g/dl	35.3	35.3			± 3.0	36.1	35.8			± 3.0	35.8	35.8			± 3.0						
		g/l	353	353			± 30	361	358			± 30	358	358			± 30						
		mmol/l	21.91	21.91			± 1.86	22.43	22.21			± 1.86	22.25	22.21			± 1.86						
IDR	RDW	%	12.9	13.5			± 4.0	12.5	13.4			± 4.0	12.5	13.4			± 4.0						
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75			± 20	258	258			± 30	495	500			± 50						
VPM	MPV	µm ³ ; fl	8.6	8.6			± 2.0	9.1	9.1			± 2.0	8.2	8.2			± 2.0						
NEUT		#	1.29	1.31			± 0.30	4.16	4.13			± 0.80	12.18	11.76			± 1.90						
		%	56.0	57.0			± 10.0	55.5	55.0			± 10.0	69.2	68.0			± 10.0						
LYMPHO		#	0.71	0.68			± 0.20	2.48	2.48			± 0.70	3.10	3.11			± 1.50						
		%	31.0	29.5			± 8.0	33.0	33.0			± 8.0	17.6	18.0			± 8.0						
MONO		#	0.09	0.09			± 0.09	0.30	0.30			± 0.30	0.70	0.70			± 0.70						
		%	4.0	4.0			± 4.0	4.0	4.0			± 4.0	4.0	4.0			± 4.0						
EOS		#	0.14	0.14			± 0.12	0.30	0.34			± 0.24	0.88	1.04			± 0.73						
		%	6.0	6.0			± 5.0	4.0	4.5			± 3.0	5.0	6.0			± 4.0						
BASO		#	0.07	0.08			± 0.06	0.26	0.26			± 0.20	0.74	0.78			± 0.37						
		%	3.0	3.5			± 2.5	3.5	3.5			± 2.5	4.2	4.5			± 2.0						

Ref: TEMP-0821 Rev.18 FRONT / RECTO 9930061-A

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 092
 Rev 1

CONTROL

(Exp.) 2012-11-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse														TOLERANCES TOLERANCE				
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX 120 DF 120			120 120 RETIC	DX 120 DF 120			120 120 RETIC	DX 120 DF 120			120 120 RETIC	DX 120 DF 120					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4		7.9	7.9			± 1.0		18.3	18.3			± 2.2		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.48	2.48			± 0.12		4.65	4.65			± 0.15		5.29	5.29			± 0.20		
HB HGB	g/dl	7.1	7.1			± 0.4		13.3	13.3			± 0.5		16.3	16.3			± 0.6		
	g/l	71	71			± 4		133	133			± 5		163	163			± 6		
	mmol/l	4.41	4.41			± 0.25		8.26	8.26			± 0.31		10.12	10.12			± 0.37		
HT HCT	%	20.3	20.3			± 1.5		38.1	38.1			± 2.0		46.6	46.6			± 2.5		
	l/l	0.203	0.203			± 0.015		0.381	0.381			± 0.020		0.466	0.466			± 0.025		
VGM MCV	µm ³ ; fl	82	82			± 5		82	82			± 5		88	88			± 5		
TGMH MCH	pg	28.6	28.6			± 2.0		28.6	28.6			± 2.0		30.8	30.8			± 2.5		
	fmol	1.78	1.78			± 0.12		1.78	1.78			± 0.12		1.91	1.91			± 0.16		
CCMH MCHC	g/dl	34.9	34.9			± 3.0		34.9	34.9			± 3.0		35.0	35.0			± 3.0		
	g/l	349	349			± 30		349	349			± 30		350	350			± 30		
	mmol/l	21.68	21.68			± 1.86		21.66	21.66			± 1.86		21.74	21.74			± 1.86		
IDR RDW	%	15.0	15.0			± 4.0		16.0	16.0			± 4.0		15.0	15.0			± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	77	77			± 20		265	265			± 30		495	495			± 50		
VPM MPV	µm ³ ; fl	8.6	8.6			± 2.0		9.0	9.0			± 2.0		8.2	8.2			± 2.0		
NEUT	#	1.36	1.36			± 0.30		4.46	4.51			± 0.80		13.10	13.00			± 1.90		
	%	56.5	56.5			± 10.0		56.5	57.1			± 10.0		71.5	71.0			± 10.0		
LYMPHO	#	0.74	0.70			± 0.20		2.65	2.46			± 0.70		3.29	3.02			± 1.50		
	%	31.0	29.0			± 8.0		33.5	31.1			± 8.0		18.0	16.5			± 8.0		
MONO	#	0.12	0.16			± 0.09		0.36	0.43			± 0.30		0.82	1.01			± 0.70		
	%	5.0	6.5			± 4.0		4.5	5.5			± 4.0		4.5	5.5			± 4.0		
EOS	#	0.12	0.13			± 0.12		0.24	0.24			± 0.24		0.73	0.73			± 0.73		
	%	5.0	5.5			± 5.0		3.0	3.0			± 3.0		4.0	4.0			± 4.0		
BASO	#	0.06	0.06			± 0.06		0.20	0.26			± 0.20		0.37	0.55			± 0.37		
	%	2.5	2.5			± 2.5		2.5	3.3			± 2.5		2.0	3.0			± 2.0		

9930066-A
 Ref: TEMP-0821 Rev.18 FRONT / RECTO