

ABX Diffrol



LOT

PX 091

CONTROL



(Exp.) **2011-11-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE	
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120		MS60	60 60C+ ES60	80 XL80	120 120 RETIC		DX120 DF120	MS60	60 60C+ ES60	80 XL80			120 120 RETIC
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4	2.4	± 0.4		7.7	7.6	8.0	8.0		± 1.0	17.6	17.4	18.1			18.1
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.42	2.36	2.40	2.40	± 0.12	4.75	4.75	4.79	4.79	± 0.15	5.02	5.02	5.12	5.12	± 0.20			
HB HGB	g/dl	6.7	6.7	6.8	6.8	± 0.4	13.5	13.5	13.5	13.5	± 0.5	15.1	15.1	15.1	15.1	± 0.6			
	g/l	67	67	68	68	± 4	135	135	135	135	± 5	151	151	151	151	± 6			
	mmol/l	4.16	4.16	4.22	4.22	± 0.25	8.38	8.38	8.38	8.38	± 0.31	9.38	9.38	9.38	9.38	± 0.37			
HT HCT	%	19.1	19.4	19.4	19.4	± 1.5	38.0	38.5	39.3	39.3	± 2.0	42.2	42.7	44.0	44.0	± 2.5			
	l/l	0.191	0.194	0.194	0.194	± 0.015	0.380	0.385	0.393	0.393	± 0.020	0.422	0.427	0.440	0.440	± 0.025			
VGM MCV	µm ³ ; fl	79	82	81	81	± 5	80	81	82	82	± 5	84	85	86	86	± 5			
TGMH MCH	pg	27.7	28.4	28.3	28.3	± 2.0	28.4	28.4	28.2	28.2	± 2.0	30.1	30.1	29.5	29.5	± 2.5			
	fmol	1.72	1.76	1.76	1.76	± 0.12	1.76	1.76	1.75	1.75	± 0.12	1.87	1.87	1.83	1.83	± 0.16			
CCMH MCHC	g/dl	35.0	34.6	35.0	35.0	± 3.0	35.5	35.1	34.4	34.4	± 3.0	35.8	35.4	34.3	34.3	± 3.0			
	g/l	350	346	350	350	± 30	355	351	344	344	± 30	358	354	343	343	± 30			
	mmol/l	21.76	21.50	21.72	21.72	± 1.86	22.06	21.79	21.34	21.34	± 1.86	22.24	21.98	21.30	21.30	± 1.86			
IDR RDW	%	13.0	13.0	14.5	14.5	± 4.0	12.9	12.7	15.0	15.0	± 4.0	12.8	12.7	15.0	15.0	± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	73	73	73	± 20	250	250	253	253	± 30	485	485	480	480	± 50			
VPM MPV	µm ³ ; fl	8.9	9.0	9.0	9.0	± 2.0	8.6	8.7	8.7	8.7	± 2.0	8.4	8.6	8.5	8.5	± 2.0			
NEUT	#	1.40	1.44	1.37	1.43	± 0.30	4.00	4.00	4.20	4.40	± 0.90	11.90	11.57	12.10	12.20	± 1.90			
	%	58.3	60.0	56.9	59.4	± 10.0	52.0	52.6	52.5	55.0	± 10.0	67.6	66.5	67.0	67.5	± 10.0			
LYMPHO	#	0.66	0.62	0.69	0.61	± 0.25	2.58	2.54	2.76	2.44	± 0.70	3.27	3.41	3.71	3.44	± 1.50			
	%	27.4	25.8	28.9	25.4	± 8.0	33.5	33.4	34.5	30.5	± 8.0	18.6	19.6	20.5	19.0	± 8.0			
MONO	#	0.14	0.13	0.14	0.15	± 0.13	0.58	0.49	0.56	0.60	± 0.49	0.88	0.87	1.09	1.27	± 0.87			
	%	6.0	5.5	5.7	6.3	± 5.5	7.5	6.5	7.0	7.5	± 6.5	5.0	5.0	6.0	7.0	± 5.0			
EOS	#	0.12	0.13	0.14	0.15	± 0.12	0.27	0.30	0.24	0.28	± 0.24	0.79	0.78	0.72	0.72	± 0.72			
	%	5.0	5.5	6.0	6.4	± 5.0	3.5	4.0	3.0	3.5	± 3.0	4.5	4.5	4.0	4.0	± 4.0			
BASO	#	0.08	0.08	0.06	0.06	± 0.06	0.27	0.27	0.24	0.28	± 0.24	0.76	0.77	0.45	0.45	± 0.45			
	%	3.3	3.2	2.5	2.5	± 2.5	3.5	3.5	3.0	3.5	± 3.0	4.3	4.4	2.5	2.5	± 2.5			

9930052-B FRONT / RECTO Ref: FORM-0821 Rev.14